

Sample S014 | Total IgE, kU/l

Methodics	<i>x_{pt}</i>	Median	sd	CV%	SEM	min	max	Outliers	n
Phadia ImmunoCAP	132	132	8	6.4	2	116	153	-	19
Roche	147	147	6	3.8	4	143	151	-	2
Siemens BN ProSpec	-	-	-	-	-	63	63	-	1
Siemens Immulite 2000/2500	117	117	4	3.0	3	114	119	-	2
All	132	132	10	7.7	2	114	153	1	24

Sample S014 | Total IgE, kU/l| histogram summaries in LabScala

Sample S015 | ECP, µg/l

Methodics	<i>x_{pt}</i>	Median	sd	CV%	SEM	min	max	Outliers	n
Phadia ImmunoCAP	76.7	76.9	5.0	6.5	1.5	67.4	85.0	-	11
All	76.7	76.9	5.0	6.5	1.5	67.4	85.0	-	11

Sample S015 | ECP, µg/l| histogram summaries in LabScala