

## LABQUALITY

Chlamydia pneumoniae antibodies 2/2017

Laboratory XXXXX

Id-number 010

(5620) Results by methods and clinical interpretations.	Nega- tive	Equi- vocal /Weak	Posi- tive	Chlamydia pneum. Acute/ recent infection	No detec- table Chlamydia pneum. antibodies	No IgM antibodies	Exposed to Chlamydia pneumo- niae	No indication of acute infection	Antibody findings not diag- nostic	Lab. does not give clinical interpre- tations	A second sample is requested	Referred	A third sample is requested	Other	Results not returned	
<b>Specimen S001</b>																
Results not returned on time	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	4
All output groups	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	4
S-CpnAbG																
Micro If (Cpn); Labsystems	-	-	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF (Cpn); Euroimmun	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF, SeroFIA; Savyon	1	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro-IF; In-house	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgG/IgM; Vircell S.L.	-	-	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
EIA; AniLabsystems	-	-> 2	-> 5	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Novalisa; Novatec	1	-> 2	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA; SeroCP, Savyon	1	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA; Novagnost, Siemens	1	-> 4	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Alegria; Orgentec Diag.	-	-	-> 5	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Virion; Virion/Serion	-	-	-> 3	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgA/IgG/IgM; Euroimmun	-	-	-> 7	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA C.pneumoniae; MEDAC	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF; Focus Diagnostica	1	-	-> 3	-	-	-	-	-	-	-	-	-	-	-	-	-
IFA C.pneumoniae; Vircell S.L.	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
VirClia; Vircell S.L.	-	-	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Chorus; Diesse	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
CFT Chorus; Diesse	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
Incorrect test kit (C.trach)	-	-	-	-	-	1	-	-	-	-	-	-	-	-	-	-
All output groups	7	9	36	-	1	-	-	-	-	-	-	-	-	-	-	-
S-CpnAbM																
Micro If (Cpn); Labsystems	-	-> 1	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF (Cpn); Euroimmun	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF, SeroFIA; Savyon	-	-	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro-IF; In-house	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgG/IgM; Vircell S.L.	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
EIA; AniLabsystems	2	-> 3	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Novalisa; Novatec	-	-	-> 4	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA; SeroCP, Savyon	-	-	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-

Own results: Grey area

Expected results: -&gt;

Acute/recent infection not excluded: \*

The level of antibodies is low (borderline/negative); themajority of the tests give negative results but the moresensitive tests can give positive or borderline results (-&gt;)

Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copy (Copyright). Labquality prohibits any use of its name, or reference to Labquality, Labquality EQA program, or materials in this report in any advertising, brochures, or other commercial publications. Labquality EQA data don't necessarily indicate the superiority of reagents/equipment or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of reagents/equipment or materials may be deceptive and misleading.

## LABQUALITY

Chlamydia pneumoniae antibodies 2/2017

Laboratory XXXXX

Id-number 000

(5620) Results by methods and clinical interpretations.	Nega- tive	Equi- vocal/ Weak	Posi- tive	Chlamydia pneum. Acute/ recent infection	No detec- table Chlamydia pneum. antibodies	No IgM antibodies	Exposed to Chlamydia pneumo- niae	No indication of acute infection	Antibody findings	Lab. does not give not diag- nostic	A second sample is requested	Referred	A third sample is requested	Other	Results not returned
<b>Specimen S001</b>															
S-CpnAbM															
ELISA; Novagnost, Siemens	-	-	->	6	-	-	-	-	-	-	-	-	-	-	-
ELISA Alegria; Orgentec Diag.	-	-	->	4	-	-	-	-	-	-	-	-	-	-	-
ELISA Virion; Virion/Serion	-	-	->	2	-	-	-	-	-	-	-	-	-	-	-
ELISA IgA/IgG/IgM; Euroimmun	-	-	->	6	-	-	-	-	-	-	-	-	-	-	-
ELISA C.pneumoniae; MEDAC	-	-	->	1	-	-	-	-	-	-	-	-	-	-	-
Micro IF; Focus Diagnostica	4	-	-	-	-	-	-	-	-	-	-	-	-	-	-
IFA C.pneumoniae; Vircell S.L.	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-
VirClia; Vircell S.L.	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Chorus; Diesse	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-
CFT Chorus; Diesse	-	-	->	2	-	-	-	-	-	-	-	-	-	-	-
All output groups	13	4	33	-	-	-	-	-	-	-	-	-	-	-	-
S-CpnAbA															
Micro If (Cpn); Labsystems	->	1	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF, SeroFIA; Savyon	->	1	-	1	-	-	-	-	-	-	-	-	-	-	-
Micro-IF; In-house	->	1	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgG/IgM; Vircell S.L.	->	1	-	-	-	-	-	-	-	-	-	-	-	-	-
EIA; AniLabsystems	->	1	2	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Novalisa; Novatec	->	2	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA; SeroCP, Savyon	-	-	1	-	-	-	-	-	-	-	-	-	-	-	-
ELISA; Novagnost, Siemens	->	5	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Alegria; Orgentec Diag.	->	2	2	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Virion; Virion/Serion	->	3	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgA/IgG/IgM; Euroimmun	->	5	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA C.pneumoniae; MEDAC	->	1	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF; Focus Diagnostica	->	2	-	-	-	-	-	-	-	-	-	-	-	-	-
All output groups	25	4	2	-	-	-	-	-	-	-	-	-	-	-	-
<b>CL. INTERPRETATION</b>															
C.pneumoniae antibodies	-	-	-	->	33	2	-	11	-	5	1	->	12	->	2
All output groups	-	-	-	-	33	2	-	11	-	5	1	->	12	->	2

Own results: Grey area      Expected results: -&gt;.

Acute/recent infection not excluded: \*

The level of antibodies is low (borderline/negative); themajority of the tests give negative results but the moresensitive tests can give positive or borderline results (-&gt;)

Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copy (Copyright). Labquality prohibits any use of its name, or reference to Labquality, Labquality EQA program, or materials in this report in any advertising, brochures, or other commercial publications. Labquality EQA data don't necessarily indicate the superiority of reagents/equipment or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of reagents/equipment or materials may be deceptive and misleading.

(5620) Results by methods and clinical interpretations.	Nega- tive	Equi- vocal /Weak	Posi- tive	Chlamydia pneum. Acute/ recent infection	No detec- table Chlamydia pneum. antibodies	No IgM antibodies	Exposed to Chlamydia pneumo- niae	No indication of acute infection	Antibody findings not diag- nostic	Lab. does not give clinical interpre- tations	A second sample is requested	Referred	A third sample is requested	Other	Results not returned	
<b>Specimen S002</b>																
Results not returned on time	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	4
All output groups	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	4
S-CpnAbG																
Micro If (Cpn); Labsystems	-	-	> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF (Cpn); Euroimmun	-	-	> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF, SeroFIA; Savyon	-	-	> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro-IF; In-house	-	-	> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgG/IgM; Vircell S.L.	-	-	> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
EIA; AniLabsystems	-	-	> 7	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Novalisa; Novatec	-	1	> 3	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA; SeroCP, Savyon	-	-	> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA; Novagnost, Siemens	-	-	> 6	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Alegria; Ongentec Diag.	-	-	> 5	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Virion; Virion/Serion	1	-	> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgA/IgG/IgM; Euroimmun	-	-	> 7	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA C.pneumoniae; MEDAC	-	-	> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF; Focus Diagnostica	-	1	> 3	-	-	-	-	-	-	-	-	-	-	-	-	-
IFA C.pneumoniae; Vircell S.L	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
VirClia; Vircell S.L	-	-	> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Chorus; Diesse	-	-	> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
CFT Chorus; Diesse	-	-	> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
Incorrect test kit (C.trach)	-	-	-	-	1	-	-	-	-	-	-	-	-	-	-	-
All output groups	2	2	48	-	1	-	-	-	-	-	-	-	-	-	-	-
S-CpnAbM																
Micro If (Cpn); Labsystems	> 1	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF (Cpn); Euroimmun	> 1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF, SeroFIA; Savyon	> 2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro-IF; In-house	> 1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgG/IgM; Vircell S.L.	> 2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
EIA; AniLabsystems	> 7	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Novalisa; Novatec	> 4	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA; SeroCP, Savyon	> 2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-

Own results: Grey area      Expected results: -&gt;

Acute/recent infection not excluded: \*

The level of antibodies is low (borderline/negative); themajority of the tests give negative results but the moresensitive tests can give positive or borderline results (-&gt;)

## LABQUALITY

Chlamydia pneumoniae antibodies 2/2017

Laboratory XXXXX

Id-number 000

(5620) Results by methods and clinical interpretations.	Nega- tive	Equi- vocal /Weak	Posi- tive	Chlamydia pneum. Acute/ recent infection	No detec- table Chlamydia pneum. antibodies	No IgM antibodies	Exposed to Chlamydia pneumo- niae	No indication of acute infection	Antibody findings not diag- nostic	Lab. does not give clinical interpre- tations	A second sample is requested	Referred	A third sample is requested	Other	Results not returned
<b>Specimen S002</b>															
S-CpnAbM	-> 6	-	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA; Novagnost, Siemens	-> 4	-	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Alegria; Ongentec Diag.	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Virion; Virion/Serion	-> 6	-	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgA/IgG/IgM; Euroimmun	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA C.pneumoniae; MEDAC	-> 3	1	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF; Focus Diagnostica	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-	-
IFA C.pneumoniae; Vircell S.L	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-	-
VirClia; Vircell S.L	-> 48	2	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Chorus; Diesse	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-	-
CFT Chorus; Diesse	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-	-
All output groups															
S-CpnAbA	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-
Micro If (Cpn); Labsystems	1	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF, SeroFIA; Savyon	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro-IF; In-house	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgG/IgM; Vircell S.L.	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-
EIA; AniLabsystems	-	-	-> 3	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Novalisa; Novatec	-	-> 1	-	-> 1	-	-	-	-	-	-	-	-	-	-	-
ELISA; SeroCP, Savyon	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-
ELISA; Novagnost, Siemens	-	-> 2	-> 3	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Alegria; Ongentec Diag.	-	-	-> 4	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Virion; Virion/Serion	-	-	-> 3	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgA/IgG/IgM; Euroimmun	1	-> 2	-> 2	-	-	-	-	-	-	-	-	-	-	-	-
ELISA C.pneumoniae; MEDAC	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF; Focus Diagnostica	1	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
All output groups	3	7	21	-	-	-	-	-	-	-	-	-	-	-	-

Own results: Grey area      Expected results: -&gt;

Acute/recent infection not excluded: \*

The level of antibodies is low (borderline/negative); themajority of the tests give negative results but the moresensitive tests can give positive or borderline results (-&gt;)

Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copy (Copyright). Labquality prohibits any use of its name, or reference to Labquality, Labquality EQA program, or materials in this report in any advertising, brochures, or other commercial publications. Labquality EQA data don't necessarily indicate the superiority of reagents/equipment or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of reagents/equipment or materials may be deceptive and misleading.

## LABQUALITY

Chlamydia pneumoniae antibodies 2/2017

Laboratory XXXXX

Id-number 000

(5620) Results by methods and clinical interpretations.	Nega- tive	Equi- vocal /Weak	Posi- tive	Chlamydia pneum. Acute/ recent infection	No detec- table Chlamydia pneum. antibodies	No IgM antibodies	Exposed to Chlamydia pneumo- niae	No indication of acute infection	Antibody findings	Lab. does not give not diag- nostic	A second sample is requested	Referred	A third sample is requested	Other	Results not returned	
<b>Specimen S003</b>																
Results not returned on time	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	4
All output groups	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	4
S-CpnAbG																
Micro If (Cpn); Labsystems	-	-	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF (Cpn); Euroimmun	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF, SeroFIA; Savyon	-	-	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro-IF; In-house	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgG/IgM; Vircell S.L.	-	-	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
EIA; AniLabsystems	-	-	-> 7	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Novalisa; Novatec	-	-	-> 4	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA; SeroCP, Savyon	-	-	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA; Novagnost, Siemens	-	-	-> 5	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Alegria; Orgentec Diag.	-	-	-> 5	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Virion; Virion/Serion	1	-	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgA/IgG/IgM; Euroimmun	-	-	-> 7	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA C.pneumoniae; MEDAC	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF; Focus Diagnostica	-	1	-> 3	-	-	-	-	-	-	-	-	-	-	-	-	-
IFA C.pneumoniae; Vircell S.L	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
VirClia; Vircell S.L	-	-	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Chorus; Diesse	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
CFT Chorus; Diesse	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
Based only on the tests results	-	-	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-
Incorrect test kit (C.trach)	-	-	-	-	-	1	-	-	-	-	-	-	-	-	-	-
All output groups	1	1	50	-	1	-	-	-	-	-	-	-	-	-	-	-
S-CpnAbM																
Micro If (Cpn); Labsystems	-> 1	-	1	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF (Cpn); Euroimmun	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF, SeroFIA; Savyon	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro-IF; In-house	-> 1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgG/IgM; Vircell S.L.	-> 2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
EIA; AniLabsystems	-> 7	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Novalisa; Novatec	-> 4	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-

Own results: Grey area      Expected results: -&gt;

Acute/recent infection not excluded: \*

The level of antibodies is low (borderline/negative); themajority of the tests give negative results but the moresensitive tests can give positive or borderline results (-&gt;)

Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copy (Copyright). Labquality prohibits any use of its name, or reference to Labquality, Labquality EQA program, or materials in this report in any advertising, brochures, or other commercial publications. Labquality EQA data don't necessarily indicate the superiority of reagents/equipment or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of reagents/equipment or materials may be deceptive and misleading.X

(5620) Results by methods and clinical interpretations.	Nega- tive	Equi- vocal /Weak	Posi- tive	Chlamydia pneum. Acute/ recent infection	No detec- table Chlamydia pneum. antibodies	No IgM antibodies	Exposed to Chlamydia pneumo- niae	No indication of acute infection	Antibody findings not diag- nostic	Lab. does not give clinical interpre- tations	A second sample is requested	Referred	A third sample is requested	Other	Results not returned
<b>Specimen S003</b>															
S-CpnAbM															
ELISA; SeroCP, Savyon	->	2	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA; Novagnost, Siemens	->	5	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Alegria; Orgentec Diag.	->	4	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Virion; Virion/Serion	->	2	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgA/IgG/IgM; Euroimmun	->	6	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA C.pneumoniae; MEDAC	->	1	-	-	-	-	-	-	-	-	-	-	-	-	-
Micro IF; Focus Diagnostica	->	3	1	-	-	-	-	-	-	-	-	-	-	-	-
IFA C.pneumoniae; Vircell S.L	->	1	-	-	-	-	-	-	-	-	-	-	-	-	-
VirClia; Vircell S.L	->	2	-	-	-	-	-	-	-	-	-	-	-	-	-
ELISA Chorus; Diesse	->	1	-	-	-	-	-	-	-	-	-	-	-	-	-
CFT Chorus; Diesse	->	2	-	-	-	-	-	-	-	-	-	-	-	-	-
Based only on the tests results	->	1	-	-	-	-	-	-	-	-	-	-	-	-	-
All output groups	48	1	1	-	-	-	-	-	-	-	-	-	-	-	-
S-CpnAbA															
Micro If (Cpn); Labsystems	-	-	->	1	-	-	-	-	-	-	-	-	-	-	-
Micro IF, SeroFIA; Savyon	1	-	->	1	-	-	-	-	-	-	-	-	-	-	-
Micro-IF; In-house	-	->	1	-	-	-	-	-	-	-	-	-	-	-	-
ELISA IgG/IgM; Vircell S.L.	-	-	->	1	-	-	-	-	-	-	-	-	-	-	-
EIA; AniLabsystems	-	-	->	3	-	-	-	-	-	-	-	-	-	-	-
ELISA Novalisa; Novatec	-	->	1	->	1	-	-	-	-	-	-	-	-	-	-
ELISA; SeroCP, Savyon	-	-	->	1	-	-	-	-	-	-	-	-	-	-	-
ELISA; Novagnost, Siemens	-	->	2	->	3	-	-	-	-	-	-	-	-	-	-
ELISA Alegria; Orgentec Diag.	-	-	->	4	-	-	-	-	-	-	-	-	-	-	-
ELISA Virion; Virion/Serion	-	-	->	3	-	-	-	-	-	-	-	-	-	-	-
ELISA IgA/IgG/IgM; Euroimmun	1	->	2	->	2	-	-	-	-	-	-	-	-	-	-
ELISA C.pneumoniae; MEDAC	-	-	->	1	-	-	-	-	-	-	-	-	-	-	-
Micro IF; Focus Diagnostica	1	->	1	-	-	-	-	-	-	-	-	-	-	-	-
All output groups	3	7	21	-	-	-	-	-	-	-	-	-	-	-	-

Own results: Grey area

Expected results: -&gt;

Acute/recent infection not excluded: \*

The level of antibodies is low (borderline/negative); themajority of the tests give negative results but the moresensitive tests can give positive or borderline results (-&gt;)

## LABQUALITY

Chlamydia pneumoniae antibodies 2/2017

Laboratory XXXXX

Id-number 000

(5620) Results by methods and clinical interpretations.	Nega- tive	Equi- vocal /Weak	Posi- tive	Chlamydia pneum. Acute/ recent infection	No detec- table Chlamydia pneum. antibodies	No IgM antibodies	Exposed to Chlamydia pneumo- niae	No indication of acute infection	Antibody findings not diag- nostic	Lab. does not give clinical interpre- tations	A second sample is requested	Referred	A third sample is requested	Other	Results not returned
<b>Paired samples S002 and S003</b>															
CL. INTERPRETATION	-	-	-	6	-	-> 1	-> 38	-> 1	-> 5	1	-	-> 3	-> 6	2	-
C.pneumoniae antibodies	-	-	-	6	-	1	38	1	5	1	-	3	6	2	-
All output groups	-	-	-	6	-										

Own results: Grey area

Expected results: -&gt;

Acute/recent infection not excluded: \*

The level of antibodies is low (borderline/negative); themajority of the tests give negative results but the moresensitive tests can give positive or borderline results (-&gt;)

Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copy (Copyright). Labquality prohibits any use of its name, or reference to Labquality, Labquality EQA program, or materials in this report in any advertising, brochures, or other commercial publications. Labquality EQA data don't necessarily indicate the superiority of reagents/equipment or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of reagents/equipment or materials may be deceptive and misleading.