


Client report

	No of participants	No of responded participants	Response percentage
Helicobacter pylori, antibodies, December, 4-2017	59	52	88.1 %


Summary



Summary	Own score	Max score	Own success rate	Difference	AVR success rate
Sample S001	6	6	100 %	1.7 %	98.3 %
Sample S002	6	6	100 %	3.4 %	96.6 %
Average:			100 %	2.6 %	97.4 %


History	Test nr.	Own success rate	Difference	AVR success rate
Round 2017-3	1-1	100 %	3.4 %	96.6 %
Round 2017-2	1-1	100 %	3.1 %	96.9 %
Round 2017-1	1-1	100 %	4.2 %	95.8 %
Round 2016-4	1-1	100 %	9.6 %	90.4 %
Round 2016-3	1-1	100 %	3.5 %	96.5 %

Sample S001




Sample S001 results	Responded	Own score	Max score	Own success rate	Difference	AVR success rate	Count
	Clinical interpretation, qualitative methods	-	-	-	-	100 %	12
	Clinical interpretation, quantitative methods	4	4	100 %	0 %	100 %	42
	HepyAb	-	-	-	-	100 %	12
	HepyAbA	-	-	-	-	70 %	10
	HepyAbG	2	2	100 %	0 %	100 %	42
Total:		6	6	100 %	1.7 %	98.3 %	118

Sample S001 Clinical interpretation, qualitative methods




Sample S001 Clinical interpretation, quantitative methods




Clinical interpretation, qualitative methods	Interpretation	Interpretation count	Own score	Max score	Own success rate	Difference	AVR success rate
	Active or past infection	12					100 %
	Total:	12	-	-	-	-	100 %

Clinical interpretation, quantitative methods	Interpretation	Interpretation count	Own score	Max score	Own success rate	Difference	AVR success rate
	☒ Active or past infection	42	4	4	100 %	0 %	100 %
	Total:	42	4	4	100 %	0 %	100 %


Sample S001 HepyAb



Sample S001 HepyAbA



Sample S001 HepyAbG



HepyAb	Interpretation	Test	Interpretation count	Test count	Own score	Max score	Own success rate	Difference	AVR success rate
	Positive		12						100 %
		AccuBiotech ACCU-TELL Rapid H. pylori Antibody Test		1					
		Biomaxima HELICO H. pylori Ab		1					
		Liaison H. pylori Ab (Diasorin)		2					
		Meridian Bioscience Immunocard H. pylori Ab		1					
		Meridian Bioscience Premier H. pylori Ab		5					
		Standard Diagnostics SD Bioline H. pylori Ab		1					
		SureScreen Diagnostics H. pylori Ab Cassette		1					
	Total:		12		-	-	-	-	100 %

HepyAbA	Interpretation	Test	Interpretation count	Test count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		7						100 %
		Euroimmun Anti-Helicobacter pylori CagA IgA ELISA		1					
		Euroimmun Helicobacter pylori IgA EIA/ELISA		3					
		IBL Helicobacter pylori IgA ELISA		1					
		Siemens Enzygnost Anti-Helicobacter pylori IgA EIA		2					
	Borderline		1						0 %
		Euroimmun Helicobacter pylori IgA EIA/ELISA		1					
	Positive		2						0 %
		DiaPro Diagnostics HP IgA ELISA		1					
		Orgentec Alegria Anti-Helicobacter pylori IgA ELISA		1					
	Total:		10		-	-	-	-	70 %

Own device: EUROIMMUN

HepyAbG	Interpretation	Test	Interpretation count	Test count	Own score	Max score	Own success rate	Difference	AVR success rate
	☒ Positive		42		2	2	100 %	0 %	100 %
		Biohit GastroPanel H. pylori IgG ELISA		1					
		bioMerieux Vidas H. pylori IgG		5					
		DiaPro Diagnostics HP IgG ELISA		1					
		Diasorin Liaison H. pylori IgG		4					
		Diesse Chorus H. pylori IgG ELISA		1					
		☒ Euroimmun Helicobacter pylori IgG EIA/ELISA		12					
		In-house H. pylori IgG ELISA		1					
		Inova Diagnostics QUANTA Lite Helicobacter pylori ELISA		2					
		Meridian Bioscience Premier H. pylori IgG EIA		1					


		Orgentec Alegria Anti-Helicobacter pylori IgG ELISA		1					
		SERION classic Helicobacter pylori IgG ELISA		1					
		Siemens Enzygnost Anti-Helicobacter pylori IgG EIA		6					
		Siemens Immulite H. pylori IgG		4					
		Vircell Helicobacter pylori VirCla IgG		1					
		ZEUS H. pylori IgG ELISA		1					
	Total:		42		2	2	100 %	0 %	100 %

Sample S002




Sample S002 results	Responded	Own score	Max score	Own success rate	Difference	AVR success rate	Count
	Clinical interpretation, qualitative methods	-	-	-	-	100 %	12
	Clinical interpretation, quantitative methods	4	4	100 %	3.6 %	96.4 %	42
	HepyAb	-	-	-	-	100 %	12
	HepyAbA	-	-	-	-	-	10
	HepyAbG	2	2	100 %	6 %	94 %	42
Total:		6	6	100 %	3.4 %	96.6 %	118

Sample S002 Clinical interpretation, qualitative methods




Sample S002 Clinical interpretation, quantitative methods




Clinical interpretation, qualitative methods	Interpretation	Interpretation count	Own score	Max score	Own success rate	Difference	AVR success rate
	Active or past infection	12					100 %
	Total:	12	-	-	-	-	100 %

Clinical interpretation, quantitative methods	Interpretation	Interpretation count	Own score	Max score	Own success rate	Difference	AVR success rate
	No indication of infection	1					0 %
	Borderline interpretation	1					50 %
	Active or past infection	40	4	4	100 %	0 %	100 %
	Total:	42	4	4	100 %	3.6 %	96.4 %


Sample S002 HepyAb



Sample S002 HepyAbA



Sample S002 HepyAbG



HepyAb	Interpretation	Test	Interpretation count	Test count	Own score	Max score	Own success rate	Difference	AVR success rate
	Positive		12						100 %
		AccuBiotech ACCU-TELL Rapid H. pylori Antibody Test		1					
		Biomaxima HELICO H. pylori Ab		1					
		Liaison H. pylori Ab (Diasorin)		2					
		Meridian Bioscience Immunocard H. pylori Ab		1					
		Meridian Bioscience Premier H. pylori Ab		5					
		Standard Diagnostics SD Bioline H. pylori Ab		1					
		SureScreen Diagnostics H. pylori Ab Cassette		1					
	Total:		12		-	-	-	-	100 %

HepyAbA	Interpretation	Test	Interpretation count	Test count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		4						-
		Euroimmun Helicobacter pylori IgA EIA/ELISA		1					
		Orgentec Alegria Anti-Helicobacter pylori IgA ELISA		1					
		Siemens Enzygnost Anti-Helicobacter pylori IgA EIA		2					
	Positive		6						-
		DiaPro Diagnostics HP IgA ELISA		1					
		Euroimmun Anti-Helicobacter pylori CagA IgA ELISA		1					
		Euroimmun Helicobacter pylori IgA EIA/ELISA		3					
		IBL Helicobacter pylori IgA ELISA		1					
	Total:		10		-	-	-	-	

Own device: EUROIMMUN

HepyAbG	Interpretation	Test	Interpretation count	Test count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative		2						0 %
		DiaPro Diagnostics HP IgG ELISA		1					
		Euroimmun Helicobacter pylori IgG EIA/ELISA		1					
	Borderline		1						50 %
		Siemens Enzygnost Anti-Helicobacter pylori IgG EIA		1					
	⦿ Positive		39		2	2	100 %	0 %	100 %
		Biohit GastroPanel H. pylori IgG ELISA		1					
		bioMerieux Vidas H. pylori IgG		5					
		Diasorin Liaison H. pylori IgG		4					
		Diesse Chorus H. pylori IgG ELISA		1					
		⦿ Euroimmun Helicobacter pylori IgG EIA/ELISA		11					

Copyright © Labquality Oy
Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copyright. Labquality prohibits any use of its name, or reference to Labquality EQA program, or material in this report in any advertising, brochures or other commercial publications. Labquality EQA data do not necessarily indicate the superiority of instruments, reagents, testing equipments or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of equipments or materials may be deceptive and misleading. Proficiency test results are handled confidentially. Labquality will not issue any statements to third parties of the performance of laboratories in external quality assessment schemes unless otherwise agreed.

		In-house H. pylori IgG ELISA		1					
		Inova Diagnostics QUANTA Lite Helicobacter pylori ELISA		2					
		Meridian Bioscience Premier H. pylori IgG EIA		1					
		Orgentec Alegria Anti-Helicobacter pylori IgG ELISA		1					
		SERION classic Helicobacter pylori IgG ELISA		1					
		Siemens Enzygnost Anti-Helicobacter pylori IgG EIA		5					
		Siemens Immulite H. pylori IgG		4					
		Vircell Helicobacter pylori VirClia IgG		1					
		ZEUS H. pylori IgG ELISA		1					
	Total:		42		2	2	100 %	6 %	94 %