

	In all	Posi- tive	mecA+	mecC-	mecA/C+
001 MRSA screening: positive					
MRSA screening					
Positive, sent to reference laboratory	=> 15	-	-	-	-
Positive, not sent to reference laboratory	=> 26	-	-	-	-
Suspect positive, sent to reference laboratory	=> 2	-	-	-	-
Results not returned	1	-	-	-	-
Plate culture					
Selective medium	20	-	-	-	-
Chromogenic medium	37	-	-	-	-
Agglutination test					
Diamondial Staph Plus Latex Kit (DiaMondialL)	-	1	-	-	-
Pastorex Staph Plus (Bio-Rad)	-	3	-	-	-
PBP2a Culture Colony Test (Alere)	-	3	-	-	-
PBP2a Latex Agglutination test (Oxoid, Thermo Scientific)	-	2	-	-	-
Test name not reported	-	1	-	-	-
Confirmation test					
BD MAX MRSA XT Assay (Becton Dickinson)	-	-	-	-	1
GenomEra MRSA/SA (Abacus Diagnostica)	-	-	4	-	1
GenoType MRSA (Hain Lifescience)	-	-	1	-	-
Xpert MRSA (Cepheid)	-	-	1	-	-
Xpert MRSA Gen3 (Cepheid)	-	-	-	-	2
Xpert MRSA NxG (Cepheid)	-	-	-	-	3
Xpert SA Nasal Complete (Cepheid)	-	-	1	-	2
PCR (In-house)	-	-	5	3	-
Direct detection					
BD MAX MRSA XT Assay (Becton Dickinson)	-	-	-	-	1
Xpert MRSA (Cepheid)	-	-	1	-	-
Xpert MRSA NxG (Cepheid)	-	1	-	-	2
Xpert MRSA/SA BC (Cepheid)	-	-	1	-	-
Xpert SA Nasal Complete (Cepheid)	-	2	-	-	1
PCR (In-house)	-	1	1	-	-

own result: grey are

expected result: =>

Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copy (Copyright). Labquality prohibits any use of its name, or reference to Labquality, Labquality EQA program, or materials in this report in any advertising, brochures, or other commercial publications. Labquality EQA data don't necessarily indicate the superiority of reagents/equipment or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of reagents/equipment or materials may be deceptive and misleading.