

LABQUALITY Surveillance for multidrug resistant bacteria, MRSA, 2, 2017.
NUMERICAL SUMMARY Sample 001 *Staphylococcus aureus* ATCC 43300 (MRSA, mecA)

Assay	Method	$\bar{x}$	SD	Interpret. pcs (%)					All
		mm	mm	S	I	R	pcs		
Cefoxitin	CLSI	10	4	0	(0 %)	0	(0 %)	6 (100 %)	6
	CA-SFM	13	2	0	(0 %)	0	(0 %)	2 (100 %)	2
	EUCAST	10	3	0	(0 %)	0	(0 %)	19 (100 %)	19
	NORDIC AST	9	-	0	(0 %)	0	(0 %)	1 (100 %)	1
	All results			0	(0 %)	0	(0 %)	28 (100 %)	28

Results by disk diffusion method.

- Guidelines:
- CA-SFM, Comité Antibiotogramme - Société Française de Microbiologie
 - CLSI, Clinical and Laboratory Standards Institute
 - EUCAST, European Committee on Antimicrobial Susceptibility Testing
 - NORDIC AST, Nordic Committee on Antimicrobial Susceptibility Testing