

LABQUALITY

Neisseria gonorrhoeae culture, identification tests 4/2017

Results of Gramstaining and identification tests.		In all	Positive +	Negative -	Neisseria sp.	Neisseria gonorrhoeae	Other genus than Neisseria
S001: K. denitrificans, Lactobacillus sp.							
Gram staining:	gramnegative coccobacilli	6	-	-	-	-	-
	gramnegative diplococcus	2	-	-	-	-	-
	gramnegative rod	4	-	-	-	-	-
	grampositive cocci in chains	1	-	-	-	-	-
	grampositive rod	3	-	-	-	-	-
Oxidase		-	9	3	-	-	-
Commercial id.kits:	API NH (bioMerieux)	-	-	-	1	1	-
	VITEK 2 (bioMerieux)	-	-	-	1	-	3
Other tests:	Aptima Combo 2 (GenProbe)	-	-	1	-	-	-
	BD MAX CT/GC/TV (BD)	-	-	1	-	-	-
	VITEK MS (bioMerieux)	-	-	-	-	-	8
	MALDI-TOF (Bruker Daltonics)	-	-	-	-	-	10

Own result: grey area

Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copy (Copyright). Labquality prohibits any use of its name, or reference to Labquality, Labquality EQA program, or materials in this report in any advertising, brochures, or other commercial publications. Labquality EQA data don't necessarily indicate the superiority of reagents/equipment or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of reagents/equipment or materials may be deceptive and misleading.

LABQUALITY

Neisseria gonorrhoeae culture, identification tests 4/2017

Results of Gramstaining and identification tests.		In all	Positive +	Negative -	Neisseria sp.	Neisseria gonorrhoeae	Other genus than Neisseria
S002: N. gonorrhoeae, P. aeruginosa							
Gram staining:	gramnegative coccobacilli	1	-	-	-	-	-
	gramnegative coccus	7	-	-	-	-	-
	gramnegative diplococcus	21	-	-	-	-	-
Oxidase		-	29	-	-	-	-
Commercial id.kits: API NH (bioMerieux)		-	-	-	-	4	-
	BBL Crystal N/H (Becton Dickinson)	-	-	-	-	1	-
	VITEK 2 (bioMerieux)	-	-	-	-	5	1
Antigen detection: Phadebact Monocl. GC Test (MKL Diagn.)		-	-	-	-	2	-
Other tests:	Aptima Combo 2 (GenProbe)	-	-	-	-	4	-
	Abbott RealTime CT/NG (Abbott)	-	-	-	-	2	-
	BD MAX CT/GC/TV (BD)	-	1	-	-	-	-
	VITEK MS (bioMerieux)	-	-	-	-	14	1
	MALDI-TOF (Bruker Daltonics)	-	-	-	-	16	3
	Cobas 4800 System (Roche Diagnostics)	-	1	-	-	-	-

Own result: grey area

Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copy (Copyright). Labquality prohibits any use of its name, or reference to Labquality, Labquality EQA program, or materials in this report in any advertising, brochures, or other commercial publications. Labquality EQA data don't necessarily indicate the superiority of reagents/equipment or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of reagents/equipment or materials may be deceptive and misleading.