

## Veterinary basic chemistry 1-2018

**Urgent!** Please analyze the samples as soon as possible, - preferably on the arrival day or the following day at the latest!

### Specimens

Specimens S001 (LQ780518011) and S002 (LQ780518012) are animal origin serum samples. Both samples are ready for use. Specimens should be handled as carefully as patient samples.

### Storage

If you are not able to analyse the samples on the arrival day, the samples should be transferred immediately to a refrigerator and stored in the dark.

### Use

Allow the samples to reach room temperature. Mix the specimens by quickly inverting the tube several times, until the suspension appears homogeneous.

### Analysis

Analyze the chemistry parameters that are in your laboratory's test selection for serum and report your results on our electronic form at [my.labscala.eu](http://my.labscala.eu).

### Reporting the results

Fill your results on [my.labscala.fi](http://my.labscala.fi). In case you have forgotten your password, please contact our client services at [info@labquality.fi](mailto:info@labquality.fi). Instructions on how to use the electronic form can be found on our YouTube channel at <https://www.youtube.com/watch?v=SyRtvTEV8iY>.

### Report

Participants who have sent their results by November 25 will receive a laboratory specific report after the scheme is closed.

## Instructions

Product no. 8530  
LQ780518011-012 (UK)

**The shipment includes**  
- 2 serum samples

If the kit is incomplete or contains damaged specimens, please report immediately to the EQA coordinator.

### Please note:

The results should be at Labquality no later than **November 25, 2018.**

### Inquiries

#### EQA Coordinator

Jonna Pelanti  
tel. +358 50 3706110  
[jonna.pelanti@labquality.fi](mailto:jonna.pelanti@labquality.fi)

**Labquality**  
Kumpulantie 15  
FI-00520 HELSINKI

**Telephone**  
+358 9 8566 8200

**Fax**  
+358 9 8566 8280

**Info@labquality.fi**  
**[www.labquality.fi](http://www.labquality.fi)**

© Labquality



Specimen S001 | Alkaline phosphatase, U/l

| Methodics                               | $x_{pt}$ | Median | sd    | CV%  | SEM   | min   | max   | Outliers | n |
|-----------------------------------------|----------|--------|-------|------|-------|-------|-------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -     | -    | -     | 12.00 | 12.00 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -     | -    | -     | 3.00  | 3.00  | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 12.00    | 12.00  | <0.01 | <0.1 | <0.01 | 12.00 | 12.00 | -        | 2 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -     | -    | -     | 12.00 | 12.00 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -     | -    | -     | 5.00  | 5.00  | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -     | -    | -     | 13.00 | 13.00 | -        | 1 |
| All                                     | 9.86     | 12.00  | 4.06  | 41.2 | 1.53  | 3.00  | 13.00 | -        | 7 |

Specimen S001 | Alkaline phosphatase, U/l| histogram summaries in LabScala

Specimen S001 | Alanine aminotransferase, U/l

| Methodics                               | $x_{pt}$ | Median | sd    | CV%  | SEM   | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|-------|------|-------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -     | -    | -     | 4.00 | 4.00 | -        | 1 |
| BS-120 (Mindray)                        | 3.00     | 3.00   | <0.01 | <0.1 | <0.01 | 3.00 | 3.00 | -        | 2 |
| Indiko (Thermo Fischer Scientific)      | 3.00     | 3.00   | 1.41  | 47.1 | 1.00  | 2.00 | 4.00 | -        | 2 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -     | -    | -     | 2.00 | 2.00 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -     | -    | -     | 6.00 | 6.00 | -        | 1 |
| All                                     | 3.43     | 3.00   | 1.40  | 40.8 | 0.53  | 2.00 | 6.00 | -        | 7 |

Specimen S001 | Alanine aminotransferase, U/l| histogram summaries in LabScala

Specimen S001 | Glutamyl transferase, U/l

| Methodics                          | $x_{pt}$ | Median | sd    | CV%  | SEM   | min  | max  | Outliers | n |
|------------------------------------|----------|--------|-------|------|-------|------|------|----------|---|
| Advia 1800 (Siemens)               | -        | -      | -     | -    | -     | 1.00 | 1.00 | -        | 1 |
| Indiko (Thermo Fischer Scientific) | -        | -      | -     | -    | -     | 1.00 | 1.00 | -        | 1 |
| All                                | 1.00     | 1.00   | <0.01 | <0.1 | <0.01 | 1.00 | 1.00 | -        | 2 |

Specimen S001 | Glutamyl transferase, U/l| histogram summaries in LabScala

Specimen S001 | Amylase, U/l

| Methodics                          | $x_{pt}$ | Median | sd    | CV% | SEM   | min    | max    | Outliers | n |
|------------------------------------|----------|--------|-------|-----|-------|--------|--------|----------|---|
| Advia 1800 (Siemens)               | -        | -      | -     | -   | -     | 570.00 | 570.00 | -        | 1 |
| Indiko (Thermo Fischer Scientific) | -        | -      | -     | -   | -     | 598.00 | 598.00 | -        | 1 |
| Vetscan VS2 (Abaxis)               | -        | -      | -     | -   | -     | 544.00 | 544.00 | -        | 1 |
| All                                | 570.67   | 570.00 | 27.01 | 4.7 | 15.59 | 544.00 | 598.00 | -        | 3 |

Specimen S001 | Amylase, U/l| histogram summaries in LabScala

Specimen S001 | Albumin, g/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV%  | SEM  | min   | max   | Outliers | n  |
|-----------------------------------------|----------|--------|------|------|------|-------|-------|----------|----|
| Advia 1800 (Siemens)                    | -        | -      | -    | -    | -    | 27.60 | 27.60 | -        | 1  |
| BS-120 (Mindray)                        | 29.45    | 29.45  | 1.34 | 4.6  | 0.95 | 28.50 | 30.40 | -        | 2  |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -    | -    | 23.00 | 23.00 | -        | 1  |
| Indiko (Thermo Fischer Scientific)      | 24.13    | 24.00  | 0.42 | 1.7  | 0.24 | 23.80 | 24.60 | -        | 3  |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -    | -    | -    | 25.00 | 25.00 | -        | 1  |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -    | -    | 24.80 | 24.80 | -        | 1  |
| Vetscan VS2 (Abaxis)                    | -        | -      | -    | -    | -    | 20.00 | 20.00 | -        | 1  |
| All                                     | 25.17    | 24.70  | 2.97 | 11.8 | 0.94 | 20.00 | 30.40 | -        | 10 |

Specimen S001 | Albumin, g/l| histogram summaries in LabScala

Specimen S001 | Protein, g/l

| Methodics | $x_{pt}$ | Median | sd | CV% | SEM | min | max | Outliers | n |
|-----------|----------|--------|----|-----|-----|-----|-----|----------|---|
|-----------|----------|--------|----|-----|-----|-----|-----|----------|---|

Copyright © Labquality Oy  
Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copyright. Labquality prohibits any use of its name, or reference to Labquality EQA program, or material in this report in any advertising, brochures or other commercial publications. Labquality EQA data do not necessarily indicate the superiority of instruments, reagents, testing equipments or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of equipments or materials may be deceptive and misleading. Proficiency test results are handled confidentially. Labquality will not issue any statements to third parties of the performance of laboratories in external quality assessment schemes unless otherwise agreed.

|                                         |       |       |      |     |      |       |       |   |    |
|-----------------------------------------|-------|-------|------|-----|------|-------|-------|---|----|
| Advia 1800 (Siemens)                    | -     | -     | -    | -   | -    | 61.70 | 61.70 | - | 1  |
| BS-120 (Mindray)                        | 61.95 | 61.95 | 1.77 | 2.9 | 1.25 | 60.70 | 63.20 | - | 2  |
| Fuji NX500 (Fujifilm)                   | -     | -     | -    | -   | -    | 59.00 | 59.00 | - | 1  |
| Indiko (Thermo Fischer Scientific)      | 59.00 | 59.00 | 1.00 | 1.7 | 0.58 | 58.00 | 60.00 | - | 3  |
| Konelab 30 (Thermo Fischer Scientific)  | -     | -     | -    | -   | -    | 57.00 | 57.00 | - | 1  |
| Konelab 60i (Thermo Fischer Scientific) | -     | -     | -    | -   | -    | 59.00 | 59.00 | - | 1  |
| Vetscan VS2 (Abaxis)                    | -     | -     | -    | -   | -    | 61.00 | 61.00 | - | 1  |
| All                                     | 59.86 | 59.50 | 1.84 | 3.1 | 0.58 | 57.00 | 63.20 | - | 10 |

Specimen S001 | Protein, g/l| histogram summaries in LabScala

Specimen S001 | Bilirubin total, µmol/l

| Methodics                               | $x_{pt}$ | Median | sd  | CV%  | SEM | min | max | Outliers | n |
|-----------------------------------------|----------|--------|-----|------|-----|-----|-----|----------|---|
| BS-120 (Mindray)                        | -        | -      | -   | -    | -   | 0.9 | 0.9 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 1.1      | 1.1    | 0.6 | 60.6 | 0.5 | 0.6 | 1.5 | -        | 2 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -   | -    | -   | 1.7 | 1.7 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -   | -    | -   | 1.6 | 1.6 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -   | -    | -   | 4.0 | 4.0 | -        | 1 |
| All                                     | 1.7      | 1.6    | 1.2 | 70.5 | 0.5 | 0.6 | 4.0 | -        | 6 |

Specimen S001 | Bilirubin total, µmol/l| histogram summaries in LabScala

Specimen S001 | Urea, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV% | SEM  | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|------|-----|------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -   | -    | 8.90 | 8.90 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -    | -   | -    | 8.66 | 8.66 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -   | -    | 8.43 | 8.43 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 8.07     | 7.90   | 0.67 | 8.3 | 0.38 | 7.50 | 8.80 | -        | 3 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -    | -   | -    | 7.20 | 7.20 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -   | -    | 8.44 | 8.44 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -    | -   | -    | 7.40 | 7.40 | -        | 1 |
| All                                     | 8.14     | 8.43   | 0.65 | 8.0 | 0.22 | 7.20 | 8.90 | -        | 9 |

Specimen S001 | Urea, mmol/l| histogram summaries in LabScala

Specimen S001 | Creatinine, µmol/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV%  | SEM | min   | max   | Outliers | n  |
|-----------------------------------------|----------|--------|------|------|-----|-------|-------|----------|----|
| Advia 1800 (Siemens)                    | -        | -      | -    | -    | -   | 100.0 | 100.0 | -        | 1  |
| BS-120 (Mindray)                        | 93.7     | 93.7   | 8.7  | 9.3  | 6.2 | 87.5  | 99.9  | -        | 2  |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -    | -   | 127.0 | 127.0 | -        | 1  |
| Indiko (Thermo Fischer Scientific)      | 99.7     | 100.0  | 3.5  | 3.5  | 2.0 | 96.0  | 103.0 | -        | 3  |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -    | -    | -   | 95.0  | 95.0  | -        | 1  |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -    | -   | 102.0 | 102.0 | -        | 1  |
| Vetscan VS2 (Abaxis)                    | -        | -      | -    | -    | -   | 100.0 | 100.0 | -        | 1  |
| All                                     | 101.0    | 100.0  | 10.2 | 10.0 | 3.2 | 87.5  | 127.0 | -        | 10 |

Specimen S001 | Creatinine, µmol/l| histogram summaries in LabScala

Specimen S001 | Calcium, ionized, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV% | SEM  | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|------|-----|------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -   | -    | 2.02 | 2.02 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -   | -    | 2.15 | 2.15 | -        | 1 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -    | -   | -    | 1.95 | 1.95 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -   | -    | 2.09 | 2.09 | -        | 1 |
| All                                     | 2.05     | 2.06   | 0.09 | 4.2 | 0.04 | 1.95 | 2.15 | -        | 4 |

Specimen S001 | Calcium, ionized, mmol/l| histogram summaries in LabScala

Specimen S001 | Glucose, mmol/l

Copyright © Labquality Oy  
Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copyright. Labquality prohibits any use of its name, or reference to Labquality EQA program, or material in this report in any advertising, brochures or other commercial publications. Labquality EQA data do not necessarily indicate the superiority of instruments, reagents, testing equipments or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of equipments or materials may be deceptive and misleading. Proficiency test results are handled confidentially. Labquality will not issue any statements to third parties of the performance of laboratories in external quality assessment schemes unless otherwise agreed.

| Methodics                               | $x_{pt}$ | Median | sd   | CV% | SEM  | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|------|-----|------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -   | -    | 3.00 | 3.00 | -        | 1 |
| BS-120 (Mindray)                        | 3.31     | 3.31   | 0.04 | 1.1 | 0.03 | 3.28 | 3.33 | -        | 2 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -   | -    | 2.90 | 2.90 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 2.97     | 3.00   | 0.06 | 1.9 | 0.03 | 2.90 | 3.00 | -        | 3 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -   | -    | 3.10 | 3.10 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -    | -   | -    | 3.10 | 3.10 | -        | 1 |
| All                                     | 3.07     | 3.00   | 0.15 | 5.0 | 0.05 | 2.90 | 3.33 | -        | 9 |

Specimen S001 | Glucose, mmol/l| histogram summaries in LabScala

Specimen S001 | Sodium, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd  | CV% | SEM | min   | max   | Outliers | n |
|-----------------------------------------|----------|--------|-----|-----|-----|-------|-------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -   | -   | -   | 142.4 | 142.4 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -   | -   | -   | 145.0 | 145.0 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -   | -   | -   | 142.0 | 142.0 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 146.3    | 145.0  | 2.3 | 1.6 | 1.3 | 145.0 | 149.0 | -        | 3 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -   | -   | -   | 144.0 | 144.0 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -   | -   | -   | 144.0 | 144.0 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -   | -   | -   | 137.0 | 137.0 | -        | 1 |
| All                                     | 143.7    | 144.0  | 3.2 | 2.2 | 1.1 | 137.0 | 149.0 | -        | 9 |

Specimen S001 | Sodium, mmol/l| histogram summaries in LabScala

Specimen S001 | Potassium, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV% | SEM  | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|------|-----|------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -   | -    | 4.96 | 4.96 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -    | -   | -    | 5.20 | 5.20 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -   | -    | 4.80 | 4.80 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 4.90     | 4.90   | 0.10 | 2.0 | 0.06 | 4.80 | 5.00 | -        | 3 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -   | -    | 4.97 | 4.97 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -    | -   | -    | 5.00 | 5.00 | -        | 1 |
| All                                     | 4.95     | 4.97   | 0.13 | 2.6 | 0.05 | 4.80 | 5.20 | -        | 8 |

Specimen S001 | Potassium, mmol/l| histogram summaries in LabScala

Specimen S001 | Phosphorus, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV% | SEM  | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|------|-----|------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -   | -    | 1.62 | 1.62 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -    | -   | -    | 1.58 | 1.58 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -   | -    | 1.70 | 1.70 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 1.57     | 1.55   | 0.08 | 4.9 | 0.04 | 1.50 | 1.65 | -        | 3 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -    | -   | -    | 1.43 | 1.43 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -   | -    | 1.56 | 1.56 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -    | -   | -    | 1.79 | 1.79 | -        | 1 |
| All                                     | 1.60     | 1.58   | 0.11 | 6.7 | 0.04 | 1.43 | 1.79 | -        | 9 |

Specimen S001 | Phosphorus, mmol/l| histogram summaries in LabScala

Specimen S001 | GLOB, g/l

| Methodics             | $x_{pt}$ | Median | sd | CV% | SEM | min | max | Outliers | n |
|-----------------------|----------|--------|----|-----|-----|-----|-----|----------|---|
| Advia 1800 (Siemens)  | -        | -      | -  | -   | -   | 34  | 34  | -        | 1 |
| Fuji NX500 (Fujifilm) | -        | -      | -  | -   | -   | 36  | 36  | -        | 1 |
| Vetscan VS2 (Abaxis)  | -        | -      | -  | -   | -   | 41  | 41  | -        | 1 |
| All                   | 37       | 36     | 4  | 9.6 | 2   | 34  | 41  | -        | 3 |

Specimen S001 | GLOB, g/l| histogram summaries in LabScala

Specimen S001 | Aspartate aminotransferase, U/l

Copyright © Labquality Oy  
Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copyright. Labquality prohibits any use of its name, or reference to Labquality EQA program, or material in this report in any advertising, brochures or other commercial publications. Labquality EQA data do not necessarily indicate the superiority of instruments, reagents, testing equipments or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of equipments or materials may be deceptive and misleading. Proficiency test results are handled confidentially. Labquality will not issue any statements to third parties of the performance of laboratories in external quality assessment schemes unless otherwise agreed.

| Methodics                               | $x_{pt}$ | Median | sd   | CV%  | SEM  | min   | max   | Outliers | n |
|-----------------------------------------|----------|--------|------|------|------|-------|-------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -    | -    | 24.00 | 24.00 | -        | 1 |
| BS-120 (Mindray)                        | 11.50    | 11.50  | 0.71 | 6.1  | 0.50 | 11.00 | 12.00 | -        | 2 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -    | -    | 23.00 | 23.00 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 20.00    | 20.00  | 1.00 | 5.0  | 0.58 | 19.00 | 21.00 | -        | 3 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -    | -    | 24.00 | 24.00 | -        | 1 |
| All                                     | 19.25    | 20.50  | 5.12 | 26.6 | 1.81 | 11.00 | 24.00 | -        | 8 |

Specimen S001 | Aspartate aminotransferase, U/l| histogram summaries in LabScala

Specimen S001 | Creatine kinase, U/l

| Methodics                               | $x_{pt}$ | Median | sd    | CV% | SEM  | min    | max    | Outliers | n |
|-----------------------------------------|----------|--------|-------|-----|------|--------|--------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -     | -   | -    | 230.00 | 230.00 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -     | -   | -    | 265.00 | 265.00 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 261.00   | 259.00 | 7.21  | 2.8 | 4.16 | 255.00 | 269.00 | -        | 3 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -     | -   | -    | 277.00 | 277.00 | -        | 1 |
| All                                     | 259.17   | 262.00 | 16.23 | 6.3 | 6.63 | 230.00 | 277.00 | -        | 6 |

Specimen S001 | Creatine kinase, U/l| histogram summaries in LabScala

Specimen S001 | Thyroxine, nmol/l

| Methodics            | $x_{pt}$ | Median | sd   | CV%  | SEM  | min   | max   | Outliers | n |
|----------------------|----------|--------|------|------|------|-------|-------|----------|---|
| Advia 1800 (Siemens) | -        | -      | -    | -    | -    | 15.30 | 15.30 | -        | 1 |
| BS-120 (Mindray)     | -        | -      | -    | -    | -    | 19.62 | 19.62 | -        | 1 |
| Vetscan VS2 (Abaxis) | -        | -      | -    | -    | -    | 26.00 | 26.00 | -        | 1 |
| All                  | 20.31    | 19.62  | 5.38 | 26.5 | 3.11 | 15.30 | 26.00 | -        | 3 |

Specimen S001 | Thyroxine, nmol/l| histogram summaries in LabScala

Specimen S001 | Cholesterol, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV% | SEM  | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|------|-----|------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -   | -    | 2.37 | 2.37 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -    | -   | -    | 2.54 | 2.54 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -   | -    | 2.42 | 2.42 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 2.45     | 2.45   | 0.07 | 2.9 | 0.05 | 2.40 | 2.50 | -        | 2 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -    | -   | -    | 2.40 | 2.40 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -   | -    | 2.50 | 2.50 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -    | -   | -    | 2.30 | 2.30 | -        | 1 |
| All                                     | 2.43     | 2.41   | 0.08 | 3.3 | 0.03 | 2.30 | 2.54 | -        | 8 |

Specimen S001 | Cholesterol, mmol/l| histogram summaries in LabScala

Specimen S001 | Magnesium, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV% | SEM   | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|------|-----|-------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -   | -     | 0.90 | 0.90 | -        | 1 |
| BS-120 (Mindray)                        | 0.93     | 0.93   | 0.03 | 3.2 | 0.02  | 0.91 | 0.95 | -        | 2 |
| Indiko (Thermo Fischer Scientific)      | 0.90     | 0.90   | 0.02 | 2.4 | 0.02  | 0.88 | 0.91 | -        | 2 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -   | -     | 0.91 | 0.91 | -        | 1 |
| All                                     | 0.91     | 0.91   | 0.02 | 2.5 | <0.01 | 0.88 | 0.95 | -        | 6 |

Specimen S001 | Magnesium, mmol/l| histogram summaries in LabScala

Specimen S001 | Triglycerides, mmol/l

| Methodics             | $x_{pt}$ | Median | sd | CV% | SEM | min  | max  | Outliers | n |
|-----------------------|----------|--------|----|-----|-----|------|------|----------|---|
| Advia 1800 (Siemens)  | -        | -      | -  | -   | -   | 0.45 | 0.45 | -        | 1 |
| BS-120 (Mindray)      | -        | -      | -  | -   | -   | 0.47 | 0.47 | -        | 1 |
| Fuji NX500 (Fujifilm) | -        | -      | -  | -   | -   | 0.34 | 0.34 | -        | 1 |

Copyright © Labquality Oy  
Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copyright. Labquality prohibits any use of its name, or reference to Labquality EQA program, or material in this report in any advertising, brochures or other commercial publications. Labquality EQA data do not necessarily indicate the superiority of instruments, reagents, testing equipments or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of equipments or materials may be deceptive and misleading. Proficiency test results are handled confidentially. Labquality will not issue any statements to third parties of the performance of laboratories in external quality assessment schemes unless otherwise agreed.

|                                         |      |      |      |      |      |      |      |   |   |
|-----------------------------------------|------|------|------|------|------|------|------|---|---|
| Indiko (Thermo Fischer Scientific)      | 0.36 | 0.36 | 0.01 | 3.9  | 0.01 | 0.35 | 0.37 | - | 2 |
| Konelab 60i (Thermo Fischer Scientific) | -    | -    | -    | -    | -    | 0.37 | 0.37 | - | 1 |
| All                                     | 0.39 | 0.37 | 0.05 | 13.9 | 0.02 | 0.34 | 0.47 | - | 6 |

Specimen S001 | Triglycerides, mmol/l| histogram summaries in LabScala

Specimen S001 | Chloride, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd  | CV% | SEM | min   | max   | Outliers | n |
|-----------------------------------------|----------|--------|-----|-----|-----|-------|-------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -   | -   | -   | 117.2 | 117.2 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -   | -   | -   | 108.0 | 108.0 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -   | -   | -   | 111.0 | 111.0 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 115.6    | 115.6  | 0.6 | 0.6 | 0.5 | 115.1 | 116.0 | -        | 2 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -   | -   | -   | 110.0 | 110.0 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -   | -   | -   | 118.6 | 118.6 | -        | 1 |
| All                                     | 113.7    | 115.1  | 4.0 | 3.5 | 1.5 | 108.0 | 118.6 | -        | 7 |

Specimen S001 | Chloride, mmol/l| histogram summaries in LabScala

Specimen S001 | Osmol, mosm/kg

| Methodics            | $x_{pt}$ | Median | sd | CV% | SEM | min    | max    | Outliers | n |
|----------------------|----------|--------|----|-----|-----|--------|--------|----------|---|
| Advia 1800 (Siemens) | -        | -      | -  | -   | -   | 303.00 | 303.00 | -        | 1 |
| All                  | -        | -      | -  | -   | -   | 303.00 | 303.00 | -        | 1 |

Specimen S001 | Osmol, mosm/kg| histogram summaries in LabScala

Specimen S001 | Calcium, mmol/l

| Methodics                          | $x_{pt}$ | Median | sd | CV% | SEM | min  | max  | Outliers | n |
|------------------------------------|----------|--------|----|-----|-----|------|------|----------|---|
| Indiko (Thermo Fischer Scientific) | -        | -      | -  | -   | -   | 2.04 | 2.04 | -        | 1 |
| All                                | -        | -      | -  | -   | -   | 2.04 | 2.04 | -        | 1 |

Specimen S001 | Calcium, mmol/l| histogram summaries in LabScala

Specimen S001 | Iron, µmol/l

| Methodics                          | $x_{pt}$ | Median | sd | CV% | SEM | min   | max   | Outliers | n |
|------------------------------------|----------|--------|----|-----|-----|-------|-------|----------|---|
| Indiko (Thermo Fischer Scientific) | -        | -      | -  | -   | -   | 13.10 | 13.10 | -        | 1 |
| All                                | -        | -      | -  | -   | -   | 13.10 | 13.10 | -        | 1 |

Specimen S001 | Iron, µmol/l| histogram summaries in LabScala

Specimen S001 | Lactate dehydrogenase, U/l

| Methodics                          | $x_{pt}$ | Median | sd | CV% | SEM | min   | max   | Outliers | n |
|------------------------------------|----------|--------|----|-----|-----|-------|-------|----------|---|
| Indiko (Thermo Fischer Scientific) | -        | -      | -  | -   | -   | 53.00 | 53.00 | -        | 1 |
| All                                | -        | -      | -  | -   | -   | 53.00 | 53.00 | -        | 1 |

Specimen S001 | Lactate dehydrogenase, U/l| histogram summaries in LabScala

Specimen S001 | Uric acid, µmol/l

| Methodics                          | $x_{pt}$ | Median | sd | CV% | SEM | min   | max   | Outliers | n |
|------------------------------------|----------|--------|----|-----|-----|-------|-------|----------|---|
| Indiko (Thermo Fischer Scientific) | -        | -      | -  | -   | -   | 18.00 | 18.00 | -        | 1 |
| All                                | -        | -      | -  | -   | -   | 18.00 | 18.00 | -        | 1 |

Specimen S001 | Uric acid, µmol/l| histogram summaries in LabScala

Specimen S002 | Alkaline phosphatase, U/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV%  | SEM  | min   | max   | Outliers | n |
|-----------------------------------------|----------|--------|------|------|------|-------|-------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -    | -    | 37.00 | 37.00 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -    | -    | -    | 26.00 | 26.00 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -    | -    | 39.00 | 39.00 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 41.00    | 40.00  | 3.61 | 8.8  | 2.08 | 38.00 | 45.00 | -        | 3 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -    | -    | -    | 47.00 | 47.00 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -    | -    | 38.00 | 38.00 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -    | -    | -    | 38.00 | 38.00 | -        | 1 |
| All                                     | 38.67    | 38.00  | 5.87 | 15.2 | 1.96 | 26.00 | 47.00 | -        | 9 |

Specimen S002 | Alkaline phosphatase, U/l | histogram summaries in LabScala

Specimen S002 | Alanine aminotransferase, U/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV% | SEM  | min   | max   | Outliers | n |
|-----------------------------------------|----------|--------|------|-----|------|-------|-------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -   | -    | 35.00 | 35.00 | -        | 1 |
| BS-120 (Mindray)                        | 34.50    | 34.50  | 0.71 | 2.0 | 0.50 | 34.00 | 35.00 | -        | 2 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -   | -    | 43.00 | 43.00 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 37.00    | 37.00  | 1.00 | 2.7 | 0.58 | 36.00 | 38.00 | -        | 3 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -   | -    | 39.00 | 39.00 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -    | -   | -    | 36.00 | 36.00 | -        | 1 |
| All                                     | 37.00    | 36.00  | 2.74 | 7.4 | 0.91 | 34.00 | 43.00 | -        | 9 |

Specimen S002 | Alanine aminotransferase, U/l | histogram summaries in LabScala

Specimen S002 | Glutamyl transferase, U/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV%  | SEM  | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|------|------|------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -    | -    | 4.00 | 4.00 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -    | -    | -    | 3.00 | 3.00 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 4.00     | 4.00   | 1.41 | 35.4 | 1.00 | 3.00 | 5.00 | -        | 2 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -    | -    | 3.00 | 3.00 | -        | 1 |
| All                                     | 3.60     | 3.00   | 0.89 | 24.8 | 0.40 | 3.00 | 5.00 | -        | 5 |

Specimen S002 | Glutamyl transferase, U/l | histogram summaries in LabScala

Specimen S002 | Amylase, U/l

| Methodics                          | $x_{pt}$ | Median | sd     | CV%  | SEM   | min    | max    | Outliers | n |
|------------------------------------|----------|--------|--------|------|-------|--------|--------|----------|---|
| Advia 1800 (Siemens)               | -        | -      | -      | -    | -     | 620.00 | 620.00 | -        | 1 |
| Indiko (Thermo Fischer Scientific) | -        | -      | -      | -    | -     | 693.00 | 693.00 | -        | 1 |
| Vetscan VS2 (Abaxis)               | -        | -      | -      | -    | -     | 491.00 | 491.00 | -        | 1 |
| All                                | 601.33   | 620.00 | 102.29 | 17.0 | 59.05 | 491.00 | 693.00 | -        | 3 |

Specimen S002 | Amylase, U/l | histogram summaries in LabScala

Specimen S002 | Albumin, g/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV% | SEM  | min   | max   | Outliers | n  |
|-----------------------------------------|----------|--------|------|-----|------|-------|-------|----------|----|
| Advia 1800 (Siemens)                    | -        | -      | -    | -   | -    | 30.70 | 30.70 | -        | 1  |
| BS-120 (Mindray)                        | 32.05    | 32.05  | 1.20 | 3.8 | 0.85 | 31.20 | 32.90 | -        | 2  |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -   | -    | 30.00 | 30.00 | -        | 1  |
| Indiko (Thermo Fischer Scientific)      | 28.23    | 28.00  | 0.40 | 1.4 | 0.23 | 28.00 | 28.70 | -        | 3  |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -    | -   | -    | 29.00 | 29.00 | -        | 1  |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -   | -    | 28.90 | 28.90 | -        | 1  |
| Vetscan VS2 (Abaxis)                    | -        | -      | -    | -   | -    | 58.00 | 58.00 | -        | 1  |
| All                                     | 29.71    | 29.00  | 1.64 | 5.5 | 0.55 | 28.00 | 32.90 | 1        | 10 |

Specimen S002 | Albumin, g/l | histogram summaries in LabScala

Specimen S002 | Protein, g/l

| Methodics | $x_{pt}$ | Median | sd | CV% | SEM | min | max | Outliers | n |
|-----------|----------|--------|----|-----|-----|-----|-----|----------|---|
|-----------|----------|--------|----|-----|-----|-----|-----|----------|---|

Copyright © Labquality Oy  
Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copyright. Labquality prohibits any use of its name, or reference to Labquality EQA program, or material in this report in any advertising, brochures or other commercial publications. Labquality EQA data do not necessarily indicate the superiority of instruments, reagents, testing equipments or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of equipments or materials may be deceptive and misleading. Proficiency test results are handled confidentially. Labquality will not issue any statements to third parties of the performance of laboratories in external quality assessment schemes unless otherwise agreed.

|                                         |       |       |      |     |      |       |       |   |    |
|-----------------------------------------|-------|-------|------|-----|------|-------|-------|---|----|
| Advia 1800 (Siemens)                    | -     | -     | -    | -   | -    | 59.90 | 59.90 | - | 1  |
| BS-120 (Mindray)                        | 61.40 | 61.40 | 3.25 | 5.3 | 2.30 | 59.10 | 63.70 | - | 2  |
| Fuji NX500 (Fujifilm)                   | -     | -     | -    | -   | -    | 57.00 | 57.00 | - | 1  |
| Indiko (Thermo Fischer Scientific)      | 58.33 | 58.00 | 0.58 | 1.0 | 0.33 | 58.00 | 59.00 | - | 3  |
| Konelab 30 (Thermo Fischer Scientific)  | -     | -     | -    | -   | -    | 56.00 | 56.00 | - | 1  |
| Konelab 60i (Thermo Fischer Scientific) | -     | -     | -    | -   | -    | 57.00 | 57.00 | - | 1  |
| Vetscan VS2 (Abaxis)                    | -     | -     | -    | -   | -    | 61.00 | 61.00 | - | 1  |
| All                                     | 58.87 | 58.50 | 2.25 | 3.8 | 0.71 | 56.00 | 63.70 | - | 10 |

Specimen S002 | Protein, g/l| histogram summaries in LabScala

Specimen S002 | Bilirubin total, µmol/l

| Methodics                               | $x_{pt}$ | Median | sd  | CV%  | SEM | min | max | Outliers | n |
|-----------------------------------------|----------|--------|-----|------|-----|-----|-----|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -   | -    | -   | 0.3 | 0.3 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -   | -    | -   | 1.0 | 1.0 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -   | -    | -   | 4.0 | 4.0 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 1.7      | 1.7    | 0.5 | 30.0 | 0.4 | 1.3 | 2.0 | -        | 2 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -   | -    | -   | 2.1 | 2.1 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -   | -    | -   | 2.2 | 2.2 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -   | -    | -   | 4.0 | 4.0 | -        | 1 |
| All                                     | 2.1      | 2.1    | 1.3 | 62.6 | 0.5 | 0.3 | 4.0 | -        | 8 |

Specimen S002 | Bilirubin total, µmol/l| histogram summaries in LabScala

Specimen S002 | Urea, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV% | SEM  | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|------|-----|------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -   | -    | 5.50 | 5.50 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -    | -   | -    | 5.66 | 5.66 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -   | -    | 5.41 | 5.41 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 5.20     | 5.30   | 0.46 | 8.8 | 0.26 | 4.70 | 5.60 | -        | 3 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -    | -   | -    | 4.60 | 4.60 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -   | -    | 5.27 | 5.27 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -    | -   | -    | 4.70 | 4.70 | -        | 1 |
| All                                     | 5.19     | 5.30   | 0.42 | 8.0 | 0.14 | 4.60 | 5.66 | -        | 9 |

Specimen S002 | Urea, mmol/l| histogram summaries in LabScala

Specimen S002 | Creatinine, µmol/l

| Methodics                               | $x_{pt}$ | Median | sd  | CV%  | SEM | min  | max  | Outliers | n  |
|-----------------------------------------|----------|--------|-----|------|-----|------|------|----------|----|
| Advia 1800 (Siemens)                    | -        | -      | -   | -    | -   | 68.0 | 68.0 | -        | 1  |
| BS-120 (Mindray)                        | 48.6     | 48.6   | 7.5 | 15.4 | 5.3 | 43.3 | 53.9 | -        | 2  |
| Fuji NX500 (Fujifilm)                   | -        | -      | -   | -    | -   | 51.0 | 51.0 | -        | 1  |
| Indiko (Thermo Fischer Scientific)      | 57.0     | 57.0   | 1.0 | 1.8  | 0.6 | 56.0 | 58.0 | -        | 3  |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -   | -    | -   | 58.0 | 58.0 | -        | 1  |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -   | -    | -   | 58.0 | 58.0 | -        | 1  |
| Vetscan VS2 (Abaxis)                    | -        | -      | -   | -    | -   | 65.0 | 65.0 | -        | 1  |
| All                                     | 56.8     | 57.5   | 6.8 | 12.0 | 2.2 | 43.3 | 68.0 | -        | 10 |

Specimen S002 | Creatinine, µmol/l| histogram summaries in LabScala

Specimen S002 | Calcium, ionized, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV% | SEM  | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|------|-----|------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -   | -    | 2.48 | 2.48 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -   | -    | 2.73 | 2.73 | -        | 1 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -    | -   | -    | 2.50 | 2.50 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -   | -    | 2.70 | 2.70 | -        | 1 |
| All                                     | 2.60     | 2.60   | 0.13 | 5.0 | 0.07 | 2.48 | 2.73 | -        | 4 |

Specimen S002 | Calcium, ionized, mmol/l| histogram summaries in LabScala

Specimen S002 | Glucose, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd    | CV%  | SEM   | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|-------|------|-------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -     | -    | -     | 3.20 | 3.20 | -        | 1 |
| BS-120 (Mindray)                        | 3.55     | 3.55   | <0.01 | <0.1 | <0.01 | 3.55 | 3.55 | -        | 2 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -     | -    | -     | 3.20 | 3.20 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 3.33     | 3.40   | 0.12  | 3.5  | 0.07  | 3.20 | 3.40 | -        | 3 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -     | -    | -     | 3.40 | 3.40 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -     | -    | -     | 3.30 | 3.30 | -        | 1 |
| All                                     | 3.36     | 3.40   | 0.14  | 4.2  | 0.05  | 3.20 | 3.55 | -        | 9 |

Specimen S002 | Glucose, mmol/l| histogram summaries in LabScala

Specimen S002 | Sodium, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd  | CV% | SEM | min   | max   | Outliers | n |
|-----------------------------------------|----------|--------|-----|-----|-----|-------|-------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -   | -   | -   | 145.3 | 145.3 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -   | -   | -   | 148.0 | 148.0 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -   | -   | -   | 143.0 | 143.0 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 148.7    | 147.0  | 2.9 | 1.9 | 1.7 | 147.0 | 152.0 | -        | 3 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -   | -   | -   | 148.0 | 148.0 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -   | -   | -   | 146.0 | 146.0 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -   | -   | -   | 138.0 | 138.0 | -        | 1 |
| All                                     | 146.0    | 147.0  | 3.9 | 2.6 | 1.3 | 138.0 | 152.0 | -        | 9 |

Specimen S002 | Sodium, mmol/l| histogram summaries in LabScala

Specimen S002 | Potassium, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV% | SEM  | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|------|-----|------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -   | -    | 4.56 | 4.56 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -    | -   | -    | 4.70 | 4.70 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -   | -    | 4.30 | 4.30 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 4.53     | 4.50   | 0.06 | 1.3 | 0.03 | 4.50 | 4.60 | -        | 3 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -   | -    | 4.52 | 4.52 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -    | -   | -    | 4.90 | 4.90 | -        | 1 |
| All                                     | 4.57     | 4.54   | 0.17 | 3.8 | 0.06 | 4.30 | 4.90 | -        | 8 |

Specimen S002 | Potassium, mmol/l| histogram summaries in LabScala

Specimen S002 | Phosphorus, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV% | SEM  | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|------|-----|------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -   | -    | 1.46 | 1.46 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -    | -   | -    | 1.39 | 1.39 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -   | -    | 1.47 | 1.47 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 1.42     | 1.40   | 0.08 | 5.8 | 0.05 | 1.35 | 1.51 | -        | 3 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -    | -   | -    | 1.34 | 1.34 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -   | -    | 1.44 | 1.44 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -    | -   | -    | 1.56 | 1.56 | -        | 1 |
| All                                     | 1.44     | 1.44   | 0.07 | 5.1 | 0.02 | 1.34 | 1.56 | -        | 9 |

Specimen S002 | Phosphorus, mmol/l| histogram summaries in LabScala

Specimen S002 | GLOB, g/l

| Methodics             | $x_{pt}$ | Median | sd | CV%  | SEM | min | max | Outliers | n |
|-----------------------|----------|--------|----|------|-----|-----|-----|----------|---|
| Advia 1800 (Siemens)  | -        | -      | -  | -    | -   | 29  | 29  | -        | 1 |
| Fuji NX500 (Fujifilm) | -        | -      | -  | -    | -   | 27  | 27  | -        | 1 |
| Vetscan VS2 (Abaxis)  | -        | -      | -  | -    | -   | 3   | 3   | -        | 1 |
| All                   | 20       | 27     | 15 | 73.6 | 8   | 3   | 29  | -        | 3 |

Specimen S002 | GLOB, g/l| histogram summaries in LabScala

Copyright © Labquality Oy  
Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copyright. Labquality prohibits any use of its name, or reference to Labquality EQA program, or material in this report in any advertising, brochures or other commercial publications. Labquality EQA data do not necessarily indicate the superiority of instruments, reagents, testing equipments or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of equipments or materials may be deceptive and misleading. Proficiency test results are handled confidentially. Labquality will not issue any statements to third parties of the performance of laboratories in external quality assessment schemes unless otherwise agreed.

Specimen S002 | Aspartate aminotransferase, U/l

| Methodics                               | $x_{pt}$ | Median | sd    | CV%  | SEM   | min   | max   | Outliers | n |
|-----------------------------------------|----------|--------|-------|------|-------|-------|-------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -     | -    | -     | 39.00 | 39.00 | -        | 1 |
| BS-120 (Mindray)                        | 34.00    | 34.00  | <0.01 | <0.1 | <0.01 | 34.00 | 34.00 | -        | 2 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -     | -    | -     | 37.00 | 37.00 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 41.33    | 41.00  | 2.52  | 6.1  | 1.45  | 39.00 | 44.00 | -        | 3 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -     | -    | -     | 46.00 | 46.00 | -        | 1 |
| All                                     | 39.25    | 39.00  | 4.33  | 11.0 | 1.53  | 34.00 | 46.00 | -        | 8 |

Specimen S002 | Aspartate aminotransferase, U/l| histogram summaries in LabScala

Specimen S002 | Creatine kinase, U/l

| Methodics                               | $x_{pt}$ | Median | sd    | CV% | SEM  | min    | max    | Outliers | n |
|-----------------------------------------|----------|--------|-------|-----|------|--------|--------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -     | -   | -    | 232.00 | 232.00 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -     | -   | -    | 281.00 | 281.00 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 272.00   | 271.00 | 11.53 | 4.2 | 6.66 | 261.00 | 284.00 | -        | 3 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -     | -   | -    | 289.00 | 289.00 | -        | 1 |
| All                                     | 269.67   | 276.00 | 20.99 | 7.8 | 8.57 | 232.00 | 289.00 | -        | 6 |

Specimen S002 | Creatine kinase, U/l| histogram summaries in LabScala

Specimen S002 | Thyroxine, nmol/l

| Methodics            | $x_{pt}$ | Median | sd   | CV% | SEM  | min   | max   | Outliers | n |
|----------------------|----------|--------|------|-----|------|-------|-------|----------|---|
| Advia 1800 (Siemens) | -        | -      | -    | -   | -    | 19.50 | 19.50 | -        | 1 |
| BS-120 (Mindray)     | -        | -      | -    | -   | -    | 19.65 | 19.65 | -        | 1 |
| Vetscan VS2 (Abaxis) | -        | -      | -    | -   | -    | 19.00 | 19.00 | -        | 1 |
| All                  | 19.38    | 19.50  | 0.34 | 1.8 | 0.20 | 19.00 | 19.65 | -        | 3 |

Specimen S002 | Thyroxine, nmol/l| histogram summaries in LabScala

Specimen S002 | Cholesterol, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV% | SEM  | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|------|-----|------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -   | -    | 4.59 | 4.59 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -    | -   | -    | 4.58 | 4.58 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -   | -    | 5.14 | 5.14 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 4.85     | 4.85   | 0.07 | 1.5 | 0.05 | 4.80 | 4.90 | -        | 2 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -    | -   | -    | 4.80 | 4.80 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -   | -    | 4.80 | 4.80 | -        | 1 |
| Vetscan VS2 (Abaxis)                    | -        | -      | -    | -   | -    | 4.60 | 4.60 | -        | 1 |
| All                                     | 4.78     | 4.80   | 0.19 | 4.0 | 0.07 | 4.58 | 5.14 | -        | 8 |

Specimen S002 | Cholesterol, mmol/l| histogram summaries in LabScala

Specimen S002 | Magnesium, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd    | CV%  | SEM   | min  | max  | Outliers | n |
|-----------------------------------------|----------|--------|-------|------|-------|------|------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -     | -    | -     | 0.81 | 0.81 | -        | 1 |
| BS-120 (Mindray)                        | 1.20     | 1.20   | 0.52  | 43.2 | 0.37  | 0.83 | 1.56 | -        | 2 |
| Indiko (Thermo Fischer Scientific)      | 0.81     | 0.81   | <0.01 | 0.9  | <0.01 | 0.80 | 0.81 | -        | 2 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -     | -    | -     | 0.87 | 0.87 | -        | 1 |
| All                                     | 0.95     | 0.82   | 0.30  | 31.8 | 0.12  | 0.80 | 1.56 | -        | 6 |

Specimen S002 | Magnesium, mmol/l| histogram summaries in LabScala

Specimen S002 | Triglycerides, mmol/l

| Methodics            | $x_{pt}$ | Median | sd | CV% | SEM | min  | max  | Outliers | n |
|----------------------|----------|--------|----|-----|-----|------|------|----------|---|
| Advia 1800 (Siemens) | -        | -      | -  | -   | -   | 0.49 | 0.49 | -        | 1 |

|                                         |      |      |      |     |      |      |      |   |   |
|-----------------------------------------|------|------|------|-----|------|------|------|---|---|
| BS-120 (Mindray)                        | -    | -    | -    | -   | -    | 0.44 | 0.44 | - | 1 |
| Fuji NX500 (Fujifilm)                   | -    | -    | -    | -   | -    | 0.38 | 0.38 | - | 1 |
| Indiko (Thermo Fischer Scientific)      | 0.43 | 0.43 | 0.01 | 3.3 | 0.01 | 0.42 | 0.44 | - | 2 |
| Konelab 60i (Thermo Fischer Scientific) | -    | -    | -    | -   | -    | 0.46 | 0.46 | - | 1 |
| All                                     | 0.44 | 0.44 | 0.04 | 8.5 | 0.02 | 0.38 | 0.49 | - | 6 |

Specimen S002 | Triglycerides, mmol/l| histogram summaries in LabScala

Specimen S002 | Chloride, mmol/l

| Methodics                               | $x_{pt}$ | Median | sd   | CV%  | SEM  | min   | max   | Outliers | n |
|-----------------------------------------|----------|--------|------|------|------|-------|-------|----------|---|
| Advia 1800 (Siemens)                    | -        | -      | -    | -    | -    | 112.7 | 112.7 | -        | 1 |
| BS-120 (Mindray)                        | -        | -      | -    | -    | -    | 104.0 | 104.0 | -        | 1 |
| Fuji NX500 (Fujifilm)                   | -        | -      | -    | -    | -    | 102.0 | 102.0 | -        | 1 |
| Indiko (Thermo Fischer Scientific)      | 112.0    | 112.0  | <0.1 | <0.1 | <0.1 | 112.0 | 112.0 | -        | 2 |
| Konelab 30 (Thermo Fischer Scientific)  | -        | -      | -    | -    | -    | 107.0 | 107.0 | -        | 1 |
| Konelab 60i (Thermo Fischer Scientific) | -        | -      | -    | -    | -    | 113.7 | 113.7 | -        | 1 |
| All                                     | 109.1    | 112.0  | 4.7  | 4.3  | 1.8  | 102.0 | 113.7 | -        | 7 |

Specimen S002 | Chloride, mmol/l| histogram summaries in LabScala

Specimen S002 | Osmol, mosm/kg

| Methodics            | $x_{pt}$ | Median | sd | CV% | SEM | min    | max    | Outliers | n |
|----------------------|----------|--------|----|-----|-----|--------|--------|----------|---|
| Advia 1800 (Siemens) | -        | -      | -  | -   | -   | 304.00 | 304.00 | -        | 1 |
| All                  | -        | -      | -  | -   | -   | 304.00 | 304.00 | -        | 1 |

Specimen S002 | Osmol, mosm/kg| histogram summaries in LabScala

Specimen S002 | Calcium, mmol/l

| Methodics                          | $x_{pt}$ | Median | sd | CV% | SEM | min  | max  | Outliers | n |
|------------------------------------|----------|--------|----|-----|-----|------|------|----------|---|
| Indiko (Thermo Fischer Scientific) | -        | -      | -  | -   | -   | 2.63 | 2.63 | -        | 1 |
| All                                | -        | -      | -  | -   | -   | 2.63 | 2.63 | -        | 1 |

Specimen S002 | Calcium, mmol/l| histogram summaries in LabScala

Specimen S002 | Iron, µmol/l

| Methodics                          | $x_{pt}$ | Median | sd | CV% | SEM | min   | max   | Outliers | n |
|------------------------------------|----------|--------|----|-----|-----|-------|-------|----------|---|
| Indiko (Thermo Fischer Scientific) | -        | -      | -  | -   | -   | 23.50 | 23.50 | -        | 1 |
| All                                | -        | -      | -  | -   | -   | 23.50 | 23.50 | -        | 1 |

Specimen S002 | Iron, µmol/l| histogram summaries in LabScala

Specimen S002 | Lactate dehydrogenase, U/l

| Methodics                          | $x_{pt}$ | Median | sd | CV% | SEM | min    | max    | Outliers | n |
|------------------------------------|----------|--------|----|-----|-----|--------|--------|----------|---|
| Indiko (Thermo Fischer Scientific) | -        | -      | -  | -   | -   | 289.00 | 289.00 | -        | 1 |
| All                                | -        | -      | -  | -   | -   | 289.00 | 289.00 | -        | 1 |

Specimen S002 | Lactate dehydrogenase, U/l| histogram summaries in LabScala

Specimen S002 | Lipase, U/l

| Methodics                          | $x_{pt}$ | Median | sd | CV% | SEM | min   | max   | Outliers | n |
|------------------------------------|----------|--------|----|-----|-----|-------|-------|----------|---|
| Indiko (Thermo Fischer Scientific) | -        | -      | -  | -   | -   | 59.00 | 59.00 | -        | 1 |
| All                                | -        | -      | -  | -   | -   | 59.00 | 59.00 | -        | 1 |

Specimen S002 | Lipase, U/l| histogram summaries in LabScala



Specimen S002 | Uric acid, µmol/l

Copyright © Labquality Oy  
Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copyright. Labquality prohibits any use of its name, or reference to Labquality EQA program, or material in this report in any advertising, brochures or other commercial publications. Labquality EQA data do not necessarily indicate the superiority of instruments, reagents, testing equipments or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of equipments or materials may be deceptive and misleading. Proficiency test results are handled confidentially. Labquality will not issue any statements to third parties of the performance of laboratories in external quality assessment schemes unless otherwise agreed.

| Methodics                          | $x_{pt}$ | Median | sd | CV% | SEM | min   | max   | Outliers | n |
|------------------------------------|----------|--------|----|-----|-----|-------|-------|----------|---|
| Indiko (Thermo Fischer Scientific) | -        | -      | -  | -   | -   | 36.00 | 36.00 | -        | 1 |
| All                                | -        | -      | -  | -   | -   | 36.00 | 36.00 | -        | 1 |

Specimen S002 | Uric acid, µmol/l| histogram summaries in LabScala



Alanine aminotransferase |1



|                  | $x_{pt}$ | sd    | SEM   | CV%  | n |
|------------------|----------|-------|-------|------|---|
| BS-120 (Mindray) | 3.00 U/l | <0.01 | <0.01 | <0.1 | 2 |
| All methods      | 3.43 U/l | 1.40  | 0.53  | 40.8 | 7 |

|                  | $x_{pt}$  | sd   | SEM  | CV% | n |
|------------------|-----------|------|------|-----|---|
| BS-120 (Mindray) | 34.50 U/l | 0.71 | 0.50 | 2.0 | 2 |
| All methods      | 37.00 U/l | 2.74 | 0.91 | 7.4 | 9 |



Albumin |1



|                  | $x_{pt}$  | sd   | SEM  | CV%  | n  |
|------------------|-----------|------|------|------|----|
| BS-120 (Mindray) | 29.45 g/l | 1.34 | 0.95 | 4.6  | 2  |
| All methods      | 25.17 g/l | 2.97 | 0.94 | 11.8 | 10 |

|                  | $x_{pt}$  | sd   | SEM  | CV% | n  |
|------------------|-----------|------|------|-----|----|
| BS-120 (Mindray) | 32.05 g/l | 1.20 | 0.85 | 3.8 | 2  |
| All methods      | 29.71 g/l | 1.64 | 0.55 | 5.5 | 10 |



Protein |1



|                  | <i>x<sub>pt</sub></i> | sd   | SEM  | CV% | n  |
|------------------|-----------------------|------|------|-----|----|
| BS-120 (Mindray) | 61.95 g/l             | 1.77 | 1.25 | 2.9 | 2  |
| All methods      | 59.86 g/l             | 1.84 | 0.58 | 3.1 | 10 |

|                  | <i>x<sub>pt</sub></i> | sd   | SEM  | CV% | n  |
|------------------|-----------------------|------|------|-----|----|
| BS-120 (Mindray) | 61.40 g/l             | 3.25 | 2.30 | 5.3 | 2  |
| All methods      | 58.87 g/l             | 2.25 | 0.71 | 3.8 | 10 |

Urea |1



# - manually removed


# - manually removed

Coordinator: Please check your unit. J.Pelanti/Labquality

|                  | <i>x<sub>pt</sub></i> | sd   | SEM  | CV% | n |
|------------------|-----------------------|------|------|-----|---|
| BS-120 (Mindray) | -                     | -    | -    | -   | 1 |
| All methods      | 8.14 mmol/l           | 0.65 | 0.22 | 8.0 | 9 |

|                  | <i>x<sub>pt</sub></i> | sd   | SEM  | CV% | n |
|------------------|-----------------------|------|------|-----|---|
| BS-120 (Mindray) | -                     | -    | -    | -   | 1 |
| All methods      | 5.19 mmol/l           | 0.42 | 0.14 | 8.0 | 9 |


Creatinine |1



|                  | $x_{pt}$     | sd   | SEM | CV%  | n  |
|------------------|--------------|------|-----|------|----|
| BS-120 (Mindray) | 93.7 µmol/l  | 8.7  | 6.2 | 9.3  | 2  |
| All methods      | 101.0 µmol/l | 10.2 | 3.2 | 10.0 | 10 |

|                  | $x_{pt}$    | sd  | SEM | CV%  | n  |
|------------------|-------------|-----|-----|------|----|
| BS-120 (Mindray) | 48.6 µmol/l | 7.5 | 5.3 | 15.4 | 2  |
| All methods      | 56.8 µmol/l | 6.8 | 2.2 | 12.0 | 10 |



Glucose |1



|                  | $x_{pt}$    | sd   | SEM  | CV% | n |
|------------------|-------------|------|------|-----|---|
| BS-120 (Mindray) | 3.31 mmol/l | 0.04 | 0.03 | 1.1 | 2 |
| All methods      | 3.07 mmol/l | 0.15 | 0.05 | 5.0 | 9 |

|                  | $x_{pt}$    | sd    | SEM   | CV%  | n |
|------------------|-------------|-------|-------|------|---|
| BS-120 (Mindray) | 3.55 mmol/l | <0.01 | <0.01 | <0.1 | 2 |
| All methods      | 3.36 mmol/l | 0.14  | 0.05  | 4.2  | 9 |



Aspartate aminotransferase |1



|                  | $x_{pt}$  | sd   | SEM  | CV%  | n |
|------------------|-----------|------|------|------|---|
| BS-120 (Mindray) | 11.50 U/l | 0.71 | 0.50 | 6.1  | 2 |
| All methods      | 19.25 U/l | 5.12 | 1.81 | 26.6 | 8 |

|                  | $x_{pt}$  | sd    | SEM   | CV%  | n |
|------------------|-----------|-------|-------|------|---|
| BS-120 (Mindray) | 34.00 U/l | <0.01 | <0.01 | <0.1 | 2 |
| All methods      | 39.25 U/l | 4.33  | 1.53  | 11.0 | 8 |

Magnesium |1



|                  | $x_{pt}$    | sd   | SEM   | CV% | n |
|------------------|-------------|------|-------|-----|---|
| BS-120 (Mindray) | 0.93 mmol/l | 0.03 | 0.02  | 3.2 | 2 |
| All methods      | 0.91 mmol/l | 0.02 | <0.01 | 2.5 | 6 |

|                  | $x_{pt}$    | sd   | SEM  | CV%  | n |
|------------------|-------------|------|------|------|---|
| BS-120 (Mindray) | 1.20 mmol/l | 0.52 | 0.37 | 43.2 | 2 |
| All methods      | 0.95 mmol/l | 0.30 | 0.12 | 31.8 | 6 |

## Veterinary basic chemistry, 1-2018

10 laboratories participated in this scheme.

### Material

Samples S001 (LQ780518011) and S002 (LQ780518012) were serum samples. Sample S001 was feline and S002 of canine origin.

The materials were sent without any temperature control packaging.

### Expert comments

In sample S001, the concentrations of many analytes is so low that the differences in the results could not be compared.

The albumin content in sample S002 is at a uniform level, except the Vetscan method. The albumin results from sample S001 varies too much to allow a consistent clinical interpretation. The total protein methods are at a very uniform level.

There are variations in creatinine methods, but some of the methods are based on the enzymatic method and some methods are based on colour reaction. There is clear distinction in these different method types. There is too much variation in the results of sodium, potassium and chloride when compared to the reference value range.

In some analytes (amylase, magnesium), the cat's sample yielded consistent results and in (thyroxine) the canine sample yielded a consistent results.

In the many methods (urea, bilirubin, glucose, creatine kinase, cholesterol, triglycerides, alanine aminotransferase and globulins), the methods give comparable results..

This round provided a good start for laboratories to monitor and check their result levels compared to other laboratories. In the future, it would be good to have pathological samples to compare the results (bilirubin, glutamyltransferase).

### How to interpret the report

Laboratories receive histogram reports in which the result distribution is shown in graphic mode. Laboratory's own result is marked with an orange dot. Black dashed line is the assigned value, the mean value of the accepted results calculated for clients using similar devices. Clear outliers of the result distribution have been removed. The yellow area is the target area which is depending on the test. The aim is for your result to be in the yellow area and if it is then it means that your laboratory's performance is acceptable compared to others using the same device. All participants receive also a numerical summary from which all the methods' performance can be seen.

2018-12-03

### Final Report

Product no. 8530  
LQ780518011-012 (UK)

Samples sent 2018-11-05  
Survey closed 2018-11-25  
Report released 2018-12-03

### The report contains

- Numerical summary
- Individual histograms (if results have been returned)

Inquiries about this survey, including questions of possible errors in result processing, should be at Labquality's office **before 2019-01-03**.

Authorized by  
EQA coordinator  
Jonna Pelanti  
tel. +358 503706110  
jonna.pelanti@labquality.fi

Scheme expert  
Kari Åkerman

**Labquality**  
Kumpulantie 15  
FI-00520 HELSINKI

**Telephone**  
+358 9 8566 8200

**Fax**  
+358 9 8566 8280

info@labquality.fi  
www.labquality.fi

© Labquality



The symbols presented in the reports are

**x<sub>pt</sub>** = Assigned value, method mean

**Med** = median value

**sd** = standard deviation showing the scattering of the results

**CV%** = coefficient of variation (standard deviation divided by the mean value)

**SEM** = Standard error of the mean (standard deviation divided by the square root of the number of results)

**min** = The smallest accepted result in the method group

**max** = The largest accepted result in the method group

**Outliers** = The number of outliers found by the calculation used

**n** = Number of accepted results in the method group

From the beginning of 2018 we have made some changes in the statistical calculations and reporting. In case the client's result is the only one in the method group, no target value will be calculated, no target area shown, and no statistics calculated. In case there are 2-12 results in a method group, the robust calculation is not used but a calculation where results deviating more than  $\pm 3$  \* standard deviation SD from the median are removed. Additionally, if the measurement uncertainty of the target value is too large ( $u(x_{pt}) < 0.1\delta E$ ) an automatic text is printed on the report: "The uncertainty of the assigned value is not negligible, and evaluations could be affected." In case there are 2-5 results in a method group, no z-score is calculated, and a text is printed on the report: "Due to the small number of results, the z score is not calculated." In case there are 6-12 results, the report has a text: "Z score is uncertain due to the small number of observations."

Guidelines how to interpret the reports can be found under "LabScala user instructions" in LabScala.

As there were only a limited amount of participants in this round all the conclusions from the report should be made with caution. In case you wish to have help in interpreting your performance please contact Labquality. We would be happy to be of assistance.

## End of report

Copyright © Labquality Oy

Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copyright. Labquality prohibits any use of its name, or reference to Labquality EQA program, or material in this report in any advertising, brochures or other commercial publications. Labquality EQA data do not necessarily indicate the superiority of instruments, reagents, testing equipment or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of equipment or materials may be deceptive and misleading. Proficiency test results are handled confidentially. Labquality will not issue any statements to third parties of the performance of laboratories in external quality assessment schemes unless otherwise agreed.