

# LABQUALITY

External Quality Assessment Scheme

## Histological staining techniques: HE and Iron (Fe) Round 1, 2023

### Specimens

Please find enclosed two unstained and unbaked slides.

### Background information

Each participating laboratory receives two unstained slides. Each slide is labeled according to the staining procedure. The samples have been fixed in 10% neutral buffered formalin.

S001: HE (hematoxylin & eosin) (small intestine, ileum)

S002: Iron (Fe) (lung and kidney)

### Storage and use

The sections are adhered to the slides for 15 min +60°C.

The slides should be stained with hematoxylin & eosin or iron. The slides can be stored at room temperature before and after staining. If the protocols of your laboratory do not include both stains, you can participate with one slide only. If needed, extra slides can be ordered from Labquality.

### Please remember to label the slides clearly with your Labquality account number

If you ordered multiple sample sets with the same client code, please separate them with the alphabet.

### Returning slides and protocol reporting

Please return the stained slides in the same slide mailer box wrapped in bubble wrap at participant's own cost to the address given in the column on the right side of the page unless our local partner has instructed otherwise. The shipment comes with an address label for return, which does not include postage.

The slides should be at Labquality's office no later than the date given in the column on the right side of the page. Please note that slides received after that will no longer be included in the evaluation.

For each antibody staining, please fill a separate staining techniques protocol form of the used method via LabScala. We can provide a more comprehensive final report that serves the customer if the method information is given as detailed as possible.

S001: HE



S002: Iron (Fe)



2023-04-25

### INSTRUCTIONS

Product no. 6543  
LQ778423011-012/FI

If the kit is incomplete or contains damaged specimens, please report immediately to [info@labquality.fi](mailto:info@labquality.fi)

---

The slides should be returned no later than **May 26, 2023.**

---

### Inquiries

EQA Coordinator  
Pia Eloranta  
[pia.eloranta@labquality.fi](mailto:pia.eloranta@labquality.fi)

### Labquality Oy

Kumpulantie 15  
FI-00520 HELSINKI  
Finland

Tel. + 358 9 8566 8200  
Fax + 358 9 8566 8280

[info@labquality.fi](mailto:info@labquality.fi)  
[www.labquality.com](http://www.labquality.com)




Client report

	No of participants	No of responded participants	Response percentage
Histological staining techniques, April, 1-2023	111	111	100 %

Summary

Overall success rate by samples




Summary	Own score	Max score	Own success rate	Difference	AVR success rate
S001	4	5	80 %	5.4 %	74.6 %
S002	4	5	80 %	-8.6 %	88.6 %
Average:			80 %	-1.6 %	81.6 %

History	Test nr.	Own success rate	Difference	AVR success rate
History not found				

S001 | Sample S001 HE


Performance




GENERAL DETAILS OF STAINING

Performance	Performance count
Automatic	80
Both options above	3
<input checked="" type="radio"/> Manual	14
Total:	97

Instrument




Instrument	Instrument count
Agilent Dako CoverStainer	6
Leica BioSystems HistoCore Spectra ST	4
Leica BioSystems Leica ST4020	1
Leica BioSystems Leica ST5010	6
Leica BioSystems Leica ST5010-CV5030	3
Leica BioSystems Leica ST5020	3
Leica BioSystems Leica ST5020-CV5030	3
Medite TST 44	1
Other, please specify in comments field	9
Roche Ventana Benchmark	1
Roche Ventana HE 600	6
Sakura Tissue-Tek DRS	2
Sakura Tissue-Tek DRS 2000	2
Sakura Tissue-Tek Prisma	15
Sakura Tissue-Tek Prisma Plus	10
Thermo Fisher Scientific Gemini AS	9
Total:	81



**STAINING METHOD**

Reagent	Reagent count
<input checked="" type="radio"/> Commercial reagent	282
Self made reagent	64
Total:	346


Reagent



■ Acros Organics 
 ■ Agilent 
 ■ Atom Scientific 
 ■ Bio Optica 
 ■ Biognost 
 ■ Dako 
 ■ Diapath 
 ■ Merck 
 ■ MicroChem 
 ■ Other, please specify in comments field 
 ■ Reagena 
 ■ Roche 
 ■ Sigma-Aldrich 
 ■ Thermo Fisher Scientific 
 ■ VWR

Manufacturer	Manufacturer count
Acros Organics	1
Agilent	2
Atom Scientific	7
Bio Optica	25
Biognost	29
<input checked="" type="radio"/> Dako	16
Diapath	10
Merck	21
MicroChem	6
Other, please specify in comments field	91
Reagena	3
Roche	12
Sigma-Aldrich	14
Thermo Fisher Scientific	10
VWR	24
Total:	271

S001 Score




SCORE

Score	Comment
4	weak contrast

Score	Score count
-	4
2	2
3	37
<input checked="" type="radio"/> 4	57
5	12
Total:	112

S002 | Sample S002 Iron


Performance



GENERAL DETAILS OF STAINING


Performance	Performance count
Automatic	23
Both options above	1
<input checked="" type="radio"/> Manual	57
Total:	81

Instrument



Instrument	Instrument count
Agilent Dako Artisan Link Pro	7
Leica BioSystems Leica ST5010-CV5030	1
Other, please specify in comments field	2
Roche Ventana Benchmark	13
Sakura Tissue-Tek Prisma Plus	1
Thermo Fisher Scientific Gemini AS	1
Total:	25


S002 Staining method



STAINING METHOD

Reagent	Reagent count
<input checked="" type="radio"/> Commercial reagent	119
Self made reagent	50
Total:	169

Reagent




Manufacturer	Manufacturer count
AFT Bratislava	1
Agilent	2
<input checked="" type="radio"/> Atom Scientific	5
Bio Optica	35



Biognost	14
Centralchem	7
Dako	9
Diapath	12
Merck	18
MicroChem	2
Other, please specify in comments field	8
Reagen	2
Roche	21
Roth	1
Sigma-Aldrich	2
VWR	6
Total:	145

S002 Score



SCORE


Score	Comment
4	background

Score	Score count
-	24
0	3
2	1
3	1
<input checked="" type="radio"/> 4	30
5	53
Total:	112

GLOBAL REPORT

	No of participants	No of responded participants	Response percentage
Histological staining techniques, April, 1-2023	111	111	100 %


Summary



Summary	AVR success rate
S001	74.6 %
S002	88.6 %
Average:	81.6 %

S001 | Sample S001 HE


Performance



GENERAL DETAILS OF STAINING

Performance	Performance count
Automatic	80
Both options above	3
Manual	14
Total:	97

Instrument




Instrument	Instrument count
Agilent Dako CoverStainer	6
Leica BioSystems HistoCore Spectra ST	4
Leica BioSystems Leica ST4020	1
Leica BioSystems Leica ST5010	6
Leica BioSystems Leica ST5010-CV5030	3
Leica BioSystems Leica ST5020	3
Leica BioSystems Leica ST5020-CV5030	3
Medite TST 44	1
Other, please specify in comments field	9
Roche Ventana Benchmark	1
Roche Ventana HE 600	6
Sakura Tissue-Tek DRS	2
Sakura Tissue-Tek DRS 2000	2
Sakura Tissue-Tek Prisma	15
Sakura Tissue-Tek Prisma Plus	10
Thermo Fisher Scientific Gemini AS	9
<b>Total:</b>	<b>81</b>




### STAINING METHOD

Reagent	Reagent count
Commercial reagent	282
Self made reagent	64
<b>Total:</b>	<b>346</b>



Manufacturer	Manufacturer count
Acros Organics	1
Agilent	2
Atom Scientific	7
Bio Optica	25
Biognost	29
Dako	16
Diapath	10
Merck	21
MicroChem	6
Other, please specify in comments field	91
Reagen	3
Roche	12
Sigma-Aldrich	14
Thermo Fisher Scientific	10
VWR	24
Total:	271

S001 Score



SCORE


Score	Comment
-	Not available
2	clearly overstained, mucus light, Paneth cells and eosinophils poorly discernible
3	dirty hematoxylin overstained hematoxylin overstained, mucus hematoxylin overstained, strong mucus light overstained overstained, mucus poor contrast strong hematoxylin strong hematoxylin, weak contrast strong hematoxylin, weak Paneth cells too strong eosin too strong hematoxylin unclear contrast, mucus weak weak contrast weak contrast, mucus too strong weak contrast, weak eosinophils weak eosin, Paneth cells not discernible weak eosinophils weak eosinophils, mucus, dirty weak eosinophils, poorly discernible weak Paneth cells and eosinophils
4	eosin too strong eosinophils poorly discernible hematoxylin strong light light, weak Paneth cells overstained slightly weak slightly weak contrast some overstained strong eosin strong hematoxylin strong mucus, very blue too strong eosin

	weak
	weak a little
	weak contrast
	weak contrast, too strong hematoxylin
	weak eosin
	weak eosinohils
	weak eosinophils
	weak eosinophils, mucus
	weak Paneth cells
	weak Paneth cells and eosinophils
5	
	Excellent
	mucus

Score	Score count
-	4
2	2
3	37
4	57
5	12
Total:	112

S002 | Sample S002 Iron


Performance



GENERAL DETAILS OF STAINING

Performance	Performance count
Automatic	23
Both options above	1
Manual	57
Total:	81


Instrument



Instrument	Instrument count
Agilent Dako Artisan Link Pro	7
Leica BioSystems Leica ST5010-CV5030	1
Other, please specify in comments field	2
Roche Ventana Benchmark	13
Sakura Tissue-Tek Prisma Plus	1
Thermo Fisher Scientific Gemini AS	1
Total:	25




### S002 Staining method



### STAINING METHOD

Reagent	Reagent count
Commercial reagent	119
Self made reagent	50
Total:	169

### Reagent




■ AFT Bratislava 
 ■ Agilent 
 ■ Atom Scientific 
 ■ Bio Optica 
 ■ Biognost 
 ■ Centralchem 
 ■ Dako 
 ■ Diapath 
 ■ Merck 
 ■ MicroChem 
 ■ Other, please specify in comments field 
 ■ Reagens 
 ■ Roche 
 ■ Roth 
 ■ Sigma-Aldrich 
 ■ VWR

Manufacturer	Manufacturer count
AFT Bratislava	1
Agilent	2
Atom Scientific	5
Bio Optica	35

Biognost	14
Centralchem	7
Dako	9
Diapath	12
Merck	18
MicroChem	2
Other, please specify in comments field	8
Reagen	2
Roche	21
Roth	1
Sigma-Aldrich	2
VWR	6
Total:	145

S002 Score



SCORE

Score	Comment
-	Not available
0	HE stain done iron not stained, probably HE done negative, iron not stained
2	clearly overstained, background
3	weak, uneven
4	almost optimal background overstained some background weak
5	Excellent

Score	Score count
-	24
0	3
2	1
3	1
4	30

5	53
Total:	112

## Histological staining techniques: HE and Iron (Fe) Round 1, 2023

### Specimens

Sample S001-S002 (LQ778423011-LQ778423012) were two slides, each with an unstained paraffin section.

S001: HE (hematoxylin & eosin) (small intestine, ileum)

S002: Iron (Fe) (lung and kidney)

### Report info

The final report contains distribution of results and used methods. Laboratory's own result and method is marked with a black radio button (☉). A numerical score given in a six-step scale 0–5 is based on consensus. The results of all participants are presented in a table form.

### Evaluation scale: 0-5

3-5 points indicate good enough staining for diagnosis.

0-2 points mean that the staining is insufficient for diagnosis or failed.

<b>5 points</b>	optimal, excellent
<b>4 points</b>	almost optimal, practically faultless, slight over/understaining, slightly uneven or patchy staining
<b>3 points</b>	acceptable for diagnosis, but distinct over/understaining, uneven or patchy staining, stain deposits etc.
<b>2 points</b>	borderline, weak staining, uncertain for diagnosis
<b>1 point</b>	poor, failed, some scanty but inadequate staining observed/notifiable overstaining.
<b>0 points</b>	negative staining or fully false positive, failed.

Points	HE (n)	Iron (n)
5	12	53
4	57	30
3	37	1
2	2	1
1		
0		3
Mean	3.73	4.43

### Comments – Expert

#### S001: Criteria for optimal HE –staining

Blue hematoxylin and red eosin should clearly differentiate tissue components. The intensity of staining was to be adequate, and overstaining was not allowed. Staining of mucus should not disturb interpretation. Cell membranes and chromatin were to be distinct. The Paneth cells and granules of eosinophils must be clearly discernible.

### Results

In the survey of HE -stain 61/108 (56.5 %) of the participants got 4 or 5 points which means excellent performance. Only 2/108 (1.9 %) showed insufficient performance (2 points). As a whole, mean of the performance was 3.73 points.

### Automation and reagents

The mean of those who used automates was 3.76 points which is almost the same as 3.64 points achieved by manual process. The mean was 3.67 points when the protocols were not reported. The participants mentioned 14 different

2023-06-28

### FINAL REPORT

Product no. 6543

Subcontracting: Sample preparation, sample pretesting

Samples sent 2023-04-25  
Round closed 2023-05-30  
Final report 2023-06-28

### Request for correction

Typing errors in laboratory's result forms are on laboratory's responsibility. Labquality accepts responsibility only for result processing. Requests must be notified by writing within three weeks from the date of this letter.

### Authorized by

EQA Coordinator  
Pia Eloranta  
pia.eloranta@labquality.fi

### Expert

Chief of department Heikki Aho, MD, PhD  
Turku university hospital, pathology

### Evaluation of the slides with experts

Chief of department Mirva Söderström, MD, PhD  
Cell biologist Iina Tuominen, PhD  
Turku university hospital, pathology

### Labquality Oy

Kumpulantie 15  
FI-00520 HELSINKI  
Finland

Tel. + 358 9 8566 8200

info@labquality.fi  
www.labquality.com




automates. Four automates achieved 5 points and 12 got 4 or 5 points which means that there were no essential differences in performance between the automates. The same automate model may result to 3, 4 or 5 points. There were 10 automates and 2 manual processes in the group of 5 points. The reagents varied from variable commercial products of different sources to home-made ones.

### **Deficiencies in the HE -staining**

Usual deficiencies were overstaining, deficient staining and disturbance caused by staining of mucus. Also deficient staining of the Paneth cells and eosinophils reduced the points. Only seldom both Paneth cells and eosinophils were clearly distinct. The chromatin was mostly well stained.

### **Protocols that produced optimal staining**

The type of automate used and the reagents, usually commercial, varied a lot and it is very difficult to give any preferable protocols. Is the final result good or less good is also often a matter of taste. The type of hematoxylin, however, influence on the staining of mucus.



**Image:** HE-stain of 5 points. Automate Sakura TissueTek Prisma Plus.

### **S002: Criteria for optimal Iron –staining**

A lot of hemosiderin, mostly in macrophages, was present in this specimen from the lung. The iron should be clearly discernible in granules of macrophages without diffuse spread to surroundings.

### **Results**

Staining of iron was successful, 4 or 5 points were achieved by 83/88 (94.3 %) of the participants. Only 4 (4.5 %) were unsuccessful (0-2 points).

### **Automation and reagents**

Automate was used in 23 laboratories, mean 4.22 points. Manual process was used in 54 laboratories, mean 4.59 points. Manual process seemed to be a little more reliable. Six automates were named and all achieved 4 or 5 points. When protocol information was not given, the performance was 4,18 points. Those who got the best 5 points result used usually commercial reagents but also with home-made reagents it was possible to get the best performance.

### **Deficiencies in the Iron staining**

One slide was heavily overstained and 3 were totally negative (only HE-stain done). Spread to surroundings was considered as a minor deficiency.

### Protocols that produced optimal staining

Optimal result was possible both with automate and by manual process. The manual process, however, was a little reliable. Incubation periods varied considerably between 20 sec to 80 min and the reagents were variable, so it is difficult to give any recommendable protocol when almost all laboratories succeeded excellently.




Image: Iron stain of 5 points. Manual process. Perls Diapath.

End of the report

Copyright © Labquality Oy

Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copyright. Labquality prohibits any use of its name, or reference to Labquality EQA program, or material in this report in any advertising, brochures or other commercial publications. Labquality EQA data do not necessarily indicate the superiority of instruments, reagents, testing equipment or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of equipment or materials may be deceptive and misleading. Proficiency test results are handled confidentially. Labquality will not issue any statements to third parties of the performance of laboratories in external quality assessment schemes unless otherwise agreed.