

External Quality Assessment Scheme

## ***Streptococcus agalactiae* (GBS), culture** **Round 1, 2023**

### **Specimens**

Please find enclosed 2 lyophilized samples S001 and S002 (in foil packages) and blue-capped vials of rehydration fluid, each 1 mL.

### **Caution**

The specimens simulate patient samples, and they should be handled in the same way as corresponding samples capable of transmitting infectious disease.

### **Examinations**

*Streptococcus agalactiae* (GBS), culture

### **Storage and use**

After arrival, the samples should be stored at +2...8 °C. Please follow the handling instructions. Excessive shaking of the vials during the rehydration is not recommendable as it may produce foam and retard the dissolution process.

1. Let the specimens and the rehydration fluids warm up to room temperature.
2. Discard the blue-coloured cap from the vial of rehydration fluid.
3. Transfer the colourless cap (on vial inside the foil package) to the vial of the rehydration fluid. The bacterial specimen is fixed in the black particles inside the colourless cap (the paper pad inside the vial with the colourless cap is a desiccant).
4. Place the vial in an inverted position, tap to be sure the liquid is in contact with the inside of the cap and place it in an incubator at +35...37 °C for 15 minutes to dissolve the preserved microbes.
5. Check that there are no black particles inside the cap to make certain that all the microbes have dissolved into the solution. If necessary, reinvert the vial and continue the rehydration process. Observe closely every 1 to 2 minutes for complete dissolution. Continue to dissolve the specimen until no black particles are left in the cap, however, note that a prolonged rehydration process may affect the bacterial count. To enhance the dissolution of the particles you may gently tap the bottom of the vial when keeping it in an inverted position. Excessive vigorous shaking will produce foam.
6. Proceed immediately with the examination as with a patient sample.

The specimens should be cultured and incubated immediately after the preparations, like patient samples. Preferably use a 1 µL culture loop (both for plate and enrichment culture). Incubate up to 48 h if needed.

### **Result reporting**

Please enter the results via LabScala ([www.labscala.com](http://www.labscala.com)). Report the findings and results according to your own test selection, leave the parts empty that are not included in your examinations. If you cannot find your test in the dropdown menu, please contact the EQA Coordinator. Merely the results reported in the "Report to the clinician" part will be scored.

S001



S002



2023-03-28

### **INSTRUCTIONS**

Product no. 5594  
LQ759423011-012/US  
UN3373

Subcontracting: Sample pretesting

If the kit is incomplete or contains damaged specimens, please report immediately to [info@labquality.fi](mailto:info@labquality.fi)

---

The results should be reported no later than  
**April 24, 2023.**

---

The expected results of the round are published in LabScala in the View reports section by April 26, 2023.

### **Inquiries**

EQA Coordinator  
Yvonne Björkman  
[yvonne.bjorkman@labquality.fi](mailto:yvonne.bjorkman@labquality.fi)

### **Labquality Oy**

Kumpulantie 15  
FI-00520 HELSINKI  
Finland

Tel. + 358 9 8566 8200  
Fax + 358 9 8566 8280


[info@labquality.fi](mailto:info@labquality.fi)  
[www.labquality.com](http://www.labquality.com)



Client report

	No of participants	No of responded participants	Response percentage
Streptococcus agalactiae (GBS), culture, April, 1-2023	82	76	92.7 %

Summary




Summary	Own score	Max score	Own success rate	Difference	AVR success rate
Specimen S001	4	4	100 %	1.3 %	98.7 %
Specimen S002	4	4	100 %	3.9 %	96.1 %
Average:			100 %	2.6 %	97.4 %

History	Test nr.	Own success rate	Difference	AVR success rate
Round 2022-4	1	100 %	4 %	96 %
Round 2022-3	1	100 %	4.3 %	95.7 %
Round 2022-2	1	100 %	4.9 %	95.1 %
Round 2021-4	1	100 %	12.3 %	87.7 %
Round 2021-3	1	100 %	5.7 %	94.3 %
Round 2021-2	1	100 %	1.9 %	98.1 %
Round 2021-1	1	100 %	4 %	96 %
Round 2020-4	1	100 %	2.8 %	97.2 %


Specimen S001 | Streptococcus agalactiae (GBS): negative

Specimen S001 success rate



Specimen S001 results	Responded	Count
	Report to the clinician	76
	Plate Culture	71
	Agglutination test	8
	Commercial identification kits and analyzers	3
	Identification tests: MALDI-TOF	18
	Identification tests on colony: NAT	8
	Chromogenic medium	25
	Identification tests: Ag detection	2


Specimen S001 Report to the clinician



REPORT TO THE CLINICIAN

Report to the clinician	Interpretation	Interpretation count	Own score	Max score	Own success rate	Difference	AVR success rate
	<input checked="" type="radio"/> Negative	75	4	4	100 %	0 %	100 %
	Positive	1	-				0 %
	Total:	76	4	4	100 %	1.3 %	98.7 %

Specimen S001 Plate Culture



Specimen S001 Agglutination test





PLATE CULTURE

Method	No growth (-)	Scarcely growth (+)	Moderate growth (++)	Heavy growth (+++)	Total:
Chromogenic medium	35	1	10	1	47
Selective medium	17	2	3	2	24
Total:	52	3	13	3	71


AGGLUTINATION TEST

Method	Negative	Positive	Total:
Microgen Strep Rapid Test (Microgen Bioproducts)	1	-	1
Oxoid Streptococcal Grouping Kit (Thermo Scientific)	2	-	2
Pastorex Strep (Bio-Rad)	1	1	2
Prolex Streptococcal Grouping Kit (Pro-Lab Diagnostics)	2	-	2
Remel PathoDextra Strep Grouping Kit (Thermo Scientific)	1	-	1
Total:	7	1	8


Specimen S001 Commercial identification kits and analyzers



Specimen S001 Identification tests: MALDI-TOF



Specimen S001 Identification tests on colony: NAT



COMMERCIAL IDENTIFICATION KITS AND ANALYZERS

Method	Result	count
VITEK 2 (bioMerieux)	Other than S. agalactiae	2
VITEK 2 Compact 60 (bioMerieux)	Other than S. agalactiae	1
Total:		3


IDENTIFICATION TESTS: MALDI-TOF

Method	Result	count
Autof MALDI-ToF (Chirus)	Streptococcus agalactiae	1
MALDI Biotyper (Bruker)	Negative	3
	Other than S. agalactiae	8
VITEK MS (bioMérieux)	Other than S. agalactiae	6
Total:		18

IDENTIFICATION TESTS ON COLONY: NAT

Method	Result	count
PCR (In-house)	Negative	1
Xpert GBS (Cepheid)	Negative	5
Xpert Xpress GBS (Cepheid)	Negative	2
Total:		8

Specimen S001 Chromogenic medium



CHROMOGENIC MEDIUM

Result	Result count
<input checked="" type="radio"/> Negative	21
Positive	3
Indeterminate result	1
Total:	25


Specimen S001 Identification tests: Ag detection



## IDENTIFICATION TESTS: AG DETECTION


Method	Result	count
Nadal Strep B Test (nal von minden)	Negative	1
StrepBStick (Novamed)	Negative	1
Total:		2

Specimen S002 | Streptococcus agalactiae (GBS): positive



Specimen S002 results	Responded	Count
	Report to the clinician	76
	Plate Culture	76
	Agglutination test	32
	Commercial identification kits and analyzers	13
	Identification tests: MALDI-TOF	50
	Identification tests on colony: NAT	12
	Chromogenic medium	31
	Identification tests: Ag detection	3


Specimen S002 Report to the clinician



REPORT TO THE CLINICIAN

Report to the clinician	Interpretation	Interpretation count	Own score	Max score	Own success rate	Difference	AVR success rate
	Negative	3	-				0 %
	<input checked="" type="radio"/> Positive	73	4	4	100 %	0 %	100 %
	Total:	76	4	4	100 %	3.9 %	96.1 %

Specimen S002 Plate Culture



■ No growth (-)  
 ■ Scarce growth (+)  
 ■ Moderate growth (++)  
 ■ Heavy growth (+++)

Specimen S002 Agglutination test



■ Positive

PLATE CULTURE

Method	No growth (-)	Scarce growth (+)	Moderate growth (++)	Heavy growth (+++)	Total:
Chromogenic medium	1	11	28	⊙ 11	51
Selective medium	1	5	15	4	25
Total:	2	16	43	15	76

AGGLUTINATION TEST


Method	Positive
Microgen Strep Rapid Test (Microgen Bioproducts)	1
Oxoid Streptococcal Grouping Kit (Thermo Scientific)	13
Pastorex Strep (Bio-Rad)	4
Prolex Streptococcal Grouping Kit (Pro-Lab Diagnostics)	⊙ 7
Remel PathoDx Strep Grouping Kit (Thermo Scientific)	2
Remel PathoDextra Strep Grouping Kit (Thermo Scientific)	4
Streptex Rapid Latex Agglutination (Thermo Scientific)	1
Total:	32

Specimen S002 Commercial identification kits and analyzers




■ Streptococcus agalactiae

Specimen S002 Identification tests: MALDI-TOF



■ Streptococcus agalactiae  
 ■ Other than S. agalactiae

Specimen S002 Identification tests on colony: NAT



■ Positive  
 ■ Streptococcus agalactiae

COMMERCIAL IDENTIFICATION KITS AND ANALYZERS

Method	Result	count
VITEK 2 (bioMerieux)	Streptococcus agalactiae	6
VITEK 2 Compact 15 (bioMerieux)	Streptococcus agalactiae	3
VITEK 2 Compact 60 (bioMerieux)		



	Streptococcus agalactiae	4
Total:		13


IDENTIFICATION TESTS: MALDI-TOF

Method	Result	count
Autof MALDI-ToF (Chirus)		
	Streptococcus agalactiae	2
MALDI Biotyper (Bruker)		
	Streptococcus agalactiae	27
	Other than S. agalactiae	1
VITEK MS (bioMérieux)		
	Streptococcus agalactiae	20
Total:		50

IDENTIFICATION TESTS ON COLONY: NAT

Method	Result	count
PCR (In-house)		
	Positive	1
	Streptococcus agalactiae	2
Xpert GBS (Cepheid)		
	Positive	5
Xpert Xpress GBS (Cepheid)		
	Positive	3
	Streptococcus agalactiae	1
Total:		12


Specimen S002 Chromogenic medium



CHROMOGENIC MEDIUM

Result	Result count
<input checked="" type="radio"/> Positive	29
Indeterminate result	2
Total:	31

Specimen S002 Identification tests: Ag detection



IDENTIFICATION TESTS: AG DETECTION

Method	Result	count
Nadal Strep B Test (nal von minden)	Weak positive	1
StrepBStick (Novamed)	Positive	2
Total:		3

## Report Info

### PARTICIPANTS

Altogether 82 laboratories from 13 countries participated in this EQA round.

### REPORT INFO

On the front page you can see a summary of the overall average success rate and sample specific success rates which have been calculated from the scores. The reported results and the scores are presented in separate tables.

In general, the expected results are marked with green color. Accepted results may also be indicated with yellow color. Laboratory's own results are indicated with a black radio button . In the participant specific report, the participant's own success is shown both as scores and as success rates (%) generated from the score values.

If you have not reported results, you will get a note: "You have not responded in time, only global report is available".

For information on report interpretation and performance evaluation, please see the "EQAS Interpretation guidelines" in LabScala User instructions. In case you have any questions regarding the reports, please contact the EQA Coordinator.

### SCORING

The results in the "Report to the clinician" part can be scored when at least 60% of the participants have reported the correct/expected result and when there are at least three reported results. Laboratory's scores have been converted to percentage (own success rate, % from maximum scores) with a target at 100%. Own success rate is compared with the success rate of all results.

The following general rules are applied:

4 points is reached by reporting the expected correct result

1-3 points is given to results that are partly correct/insufficient regarding the expected finding


0 points is given for an incorrect/false result

GLOBAL REPORT

	No of participants	No of responded participants	Response percentage
Streptococcus agalactiae (GBS), culture, April, 1-2023	82	76	92.7 %


Summary

Overall success rate by samples




Summary	AVR success rate
Specimen S001	98.7 %
Specimen S002	96.1 %
Average:	97.4 %

Specimen S001 | Streptococcus agalactiae (GBS): negative



Specimen S001 results	Responded	Count
	Report to the clinician	76
	Plate Culture	71
	Agglutination test	8
	Commercial identification kits and analyzers	3
	Identification tests: MALDI-TOF	18
	Identification tests on colony: NAT	8
	Chromogenic medium	25
	Identification tests: Ag detection	2


Specimen S001 Report to the clinician



REPORT TO THE CLINICIAN

Report to the clinician	Interpretation	Interpretation count	AVR success rate	Interpretation Score
	Negative	75	100 %	4
	Positive	1	0 %	0
	Total:	76	98.7 %	

Specimen S001 Plate Culture



Specimen S001 Agglutination test





PLATE CULTURE

Method	No growth (-)	Scarcely growth (+)	Moderate growth (++)	Heavy growth (+++)	Total:
Chromogenic medium	35	1	10	1	47
Selective medium	17	2	3	2	24
Total:	52	3	13	3	71

AGGLUTINATION TEST


Method	Negative	Positive	Total:
Microgen Strep Rapid Test (Microgen Bioproducts)	1	-	1
Oxoid Streptococcal Grouping Kit (Thermo Scientific)	2	-	2
Pastorex Strep (Bio-Rad)	1	1	2
Prolex Streptococcal Grouping Kit (Pro-Lab Diagnostics)	2	-	2
Remel PathoDextra Strep Grouping Kit (Thermo Scientific)	1	-	1
Total:	7	1	8

Specimen S001 Commercial identification kits and analyzers




Other than S. agalactiae

Specimen S001 Identification tests: MALDI-TOF



Negative Streptococcus agalactiae Other than S. agalactiae

Specimen S001 Identification tests on colony: NAT



Negative

COMMERCIAL IDENTIFICATION KITS AND ANALYZERS

Method	Result	count
VITEK 2 (bioMerieux)	Other than S. agalactiae	2
VITEK 2 Compact 60 (bioMerieux)	Other than S. agalactiae	1
Total:		3


IDENTIFICATION TESTS: MALDI-TOF

Method	Result	count
Autof MALDI-ToF (Chirus)	Streptococcus agalactiae	1
MALDI Biotyper (Bruker)	Negative	3
	Other than S. agalactiae	8
VITEK MS (bioMérieux)	Other than S. agalactiae	6
Total:		18

IDENTIFICATION TESTS ON COLONY: NAT

Method	Result	count
PCR (In-house)	Negative	1
Xpert GBS (Cepheid)	Negative	5
Xpert Xpress GBS (Cepheid)	Negative	2
Total:		8

Specimen S001 Chromogenic medium



CHROMOGENIC MEDIUM

Result	Result count
Negative	21
Positive	3
Indeterminate result	1
Total:	25

Specimen S001 Identification tests: Ag detection




## IDENTIFICATION TESTS: AG DETECTION

Method	Result	count
Nadal Strep B Test (nal von minden)	Negative	1
StrepBStick (Novamed)	Negative	1
Total:		2




Specimen S002 | Streptococcus agalactiae (GBS): positive



Specimen S002 results	Responded	Count
	Report to the clinician	76
	Plate Culture	76
	Agglutination test	32
	Commercial identification kits and analyzers	13
	Identification tests: MALDI-TOF	50
	Identification tests on colony: NAT	12
	Chromogenic medium	31
	Identification tests: Ag detection	3


Specimen S002 Report to the clinician



REPORT TO THE CLINICIAN

Report to the clinician	Interpretation	Interpretation count	AVR success rate	Interpretation Score
	Negative	3	0 %	0
	Positive	73	100 %	4
	Total:	76	96.1 %	

Specimen S002 Plate Culture



No growth (-) Scarce growth (+) Moderate growth (++)  
Heavy growth (+++)

Specimen S002 Agglutination test



Positive

PLATE CULTURE

Method	No growth (-)	Scarce growth (+)	Moderate growth (++)	Heavy growth (+++)	Total:
Chromogenic medium	1	11	28	11	51
Selective medium	1	5	15	4	25
Total:	2	16	43	15	76

AGGLUTINATION TEST


Method	Positive
Microgen Strep Rapid Test (Microgen Bioproducts)	1
Oxoid Streptococcal Grouping Kit (Thermo Scientific)	13
Pastorex Strep (Bio-Rad)	4
Prolex Streptococcal Grouping Kit (Pro-Lab Diagnostics)	7
Remel PathoDx Strep Grouping Kit (Thermo Scientific)	2
Remel PathoDextra Strep Grouping Kit (Thermo Scientific)	4
Streptex Rapid Latex Agglutination (Thermo Scientific)	1
Total:	32

Specimen S002 Commercial identification kits and analyzers




Streptococcus agalactiae

Specimen S002 Identification tests: MALDI-TOF



Streptococcus agalactiae  
Other than S. agalactiae

Specimen S002 Identification tests on colony: NAT



Positive Streptococcus agalactiae

COMMERCIAL IDENTIFICATION KITS AND ANALYZERS

Method	Result	count
VITEK 2 (bioMerieux)	Streptococcus agalactiae	6
VITEK 2 Compact 15 (bioMerieux)	Streptococcus agalactiae	3
VITEK 2 Compact 60 (bioMerieux)		

	Streptococcus agalactiae	4
Total:		13


**IDENTIFICATION TESTS: MALDI-TOF**

Method	Result	count
Autof MALDI-ToF (Chirus)		
	Streptococcus agalactiae	2
MALDI Biotyper (Bruker)		
	Streptococcus agalactiae	27
	Other than S. agalactiae	1
VITEK MS (bioMérieux)		
	Streptococcus agalactiae	20
Total:		50

**IDENTIFICATION TESTS ON COLONY: NAT**

Method	Result	count
PCR (In-house)		
	Positive	1
	Streptococcus agalactiae	2
Xpert GBS (Cepheid)		
	Positive	5
Xpert Xpress GBS (Cepheid)		
	Positive	3
	Streptococcus agalactiae	1
Total:		12


Specimen S002 Chromogenic medium



**CHROMOGENIC MEDIUM**

Result	Result count
Positive	29
Indeterminate result	2
Total:	31

Specimen S002 Identification tests: Ag detection



IDENTIFICATION TESTS: AG DETECTION

Method	Result	count
Nadal Strep B Test (nal von minden)	Weak positive	1
StrepBStick (Novamed)	Positive	2
Total:		3

## Report Info

### PARTICIPANTS

Altogether 82 laboratories from 13 countries participated in this EQA round.

### REPORT INFO

On the front page you can see a summary of the overall average success rate and sample specific success rates which have been calculated from the scores. The reported results and the scores are presented in separate tables.

In general, the expected results are marked with green color. Accepted results may also be indicated with yellow color. Laboratory's own results are indicated with a black radio button . In the participant specific report, the participant's own success is shown both as scores and as success rates (%) generated from the score values.

If you have not reported results, you will get a note: "You have not responded in time, only global report is available".

For information on report interpretation and performance evaluation, please see the "EQAS Interpretation guidelines" in LabScala User instructions. In case you have any questions regarding the reports, please contact the EQA Coordinator.

### SCORING

The results in the "Report to the clinician" part can be scored when at least 60% of the participants have reported the correct/expected result and when there are at least three reported results. Laboratory's scores have been converted to percentage (own success rate, % from maximum scores) with a target at 100%. Own success rate is compared with the success rate of all results.

The following general rules are applied:

4 points is reached by reporting the expected correct result

1-3 points is given to results that are partly correct/insufficient regarding the expected finding

0 points is given for an incorrect/false result

External Quality Assessment Scheme

## ***Streptococcus agalactiae* (GBS), culture** **Round 1, 2023**

### **Specimens**

The round included two lyophilized samples. Based on the quality controls conducted by the sample material manufacturer and the results obtained in the round, the sample lots are to be considered as homogeneous, stable and suitable for external quality assessment. The materials were sent without temperature control packaging.

The content of the samples was as follows:

Sample S001 (LQ759423011)  
*Streptococcus* sp. (viridans group) C020203  
*Gardnerella vaginalis* C04701  
Expected result: Negative

Sample S002 (LQ759423012)  
*Streptococcus agalactiae* C010502  
*Gardnerella vaginalis* C04701  
Expected result: Positive

### **Report info**

Please see the description of the data analysis on the last page of the laboratory-specific reports and global reports. It is important to read the Final report first, because it contains important information of the samples and results in each round.

### **Comments – Expert**

Sample S001 was expected to be GBS negative and included high amount of alpha hemolytic streptococci. The sample was reported incorrectly as positive by 1/76 participants.

Sample S002 was expected to be GBS positive but was reported as negative by 3/76 participants.

### **Exceptions in scoring**

No exceptions.

### **End of report**

2023-05-05

### **FINAL REPORT**

Product no. 5594

Subcontracting: sample pretesting

Samples sent	2023-03-28
Round closed	2023-04-24
Expected results	2023-04-26
Final report	2023-05-05

### **Request for correction**

Typing errors in laboratory's result forms are on laboratory's responsibility. Labquality accepts responsibility only for result processing. Requests must be notified by writing within three weeks from the date of this letter.

### **Authorized by**

EQA Coordinator  
Yvonne Björkman  
yvonne.bjorkman@labquality.fi

### **Expert**

Clinical Microbiologist, Ph.D.,  
Johanna Haiko, HUS Diagnostic  
Center, Clinical Microbiology  
Helsinki, Finland

### **Labquality Oy**

Kumpulantie 15  
FI-00520 HELSINKI  
Finland

Tel. + 358 9 8566 8200  
Fax + 358 9 8566 8280

info@labquality.fi  
www.labquality.com

