

LABQUALITY

External Quality Assessment Scheme

Hormones B Steroid and peptide hormones Round 1, 2023

Specimens

Please find enclosed 2 lyophilized serum samples S001 (B1), and S002 (B2), 3 mL each.

Caution

Quality control specimens derived from human blood must be handled with the same care as patient samples, i.e. as potential transmitters of serious diseases. The specimens are found to be HBsAg, HCVAb and HIVAgAb negative when tested with licensed reagents, but no known test method can offer complete assurance that the specimens will not transmit these or other infectious diseases.

Examinations

See page 2.

Storage and use

Lyophilized samples are stable at +2 ...8 °C until the closing date of the round.

Add 3.0 mL aqua to the samples. Ensure solubility and homogeneity. The reconstituted sample can be stored for 10 days at +2 to +8 °C (C-peptide for 3 days). The sample can be stored frozen at -20 °C for 4 weeks if the sample is frozen within 30 minutes after reconstitution. Thaw the sample shortly before analysis and mix until homogenous. Allow the sample to warm to room temperature before analysis.

Analyze both samples as patient samples.

Result reporting

Please enter the results and methods via LabScala (www.labscala.com). If you cannot find your instrument or reagent from the registry, please contact the EQA Coordinator.

S001



S002



2023-02-07

INSTRUCTIONS

Product no. 2301, 2301S, 1301
LQ732423011-012/NO

If the kit is incomplete or contains damaged specimens, please report immediately to info@labquality.fi

The results should be reported no later than **February 27, 2023.**

Inquiries

EQA Coordinator
Päivi Ranta
paivi.ranta@labquality.fi

Labquality Oy

Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200
Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.com



Examinations

Aldosterone

Androstendione

C-peptide

Cortisol

DHEAS

Estradiol

FSH

Gastrin

Growth hormone

Insulin

LH

Progesterone, 17-OH-progesterone

Prolactin




SHBG

IGF-1

TBG


Testosterone and testosterone, free

Cortisol 0030 e601 Roche cobas e, Elecsys, & Modular... 360 Roche Elecsys 60



Previous results

| Round | Sample | X | Own | Diff% | z |
|-------|---------|-----|-----|-------|------|
| 23/01 | HaD 360 | 268 | 273 | 1.9 | 0.5 |
| 23/01 | HaF 360 | 704 | 665 | -5.6 | -1.1 |
| 22/04 | HbZ 360 | 164 | 172 | 5.2 | 1.2 |
| 22/04 | HaF 360 | 714 | 794 | 11.1 | 2.1 |
| 22/02 | HbX 360 | 362 | 370 | 2.3 | 0.6 |
| 22/02 | HaF 360 | 699 | 686 | -1.9 | -0.4 |
| 21/06 | Hbl 360 | 335 | 350 | 4.3 | 1.1 |
| 21/06 | HaF 360 | 691 | 725 | 5.0 | 1.1 |






Comments

Sample B1, Own group:
The standard uncertainty of the assigned value (X_{rob}) is 1.25 x standard error of the mean (SEM) and marked as 'u'.


Sample B2, Own group:
The standard uncertainty of the assigned value (X_{rob}) is 1.25 x standard error of the mean (SEM) and marked as 'u'.

C-peptide 0608 e601 Roche cobas e, Elecsys, & Modular... 360 Roche Elecsys 60



Previous results

| Round | Sample | X | Own | Diff% | z |
|-------|---------|------|------|-------|------|
| 23/01 | HaD 360 | 0.30 | 0.29 | -2.7 | -0.8 |
| 23/01 | HaF 360 | 1.25 | 1.15 | -8.0 | -2.3 |
| 22/04 | HbZ 360 | 1.39 | 1.58 | 13.9 | 1.8 |
| 22/04 | HaF 360 | 1.26 | 1.39 | 10.6 | 2.7 |
| 22/02 | HbX 360 | 0.97 | 1.04 | 7.3 | 0.5 |
| 22/02 | HaF 360 | 1.26 | 1.32 | 5.0 | 1.3 |
| 21/06 | Hbl 360 | 1.33 | 1.41 | 6.3 | 0.7 |
| 21/06 | HaF 360 | 1.28 | 1.26 | -1.3 | -0.2 |






Comments

Sample B1, Own group:
The standard uncertainty of the assigned value (X_{rob}) is 1.25 x standard error of the mean (SEM) and marked as 'u'.


Sample B2, Own group:
The standard uncertainty of the assigned value (X_{rob}) is 1.25 x standard error of the mean (SEM) and marked as 'u'.

FSH 0633 Roche cobas e, Elecsys, & Modular... Roche Elecsys
e601 360 60



Previous results

| Round | Sample | X | Own | Diff% | z |
|-------|---------|------|------|-------|------|
| 23/01 | HaD 360 | 4.6 | 4.4 | -4.5 | -1.3 |
| 23/01 | HaF 360 | 36.1 | 33.1 | -8.3 | -1.9 |
| 22/04 | HbZ 360 | 41.9 | 41.5 | -1.2 | -0.4 |
| 22/04 | HaF 360 | 36.4 | 35.5 | -2.5 | -0.6 |
| 22/02 | HbX 360 | 4.0 | 4.0 | 0.4 | 0.1 |
| 22/02 | HaF 360 | 36.7 | 37.1 | 0.9 | 0.2 |
| 21/06 | Hbl 360 | 29.2 | 28.3 | -3.1 | -1.2 |
| 21/06 | HaF 360 | 37.2 | 36.6 | -1.7 | -0.8 |






Comments

Sample B1, Own group:
The standard uncertainty of the assigned value (X_{rob}) is 1.25 x standard error of the mean (SEM) and marked as 'u'.


Sample B2, Own group:
The standard uncertainty of the assigned value (X_{rob}) is 1.25 x standard error of the mean (SEM) and marked as 'u'.

GH 0638 Roche cobas e, Elecsys, & Modular... Roche Elecsys
e601 360 60



Previous results

| Round | Sample | X | Own | Diff% | z |
|-------|---------|-----|-----|-------|------|
| 23/01 | HaD 360 | 2.2 | 2.2 | 0.0 | 0.5 |
| 23/01 | HaF 360 | 9.1 | 8.6 | -5.4 | -1.1 |
| 22/04 | HbZ 360 | 1.4 | 1.5 | 8.6 | 0.9 |
| 22/04 | HaF 360 | 9.1 | 8.9 | -2.1 | -0.4 |
| 22/02 | HbX 360 | 0.5 | 0.5 | 5.1 | 0.7 |
| 22/02 | HaF 360 | 9.1 | 9.3 | 2.7 | 0.6 |
| 21/06 | Hbl 360 | 1.4 | 1.4 | 0.7 | 0.5 |
| 21/06 | HaF 360 | 9.0 | 8.5 | -5.2 | -1.3 |






Comments

Sample B1, Own group:
The standard measurement uncertainty (u) of the assigned value (X_{pt}) is the standard error of the mean (SEM). Z score is uncertain due to the small number of observations.


Sample B2, Own group:
The standard measurement uncertainty (u) of the assigned value (X_{pt}) is the standard error of the mean (SEM). Z score is uncertain due to the small number of observations.

IGF-1 0671 e601 Roche cobas e, Elecsys, & Modular... Roche Elecsys 360 60



Previous results

| Round | Sample | X | Own | Diff% | z |
|-------|---------|------|------|-------|------|
| 23/01 | HaD 360 | 17.2 | 16.6 | -3.7 | -1.0 |
| 23/01 | HaF 360 | 15.9 | 14.0 | -11.8 | -1.1 |
| 22/04 | HbZ 360 | 17.7 | 16.8 | -5.2 | - |
| 22/04 | HaF 360 | 15.6 | 14.8 | -4.9 | - |
| 22/02 | HbX 360 | 15.8 | 16.0 | 1.3 | 0.4 |
| 22/02 | HaF 360 | 15.6 | 15.7 | 0.8 | 0.3 |






Comments

Sample B1, Own group:
The standard measurement uncertainty (u) of the assigned value (X_{pt}) is the standard error of the mean (SEM). Z score is uncertain due to the small number of observations.


Sample B2, Own group:
The standard measurement uncertainty (u) of the assigned value (X_{pt}) is the standard error of the mean (SEM). The uncertainty (u) of the assigned value is not negligible, and evaluations could be affected. Z score is uncertain due to the small number of observations.

Insulin 0650 e601 Roche cobas e, Elecsys, & Modular... Roche Elecsys 360 60



Previous results

| Round | Sample | X | Own | Diff% | z |
|-------|---------|-----|-----|-------|------|
| 23/01 | HaD 360 | 9 | 10 | 5.5 | 1.1 |
| 23/01 | HaF 360 | 131 | 135 | 3.1 | 0.8 |
| 22/04 | HbZ 360 | 12 | 16 | 27.6 | 1.7 |
| 22/04 | HaF 360 | 131 | 135 | 3.3 | 1.1 |
| 22/02 | HbX 360 | 13 | 14 | 9.0 | 0.9 |
| 22/02 | HaF 360 | 132 | 132 | 0.2 | 0.0 |
| 21/06 | Hbl 360 | 21 | 24 | 11.4 | 0.9 |
| 21/06 | HaF 360 | 130 | 129 | -1.3 | -0.4 |






Comments

Sample B1, Own group:
The standard uncertainty of the assigned value (X_{rob}) is 1.25 x standard error of the mean (SEM) and marked as 'u'.


Sample B2, Own group:
The standard uncertainty of the assigned value (X_{rob}) is 1.25 x standard error of the mean (SEM) and marked as 'u'.

LH 0653 e601 Roche cobas e, Elecsys, & Modular... 360 Roche Elecsys 60



Previous results

| Round | Sample | X | Own | Diff% | z |
|-------|---------|------|------|-------|------|
| 23/01 | HaD 360 | 4.7 | 4.5 | -4.6 | -1.0 |
| 23/01 | HaF 360 | 50.8 | 44.7 | -12.1 | -2.8 |
| 22/04 | HbZ 360 | 17.1 | 17.4 | 1.8 | 0.7 |
| 22/04 | HaF 360 | 50.8 | 51.0 | 0.5 | 0.2 |
| 22/02 | HbX 360 | 4.4 | 4.5 | 1.6 | 0.5 |
| 22/02 | HaF 360 | 50.5 | 52.4 | 3.8 | 1.1 |
| 21/06 | Hbl 360 | 15.8 | 15.1 | -4.5 | -1.3 |
| 21/06 | HaF 360 | 51.2 | 50.0 | -2.3 | -0.6 |



Comments

Sample B1, Own group:
The standard uncertainty of the assigned value (X_{rob}) is 1.25 x standard error of the mean (SEM) and marked as 'u'.

Sample B2, Own group:
The standard uncertainty of the assigned value (X_{rob}) is 1.25 x standard error of the mean (SEM) and marked as 'u'.

Participants

108 participants from 17 countries.

Report info

Assigned value (target value) calculation and its uncertainty

Your own result should be compared to others using the same method.

The assigned values (X_{rob}) are calculated according to the robust procedure described in the standard ISO 13528 (Statistical methods for use in proficiency testing by interlaboratory comparisons, Annex C, Algorithm A).

The standard uncertainty of the assigned value is expressed as $1.25 \times$ the standard error of mean (SEM) and marked as "u" in numerical summary. Due to its iterative mode algorithm A adds the uncertainty of the assigned value and with this factor we want to adjust uncertainty accordingly.

In case there are 2-12 results in a method group, the robust calculation is not used but assigned values (X_{pt}) are means of the results where results deviating more than $\pm 3 \times$ standard deviation from the median are removed. The standard uncertainty (u) of the assigned value is reported as standard error of the mean (SEM). Additionally, if the measurement uncertainty of the target value is too large ($u > 0.1 \times$ maximum allowable error) an automatic text is printed on the report: "The uncertainty of the assigned value is not negligible, and evaluations could be affected."

Please notice also that for groups that have only 1 result only the client's own result is shown. No target value (except for reference method values) is calculated, no target areas are shown.

Z score

In case there are 2-5 results in a method group, no z-score is calculated, and a text is printed on the report: "Due to the small number of results, the z score is not calculated." In case there are 6-12 results, the report has a text: "Z score is uncertain due to the small number of observations."

Results reported with $< \text{tai} >$ -signs cannot be included in the statistics.

For information on report interpretation and performance evaluation, please see the "EQAS Interpretation guidelines" in LabScala User instructions (top right corner ?Help link).

NUMERICAL SUMMARY **Hormone Determinations B 2023/01**

| Analyte | Method group | x | med | s | CV% | u | Min | Max | Number |
|--------------------------------|-------------------------------------|--------------|-------------|--------------|-------------|--------------|-------------|-------------|-----------|
| Sample B1 | | | | | | | | | |
| Aldosterone, pmol/L | | | | | | | | | |
| | Beckman Coulter Immunotech | 159.2 | 159 | 25.2 | 15.9 | 17.9 | 141 | 177 | 2 |
| | DiaSorin Liaison | 230.7 | 225 | 30.0 | 13.0 | 10.0 | 185 | 272 | 9 |
| | All | 218.9 | 224 | 43.0 | 19.7 | 13.0 | 141 | 272 | 11 |
| Androstenedione, nmol/L | | | | | | | | | |
| | DiaSorin Liaison | 4.16 | 4.3 | 0.29 | 7.0 | 0.17 | 3.8 | 4.4 | 3 |
| | DRG Diagnostics EIA | 5.27 | 5.3 | 2.47 | 46.8 | 1.75 | 3.5 | 7.0 | 2 |
| | IBL ELISA | 5.10 | 5.1 | - | - | - | - | - | 1 |
| | Mass spectrometry | 3.89 | 3.9 | 1.39 | 35.9 | 0.99 | 2.9 | 4.9 | 2 |
| | Roche cobas e, Elecsys, & Modular E | 2.90 | 2.9 | 0.14 | 4.7 | 0.08 | 2.8 | 3.0 | 3 |
| | Siemens Immulite 1000, 2000, 2500 | 4.90 | 4.9 | 0.69 | 14.1 | 0.40 | 4.2 | 5.6 | 3 |
| | Snibe Diagnostic Maglumi | 4.99 | 5.0 | - | - | - | - | - | 1 |
| | All | 4.22 | 4.3 | 1.17 | 27.7 | 0.30 | 2.8 | 7.0 | 15 |
| Cortisol, nmol/l | | | | | | | | | |
| | Abbott Alinity | 248.8 | 250 | 14.3 | 5.7 | 3.6 | 209 | 271 | 25 |
| | Abbott Architect | 249.7 | 256 | 12.0 | 4.8 | 6.9 | 236 | 257 | 3 |
| | DiaSorin Liaison | 160.1 | 160 | - | - | - | - | - | 1 |
| | Mass spectrometry | 266.0 | 266 | - | - | - | - | - | 1 |
| | Roche cobas e, Elecsys, & Modular E | 268.2 | 267 | 10.5 | 3.9 | 2.2 | 231 | 296 | 36 |
| | Siemens Advia Centaur & Atellica | 298.9 | 283 | 34.1 | 11.4 | 11.4 | 259 | 360 | 9 |
| | Siemens Immulite 1000, 2000, 2500 | 244.7 | 245 | 22.3 | 9.1 | 15.7 | 229 | 260 | 2 |
| | Tosoh AIA | 302.0 | 302 | - | - | - | - | - | 1 |
| | All | 262.3 | 263 | 17.9 | 6.8 | 2.0 | 160 | 360 | 78 |
| C-peptide, nmol/L | | | | | | | | | |
| | Abbott Alinity | 0.225 | 0.22 | 0.014 | 6.1 | 0.005 | 0.20 | 0.25 | 12 |
| | Abbott Architect | 0.241 | 0.24 | 0.008 | 3.5 | 0.006 | 0.24 | 0.25 | 2 |
| | DiaSorin Liaison | 0.324 | 0.32 | - | - | - | - | - | 1 |
| | Roche cobas e, Elecsys, & Modular E | 0.298 | 0.30 | 0.010 | 3.3 | 0.002 | 0.28 | 0.33 | 29 |
| | Siemens Advia Centaur & Atellica | 0.239 | 0.24 | 0.022 | 9.1 | 0.008 | 0.22 | 0.28 | 7 |
| | Siemens Immulite 1000, 2000, 2500 | 0.186 | 0.19 | 0.008 | 4.3 | 0.006 | 0.18 | 0.19 | 2 |
| | Tosoh AIA | 0.180 | 0.18 | - | - | - | - | - | 1 |
| | All | 0.268 | 0.29 | 0.044 | 16.5 | 0.006 | 0.18 | 0.33 | 54 |
| DHEAS, µmol/L | | | | | | | | | |
| | Abbott Alinity | 7.04 | 6.9 | 0.61 | 8.7 | 0.22 | 6.1 | 8.2 | 8 |
| | Abbott Architect | 6.73 | 6.7 | 0.59 | 8.7 | 0.41 | 6.3 | 7.1 | 2 |
| | Roche cobas e, Elecsys, & Modular E | 7.97 | 8.0 | 0.35 | 4.3 | 0.11 | 7.5 | 8.6 | 15 |
| | Siemens Advia Centaur & Atellica | 5.10 | 5.1 | - | - | - | - | - | 1 |
| | Siemens Immulite 1000, 2000, 2500 | 6.04 | 6.0 | 0.28 | 4.6 | 0.14 | 5.7 | 6.4 | 4 |
| | Tosoh AIA | 8.05 | 8.1 | - | - | - | - | - | 1 |
| | All | 7.35 | 7.6 | 0.95 | 12.9 | 0.17 | 5.1 | 8.6 | 31 |
| Estradiol, nmol/L | | | | | | | | | |
| | Abbott Alinity | 0.316 | 0.32 | 0.031 | 9.8 | 0.009 | 0.28 | 0.39 | 19 |
| | Abbott Architect | 0.308 | 0.30 | 0.029 | 9.3 | 0.013 | 0.28 | 0.36 | 5 |
| | Mass spectrometry | 0.328 | 0.33 | - | - | - | - | - | 1 |
| | Roche cobas e, Elecsys, & Modular E | 0.321 | 0.32 | 0.017 | 5.2 | 0.004 | 0.28 | 0.37 | 33 |
| | Siemens Advia Centaur & Atellica | 0.290 | 0.30 | 0.034 | 11.6 | 0.011 | 0.23 | 0.33 | 9 |
| | Siemens Dimension & Vista | 0.290 | 0.29 | 0.015 | 5.1 | 0.011 | 0.28 | 0.30 | 2 |
| | Siemens Immulite 1000, 2000, 2500 | 0.179 | 0.18 | - | - | - | - | - | 1 |
| | Tosoh AIA | 0.249 | 0.25 | - | - | - | - | - | 1 |

NUMERICAL SUMMARY
Hormone Determinations B 2023/01, Sample B1

| Analyte | Method group | x | med | s | CV% | u | Min | Max | Number |
|-------------------------------------|--------------|--------------|-------------|--------------|-------------|--------------|-------------|-------------|-----------|
| Estradiol, nmol/L | | | | | | | | | |
| All | | 0.313 | 0.32 | 0.025 | 8.1 | 0.003 | 0.18 | 0.39 | 71 |
| FSH, U/L | | | | | | | | | |
| Abbott Alinity | | 4.51 | 4.5 | 0.20 | 4.3 | 0.05 | 4.2 | 5.0 | 25 |
| Abbott Architect | | 4.39 | 4.4 | 0.29 | 6.6 | 0.13 | 4.0 | 4.7 | 5 |
| Roche cobas e, Elecsys, & Modular E | | 4.58 | 4.6 | 0.16 | 3.4 | 0.03 | 4.1 | 4.9 | 36 |
| Siemens Advia Centaur & Atellica | | 4.53 | 4.7 | 0.57 | 12.5 | 0.19 | 3.4 | 5.2 | 9 |
| Siemens Dimension & Vista | | 4.40 | 4.4 | 0.20 | 4.5 | 0.12 | 4.2 | 4.6 | 3 |
| Siemens Immulite 1000, 2000, 2500 | | 4.22 | 4.3 | 0.19 | 4.6 | 0.11 | 4.0 | 4.3 | 3 |
| Tosoh AIA | | 5.90 | 5.9 | - | - | - | - | - | 1 |
| All | | 4.53 | 4.6 | 0.22 | 4.8 | 0.02 | 3.4 | 5.9 | 82 |
| Gastrin, pmol/L | | | | | | | | | |
| Siemens Immulite 1000, 2000, 2500 | | 9.9 | 10 | 1.0 | 10.0 | 0.4 | 8 | 11 | 8 |
| All | | 9.9 | 10 | 1.1 | 11.4 | 0.4 | 8 | 11 | 8 |
| GH, µg/l | | | | | | | | | |
| DiaSorin Liaison | | 1.71 | 1.7 | 0.08 | 4.8 | 0.05 | 1.6 | 1.8 | 3 |
| IDS-iSYS | | 1.82 | 1.8 | 0.02 | 1.2 | 0.02 | 1.8 | 1.8 | 2 |
| Roche cobas e, Elecsys, & Modular E | | 2.17 | 2.2 | 0.06 | 2.6 | 0.02 | 2.1 | 2.3 | 9 |
| Siemens Immulite 1000, 2000, 2500 | | 1.92 | 2.0 | 0.15 | 7.9 | 0.05 | 1.7 | 2.1 | 13 |
| Tosoh AIA | | 1.64 | 1.6 | - | - | - | - | - | 1 |
| All | | 1.96 | 2.0 | 0.22 | 11.2 | 0.04 | 1.6 | 2.3 | 28 |
| IGF-1, nmol/L | | | | | | | | | |
| DiaSorin Liaison | | 23.51 | 23.5 | 1.55 | 6.6 | 0.69 | 21.8 | 26.0 | 5 |
| DIAsource RIA | | 17.59 | 17.6 | - | - | - | - | - | 1 |
| IDS-iSYS | | 21.88 | 21.9 | - | - | - | - | - | 1 |
| Roche cobas e, Elecsys, & Modular E | | 17.21 | 17.4 | 0.61 | 3.5 | 0.25 | 16.5 | 17.8 | 6 |
| Siemens IGF-I Restandardized | | 17.54 | 17.5 | 1.54 | 8.8 | 0.52 | 15.0 | 20.7 | 14 |
| All | | 18.54 | 17.7 | 2.63 | 14.2 | 0.51 | 15.0 | 26.0 | 27 |
| Insulin, mU/L | | | | | | | | | |
| Abbott Alinity | | 6.1 | 6 | 0.3 | 4.2 | 0.1 | 6 | 6 | 10 |
| Abbott Architect | | 7.6 | 8 | 0.5 | 6.7 | 0.4 | 7 | 8 | 2 |
| DiaSorin Liaison | | 10.4 | 10 | 0.1 | 1.4 | 0.1 | 10 | 11 | 2 |
| Roche cobas e, Elecsys, & Modular E | | 9.1 | 9 | 0.5 | 4.9 | 0.1 | 8 | 13 | 25 |
| Siemens Advia Centaur & Atellica | | 7.5 | 8 | 0.4 | 5.8 | 0.3 | 7 | 8 | 2 |
| Siemens Immulite 1000, 2000, 2500 | | 5.0 | 5 | - | - | - | - | - | 1 |
| Tosoh AIA | | 6.5 | 7 | - | - | - | - | - | 1 |
| All | | 8.2 | 9 | 1.7 | 20.6 | 0.3 | 5 | 13 | 43 |
| LH, U/L | | | | | | | | | |
| Abbott Alinity | | 2.71 | 2.7 | 0.12 | 4.3 | 0.03 | 2.5 | 3.5 | 23 |
| Abbott Architect | | 2.88 | 2.9 | 0.15 | 5.0 | 0.07 | 2.7 | 3.1 | 5 |
| Roche cobas e, Elecsys, & Modular E | | 4.69 | 4.7 | 0.23 | 4.9 | 0.05 | 3.6 | 5.1 | 34 |
| Siemens Advia Centaur & Atellica | | 3.34 | 3.4 | 0.18 | 5.3 | 0.05 | 3.0 | 3.5 | 11 |
| Siemens Dimension & Vista | | 3.37 | 3.3 | 0.12 | 3.4 | 0.07 | 3.3 | 3.5 | 3 |
| Siemens Immulite 1000, 2000, 2500 | | 3.90 | 4.0 | 0.16 | 4.0 | 0.09 | 3.7 | 4.0 | 3 |
| Tosoh AIA | | 2.80 | 2.8 | - | - | - | - | - | 1 |
| All | | 3.72 | 3.5 | 1.01 | 27.2 | 0.11 | 2.5 | 5.1 | 80 |
| Prog, 17-OH, nmol/L | | | | | | | | | |
| Demeditec Diagnostics EIA | | 5.40 | 5.4 | - | - | - | - | - | 1 |
| DIAsource RIA | | 4.10 | 4.1 | - | - | - | - | - | 1 |
| DRG Diagnostics EIA | | 3.46 | 3.8 | 1.60 | 46.4 | 0.93 | 1.7 | 4.9 | 3 |
| IBL ELISA | | 3.60 | 3.6 | - | - | - | - | - | 1 |

NUMERICAL SUMMARY
Hormone Determinations B 2023/01, Sample B1

| Analyte | Method group | x | med | s | CV% | u | Min | Max | Number |
|-----------------------------|-------------------------------------|--------------|-------------|--------------|--------------|--------------|-------------|-------------|-----------|
| Prog, 17-OH, nmol/L | | | | | | | | | |
| | Mass spectrometry | 2.66 | 2.7 | 0.06 | 2.4 | 0.04 | 2.6 | 2.7 | 2 |
| | Novatec EIA | 3.03 | 3.0 | - | - | - | - | - | 1 |
| | Siemens Dimension | 2.25 | 2.3 | - | - | - | - | - | 1 |
| | Snibe Diagnostic Maglumi | 3.30 | 3.3 | - | - | - | - | - | 1 |
| | All | 3.38 | 3.3 | 1.22 | 36.0 | 0.37 | 1.7 | 5.4 | 11 |
| Progesterone, nmol/L | | | | | | | | | |
| | Abbott Alinity | 2.68 | 2.6 | 0.37 | 13.8 | 0.12 | 2.2 | 3.5 | 15 |
| | Abbott Architect | 2.65 | 2.7 | 0.25 | 9.4 | 0.11 | 2.2 | 2.9 | 5 |
| | Roche cobas e, Elecsys, & Modular E | 3.05 | 3.0 | 0.28 | 9.2 | 0.07 | 2.6 | 3.6 | 25 |
| | Siemens Advia Centaur & Atellica | 2.83 | 2.8 | 0.25 | 8.7 | 0.09 | 2.5 | 3.3 | 12 |
| | Siemens Dimension & Vista | 5.20 | 5.2 | 0.28 | 5.4 | 0.20 | 5.0 | 5.4 | 2 |
| | Siemens Immulite 1000, 2000, 2500 | 3.53 | 3.5 | 0.04 | 1.3 | 0.03 | 3.5 | 3.6 | 2 |
| | Tosoh AIA | 3.37 | 3.4 | - | - | - | - | - | 1 |
| | All | 2.93 | 2.9 | 0.38 | 13.1 | 0.05 | 2.2 | 5.4 | 62 |
| Prolactin, mU/L | | | | | | | | | |
| | Abbott Alinity | 177.4 | 177 | 7.4 | 4.2 | 1.9 | 167 | 197 | 24 |
| | Abbott Architect | 172.3 | 171 | 10.2 | 5.9 | 4.5 | 160 | 183 | 5 |
| | Brahms Kryptor | 126.0 | 126 | 1.8 | 1.4 | 0.9 | 124 | 128 | 4 |
| | Roche cobas e, Elecsys, & Modular E | 179.4 | 180 | 7.5 | 4.2 | 1.6 | 160 | 198 | 33 |
| | Siemens Advia Centaur & Atellica | 130.0 | 127 | 7.4 | 5.7 | 2.6 | 121 | 141 | 13 |
| | Siemens Dimension & Vista | 140.0 | 140 | 0.0 | 0.0 | 0.0 | 140 | 140 | 3 |
| | Siemens Immulite 1000, 2000, 2500 | 149.7 | 152 | 4.3 | 2.9 | 2.5 | 145 | 152 | 3 |
| | Tosoh AIA | 200.0 | 200 | - | - | - | - | - | 1 |
| | All | 167.6 | 174 | 22.6 | 13.5 | 2.4 | 121 | 200 | 86 |
| SHBG, nmol/L | | | | | | | | | |
| | Abbott Alinity | 27.4 | 28 | 1.8 | 6.4 | 0.6 | 24 | 30 | 15 |
| | Abbott Architect | 27.6 | 27 | 2.2 | 7.8 | 1.0 | 26 | 30 | 5 |
| | Roche cobas e, Elecsys, & Modular E | 26.2 | 26 | 1.0 | 3.8 | 0.2 | 24 | 28 | 29 |
| | Siemens Advia Centaur & Atellica | 26.3 | 27 | 1.8 | 6.7 | 0.6 | 24 | 30 | 9 |
| | Siemens Immulite 1000, 2000, 2500 | 22.1 | 22 | - | - | - | - | - | 1 |
| | Tosoh AIA | 29.1 | 29 | - | - | - | - | - | 1 |
| | All | 26.6 | 27 | 1.6 | 5.9 | 0.2 | 22 | 30 | 60 |
| Testo, free, pmol/L | | | | | | | | | |
| | Abbott Alinity | 630.0 | 630 | - | - | - | - | - | 1 |
| | Calculated | 393.0 | 393 | - | - | - | - | - | 1 |
| | DRG Diagnostics EIA | 65.9 | 66 | - | - | - | - | - | 1 |
| | IDS-iSYS | 51.2 | 51 | - | - | - | - | - | 1 |
| | Novatec EIA | 125.3 | 125 | - | - | - | - | - | 1 |
| | All | 253.1 | 125 | 285.6 | 112.9 | 127.7 | 51 | 630 | 5 |
| Testosterone, nmol/L | | | | | | | | | |
| | Abbott Alinity | 25.05 | 25.1 | 1.32 | 5.3 | 0.37 | 23.2 | 27.9 | 20 |
| | Abbott Architect | 24.08 | 24.6 | 1.40 | 5.8 | 0.57 | 21.9 | 25.3 | 6 |
| | Mass spectrometry | 23.03 | 23.0 | 1.41 | 6.1 | 0.63 | 21.2 | 24.5 | 5 |
| | Roche cobas e, Elecsys, & Modular E | 23.83 | 23.7 | 0.86 | 3.6 | 0.20 | 22.7 | 39.3 | 28 |
| | Siemens Advia Centaur & Atellica | 25.16 | 25.0 | 2.01 | 8.0 | 0.67 | 22.0 | 28.5 | 9 |
| | Siemens Immulite 1000, 2000, 2500 | 23.28 | 23.3 | - | - | - | - | - | 1 |
| | Tosoh AIA | 27.50 | 27.5 | - | - | - | - | - | 1 |
| | All | 24.37 | 24.3 | 1.45 | 5.9 | 0.17 | 21.2 | 39.3 | 70 |

Sample B2
Aldosterone, pmol/L

NUMERICAL SUMMARY
Hormone Determinations B 2023/01, Sample B2

| Analyte | Method group | x | med | s | CV% | u | Min | Max | Number |
|--------------------------------|-------------------------------------|----------------|--------------|--------------|-------------|--------------|--------------|--------------|-----------|
| Aldosterone, pmol/L | | | | | | | | | |
| | Beckman Coulter Immunotech | 1,103.5 | 1,104 | 106.8 | 9.7 | 75.5 | 1,028 | 1,179 | 2 |
| | DiaSorin Liaison | 1,121.9 | 1,090 | 72.6 | 6.5 | 24.2 | 1,040 | 1,222 | 9 |
| | All | 1,118.6 | 1,090 | 83.4 | 7.5 | 25.1 | 1,028 | 1,222 | 11 |
| Androstenedione, nmol/L | | | | | | | | | |
| | DiaSorin Liaison | 21.51 | 22.1 | 1.42 | 6.6 | 0.82 | 19.9 | 22.5 | 3 |
| | DRG Diagnostics EIA | 17.62 | 17.6 | 0.64 | 3.6 | 0.45 | 17.2 | 18.1 | 2 |
| | IBL ELISA | 29.18 | 29.2 | - | - | - | - | - | 1 |
| | Mass spectrometry | 18.10 | 18.1 | 2.40 | 13.3 | 1.70 | 16.4 | 19.8 | 2 |
| | Roche cobas e, Elecsys, & Modular E | 16.31 | 16.0 | 0.68 | 4.2 | 0.40 | 15.8 | 17.1 | 3 |
| | Siemens Immulite 1000, 2000, 2500 | 36.81 | 36.8 | 13.59 | 36.9 | 9.61 | 27.2 | 46.4 | 2 |
| | Snibe Diagnostic Maglumi | 17.83 | 17.8 | - | - | - | - | - | 1 |
| | All | 20.43 | 18.9 | 5.08 | 24.9 | 1.36 | 15.8 | 46.4 | 14 |
| Cortisol, nmol/l | | | | | | | | | |
| | Abbott Alinity | 606.9 | 604 | 29.2 | 4.8 | 7.3 | 545 | 647 | 25 |
| | Abbott Architect | 606.0 | 604 | 13.0 | 2.1 | 7.5 | 594 | 620 | 3 |
| | DiaSorin Liaison | 626.1 | 626 | - | - | - | - | - | 1 |
| | Mass spectrometry | 665.0 | 665 | - | - | - | - | - | 1 |
| | Roche cobas e, Elecsys, & Modular E | 703.7 | 701 | 35.1 | 5.0 | 7.2 | 618 | 790 | 37 |
| | Siemens Advia Centaur & Atellica | 810.6 | 781 | 91.8 | 11.3 | 32.4 | 715 | 1,010 | 8 |
| | Siemens Immulite 1000, 2000, 2500 | 608.3 | 608 | 21.5 | 3.5 | 15.2 | 593 | 624 | 2 |
| | Tosoh AIA | 757.5 | 758 | - | - | - | - | - | 1 |
| | All | 672.6 | 679 | 72.6 | 10.8 | 8.2 | 545 | 1,010 | 78 |
| C-peptide, nmol/L | | | | | | | | | |
| | Abbott Alinity | 1.098 | 1.10 | 0.057 | 5.2 | 0.021 | 1.03 | 1.21 | 12 |
| | Abbott Architect | 1.133 | 1.13 | 0.010 | 0.9 | 0.007 | 1.13 | 1.14 | 2 |
| | DiaSorin Liaison | 1.430 | 1.43 | - | - | - | - | - | 1 |
| | Roche cobas e, Elecsys, & Modular E | 1.249 | 1.25 | 0.043 | 3.5 | 0.010 | 1.15 | 1.40 | 29 |
| | Siemens Advia Centaur & Atellica | 1.062 | 1.04 | 0.067 | 6.3 | 0.025 | 0.99 | 1.18 | 7 |
| | Siemens Immulite 1000, 2000, 2500 | 1.112 | 1.11 | 0.098 | 8.8 | 0.069 | 1.04 | 1.18 | 2 |
| | Tosoh AIA | 1.010 | 1.01 | - | - | - | - | - | 1 |
| | All | 1.181 | 1.21 | 0.111 | 9.4 | 0.015 | 0.99 | 1.43 | 54 |
| DHEAS, µmol/L | | | | | | | | | |
| | Abbott Alinity | 23.06 | 23.0 | 1.04 | 4.5 | 0.37 | 21.0 | 24.5 | 8 |
| | Abbott Architect | 22.76 | 22.8 | 1.63 | 7.2 | 1.15 | 21.6 | 23.9 | 2 |
| | Roche cobas e, Elecsys, & Modular E | 26.67 | 26.2 | 1.50 | 5.6 | 0.50 | 25.1 | 33.9 | 14 |
| | Siemens Advia Centaur & Atellica | 18.40 | 18.4 | - | - | - | - | - | 1 |
| | Siemens Immulite 1000, 2000, 2500 | 20.52 | 20.4 | 1.52 | 7.4 | 0.76 | 19.1 | 22.3 | 4 |
| | Tosoh AIA | 26.20 | 26.2 | - | - | - | - | - | 1 |
| | All | 24.43 | 24.8 | 2.99 | 12.2 | 0.55 | 18.4 | 33.9 | 30 |
| Estradiol, nmol/L | | | | | | | | | |
| | Abbott Alinity | 1.243 | 1.25 | 0.051 | 4.1 | 0.015 | 1.12 | 1.33 | 19 |
| | Abbott Architect | 1.268 | 1.27 | 0.055 | 4.3 | 0.025 | 1.20 | 1.35 | 5 |
| | Mass spectrometry | 1.400 | 1.40 | - | - | - | - | - | 1 |
| | Roche cobas e, Elecsys, & Modular E | 1.557 | 1.55 | 0.056 | 3.6 | 0.012 | 1.45 | 1.75 | 34 |
| | Siemens Advia Centaur & Atellica | 1.504 | 1.53 | 0.074 | 4.9 | 0.025 | 1.42 | 1.63 | 9 |
| | Siemens Dimension & Vista | 1.202 | 1.20 | 0.002 | 0.2 | 0.002 | 1.20 | 1.20 | 2 |
| | Siemens Immulite 1000, 2000, 2500 | 1.130 | 1.13 | - | - | - | - | - | 1 |
| | Tosoh AIA | 1.461 | 1.46 | - | - | - | - | - | 1 |
| | All | 1.429 | 1.49 | 0.179 | 12.5 | 0.021 | 1.12 | 1.75 | 72 |
| FSH, U/L | | | | | | | | | |

NUMERICAL SUMMARY
Hormone Determinations B 2023/01, Sample B2

| Analyte | Method group | x | med | s | CV% | u | Min | Max | Number |
|----------------------------|-------------------------------------|--------------|-------------|-------------|-------------|-------------|-------------|-------------|-----------|
| FSH, U/L | | | | | | | | | |
| | Abbott Alinity | 36.18 | 36.1 | 1.76 | 4.9 | 0.44 | 32.5 | 38.8 | 25 |
| | Abbott Architect | 36.26 | 35.8 | 3.65 | 10.1 | 1.63 | 30.8 | 40.0 | 5 |
| | Roche cobas e, Elecsys, & Modular E | 36.11 | 36.2 | 1.61 | 4.5 | 0.33 | 32.1 | 40.2 | 36 |
| | Siemens Advia Centaur & Atellica | 40.19 | 41.4 | 4.29 | 10.7 | 1.36 | 28.6 | 43.6 | 10 |
| | Siemens Dimension & Vista | 34.00 | 33.9 | 0.95 | 2.8 | 0.55 | 33.1 | 35.0 | 3 |
| | Siemens Immulite 1000, 2000, 2500 | 31.57 | 31.1 | 1.46 | 4.6 | 0.84 | 30.4 | 33.2 | 3 |
| | Tosoh AIA | 40.10 | 40.1 | - | - | - | - | - | 1 |
| | All | 36.36 | 36.2 | 2.55 | 7.0 | 0.28 | 28.6 | 43.6 | 83 |
| Gastrin, pmol/L | | | | | | | | | |
| | Siemens Immulite 1000, 2000, 2500 | 10.7 | 11 | 1.0 | 9.4 | 0.4 | 9 | 12 | 8 |
| | All | 10.8 | 11 | 1.0 | 9.5 | 0.4 | 9 | 12 | 8 |
| GH, µg/l | | | | | | | | | |
| | DiaSorin Liaison | 8.00 | 7.9 | 0.37 | 4.6 | 0.21 | 7.7 | 8.4 | 3 |
| | IDS-iSYS | 8.19 | 8.2 | 0.27 | 3.3 | 0.19 | 8.0 | 8.4 | 2 |
| | Roche cobas e, Elecsys, & Modular E | 9.10 | 9.0 | 0.44 | 4.9 | 0.15 | 8.6 | 10.1 | 9 |
| | Siemens Immulite 1000, 2000, 2500 | 8.79 | 8.7 | 0.63 | 7.2 | 0.22 | 8.0 | 10.5 | 13 |
| | Tosoh AIA | 9.10 | 9.1 | - | - | - | - | - | 1 |
| | All | 8.76 | 8.8 | 0.61 | 7.0 | 0.12 | 7.7 | 10.5 | 28 |
| IGF-1, nmol/L | | | | | | | | | |
| | DiaSorin Liaison | 22.29 | 22.5 | 1.64 | 7.4 | 0.73 | 20.2 | 24.4 | 5 |
| | DIAsource RIA | 20.80 | 20.8 | - | - | - | - | - | 1 |
| | IDS-iSYS | 19.91 | 19.9 | - | - | - | - | - | 1 |
| | Roche cobas e, Elecsys, & Modular E | 15.91 | 15.7 | 1.76 | 11.0 | 0.72 | 14.0 | 19.1 | 6 |
| | Siemens IGF-I Restandardized | 16.88 | 16.8 | 1.29 | 7.6 | 0.43 | 15.0 | 20.0 | 14 |
| | All | 17.81 | 17.0 | 2.73 | 15.3 | 0.52 | 14.0 | 24.4 | 27 |
| Insulin, mU/L | | | | | | | | | |
| | Abbott Alinity | 89.2 | 90 | 5.1 | 5.7 | 1.6 | 82 | 97 | 10 |
| | Abbott Architect | 97.8 | 98 | 5.2 | 5.4 | 3.7 | 94 | 102 | 2 |
| | DiaSorin Liaison | 152.3 | 152 | 6.2 | 4.1 | 4.4 | 148 | 157 | 2 |
| | Roche cobas e, Elecsys, & Modular E | 131.2 | 130 | 5.2 | 3.9 | 1.3 | 121 | 186 | 25 |
| | Siemens Advia Centaur & Atellica | 108.2 | 108 | 6.6 | 6.1 | 4.7 | 104 | 113 | 2 |
| | Siemens Immulite 1000, 2000, 2500 | 31.5 | 32 | - | - | - | - | - | 1 |
| | Tosoh AIA | 117.6 | 118 | - | - | - | - | - | 1 |
| | All | 118.7 | 127 | 24.0 | 20.2 | 3.7 | 32 | 186 | 43 |
| LH, U/L | | | | | | | | | |
| | Abbott Alinity | 36.73 | 36.7 | 1.15 | 3.1 | 0.31 | 35.0 | 40.4 | 22 |
| | Abbott Architect | 38.83 | 39.2 | 1.44 | 3.7 | 0.64 | 37.3 | 40.8 | 5 |
| | Roche cobas e, Elecsys, & Modular E | 50.81 | 51.0 | 2.22 | 4.4 | 0.48 | 44.7 | 54.1 | 34 |
| | Siemens Advia Centaur & Atellica | 48.12 | 48.0 | 2.13 | 4.4 | 0.64 | 44.9 | 51.0 | 11 |
| | Siemens Dimension & Vista | 56.53 | 56.4 | 1.60 | 2.8 | 0.93 | 55.0 | 58.2 | 3 |
| | Siemens Immulite 1000, 2000, 2500 | 51.72 | 50.7 | 4.61 | 8.9 | 2.66 | 47.7 | 56.8 | 3 |
| | Tosoh AIA | 46.70 | 46.7 | - | - | - | - | - | 1 |
| | All | 45.91 | 47.7 | 7.72 | 16.8 | 0.87 | 35.0 | 58.2 | 79 |
| Prog, 17-OH, nmol/L | | | | | | | | | |
| | Demeditec Diagnostics EIA | 19.70 | 19.7 | - | - | - | - | - | 1 |
| | DIAsource RIA | 17.45 | 17.5 | - | - | - | - | - | 1 |
| | DRG Diagnostics EIA | 16.38 | 18.4 | 4.73 | 28.9 | 2.73 | 11.0 | 19.8 | 3 |
| | IBL ELISA | 15.00 | 15.0 | - | - | - | - | - | 1 |
| | Mass spectrometry | 14.31 | 14.3 | 2.40 | 16.8 | 1.70 | 12.6 | 16.0 | 2 |
| | Novatec EIA | 15.21 | 15.2 | - | - | - | - | - | 1 |

NUMERICAL SUMMARY
Hormone Determinations B 2023/01, Sample B2

| Analyte | Method group | x | med | s | CV% | u | Min | Max | Number |
|-----------------------------|-------------------------------------|--------------|-------------|--------------|--------------|-------------|-------------|-------------|-----------|
| Prog, 17-OH, nmol/L | | | | | | | | | |
| | Siemens Dimension | 14.06 | 14.1 | - | - | - | - | - | 1 |
| | Snibe Diagnostic Maglumi | 14.10 | 14.1 | - | - | - | - | - | 1 |
| | All | 15.75 | 15.2 | 3.22 | 20.5 | 0.97 | 11.0 | 19.8 | 11 |
| Progesterone, nmol/L | | | | | | | | | |
| | Abbott Alinity | 55.34 | 55.0 | 3.68 | 6.7 | 1.19 | 48.3 | 72.7 | 15 |
| | Abbott Architect | 56.57 | 57.2 | 3.99 | 7.1 | 1.79 | 50.6 | 61.5 | 5 |
| | Roche cobas e, Elecsys, & Modular E | 65.26 | 65.3 | 2.29 | 3.5 | 0.57 | 58.9 | 73.6 | 25 |
| | Siemens Advia Centaur & Atellica | 43.38 | 42.9 | 1.37 | 3.2 | 0.41 | 41.9 | 46.0 | 11 |
| | Siemens Dimension & Vista | 54.10 | 54.1 | 0.14 | 0.3 | 0.10 | 54.0 | 54.2 | 2 |
| | Siemens Immulite 1000, 2000, 2500 | 46.27 | 46.3 | 0.22 | 0.5 | 0.16 | 46.1 | 46.4 | 2 |
| | Tosoh AIA | 75.50 | 75.5 | - | - | - | - | - | 1 |
| | All | 57.52 | 58.0 | 10.35 | 18.0 | 1.33 | 41.9 | 75.5 | 61 |
| Prolactin, mU/L | | | | | | | | | |
| | Abbott Alinity | 666.0 | 672 | 23.5 | 3.5 | 6.0 | 634 | 734 | 24 |
| | Abbott Architect | 651.1 | 662 | 32.0 | 4.9 | 14.3 | 610 | 688 | 5 |
| | Brahms Kryptor | 546.0 | 540 | 14.3 | 2.6 | 7.1 | 537 | 567 | 4 |
| | Roche cobas e, Elecsys, & Modular E | 673.8 | 673 | 27.5 | 4.1 | 6.0 | 618 | 744 | 33 |
| | Siemens Advia Centaur & Atellica | 508.0 | 490 | 35.6 | 7.0 | 12.9 | 475 | 571 | 12 |
| | Siemens Dimension & Vista | 593.3 | 590 | 5.8 | 1.0 | 3.3 | 590 | 600 | 3 |
| | Siemens Immulite 1000, 2000, 2500 | 612.3 | 607 | 32.1 | 5.2 | 18.5 | 583 | 647 | 3 |
| | Tosoh AIA | 718.0 | 718 | - | - | - | - | - | 1 |
| | All | 644.9 | 658 | 57.5 | 8.9 | 6.2 | 475 | 744 | 85 |
| SHBG, nmol/L | | | | | | | | | |
| | Abbott Alinity | 52.9 | 53 | 2.7 | 5.2 | 0.9 | 45 | 60 | 15 |
| | Abbott Architect | 50.0 | 51 | 3.2 | 6.4 | 1.4 | 47 | 54 | 5 |
| | Roche cobas e, Elecsys, & Modular E | 50.6 | 50 | 2.4 | 4.8 | 0.6 | 47 | 56 | 29 |
| | Siemens Advia Centaur & Atellica | 52.6 | 52 | 3.8 | 7.3 | 1.3 | 47 | 61 | 9 |
| | Siemens Immulite 1000, 2000, 2500 | 43.7 | 44 | - | - | - | - | - | 1 |
| | Tosoh AIA | 58.3 | 58 | - | - | - | - | - | 1 |
| | All | 51.3 | 51 | 3.1 | 6.0 | 0.4 | 44 | 61 | 60 |
| Testo, free, pmol/L | | | | | | | | | |
| | Abbott Alinity | 140.0 | 140 | - | - | - | - | - | 1 |
| | Calculated | 124.9 | 125 | - | - | - | - | - | 1 |
| | DRG Diagnostics EIA | 27.2 | 27 | - | - | - | - | - | 1 |
| | IDS-iSYS | 25.0 | 25 | - | - | - | - | - | 1 |
| | Novatec EIA | 26.4 | 26 | - | - | - | - | - | 1 |
| | Roche cobas e, Elecsys, & Modular E | 3.6 | 4 | - | - | - | - | - | 1 |
| | All | 57.8 | 27 | 66.5 | 115.0 | 27.1 | 4 | 140 | 6 |
| Testosterone, nmol/L | | | | | | | | | |
| | Abbott Alinity | 9.75 | 9.8 | 0.42 | 4.3 | 0.12 | 9.1 | 10.8 | 20 |
| | Abbott Architect | 9.66 | 9.7 | 0.28 | 2.9 | 0.11 | 9.2 | 10.0 | 6 |
| | Mass spectrometry | 9.88 | 10.0 | 0.62 | 6.3 | 0.28 | 9.1 | 10.5 | 5 |
| | Roche cobas e, Elecsys, & Modular E | 9.99 | 10.0 | 0.33 | 3.3 | 0.08 | 9.2 | 11.0 | 28 |
| | Siemens Advia Centaur & Atellica | 7.65 | 7.8 | 0.73 | 9.5 | 0.24 | 6.0 | 8.5 | 9 |
| | Siemens Immulite 1000, 2000, 2500 | 10.90 | 10.9 | - | - | - | - | - | 1 |
| | Tosoh AIA | 13.40 | 13.4 | - | - | - | - | - | 1 |
| | All | 9.76 | 9.8 | 0.60 | 6.2 | 0.07 | 6.0 | 13.4 | 70 |

NUMERICAL SUMMARY

Hormone Determinations B 2023/01, Sample B2

| Analyte | Method group | x | med | s | CV% | u | Min | Max | Number |
|----------------------|--------------|---|-----|---|-----|---|-----|-----|--------|
| Testosterone, nmol/L | | | | | | | | | |

Participants

108 participants from 17 countries.

Report info

Assigned value (target value) calculation and its uncertainty

Your own result should be compared to others using the same method.

The assigned values (X_{rob}) are calculated according to the robust procedure described in the standard ISO 13528 (Statistical methods for use in proficiency testing by interlaboratory comparisons, Annex C, Algorithm A).

The standard uncertainty of the assigned value is expressed as 1.25 x the standard error of mean (SEM) and marked as "u" in numerical summary. Due to its iterative mode algorithm A adds the uncertainty of the assigned value and with this factor we want to adjust uncertainty accordingly.

In case there are 2-12 results in a method group, the robust calculation is not used but assigned values (X_{pt}) are means of the results where results deviating more than ± 3 * standard deviation from the median are removed. The standard uncertainty (u) of the assigned value is reported as standard error of the mean (SEM). Additionally, if the measurement uncertainty of the target value is too large ($u > 0.1$ * maximum allowable error) an automatic text is printed on the report: "The uncertainty of the assigned value is not negligible, and evaluations could be affected."

Please notice also that for groups that have only 1 result only the client's own result is shown. No target value (except for reference method values) is calculated, no target areas are shown.

Z score

In case there are 2-5 results in a method group, no z-score is calculated, and a text is printed on the report: "Due to the small number of results, the z score is not calculated." In case there are 6-12 results, the report has a text: "Z score is uncertain due to the small number of observations."

Results reported with $< \text{tai} >$ -signs cannot be included in the statistics.

For information on report interpretation and performance evaluation, please see the "EQAS Interpretation guidelines" in LabScala User instructions (top right corner ?Help link).

External Quality Assessment Scheme

Hormones B Steroid and peptide hormones Round 1, 2023

Specimens

Sample S001/B1 (LQ732423011) and sample S002/B2 (LQ732423012) were lyophilized human serum samples.

Based on the previous tests and the results of this round, the samples are homogeneous, stable and suitable for the external quality assessment scheme.

The materials were sent without temperature control packaging.

Report info

Please see the description of the data analysis on the last page of the laboratory-specific histogram and Global report. It is important to read the Final report first, because it contains important information of the samples and results in each round.

The samples have been used before on the round. The sample history can be seen in the client specific report with the histogram.

Comments – EQA Coordinator

In C-peptide, prolactin and progesterone, Siemens Advia Centaur & Atellica's level is the lowest in both samples, while Roche's level is the highest. In LH, in both samples, Abbott Alinity has the lowest level, while Roche has the highest level. The insulin level in Siemens Immulite is significantly lower than the other groups in sample S002/B2. The results of free testosterone with different methods differ from each other, just like in previous rounds.

Comments – Expert

The mean androstenedione concentration in sample B1 was in the adult reference range ($\bar{x} = 4.22$ nmol/L, $n = 15$). The range of results of different methods was large, and the overall variation of the results was large (CV% = 27.7%). The consensus mean of sample B2 was very high and clearly above the adult reference range ($\bar{x} = 20.43$ nmol/L, $n = 14$). Also in this sample, the overall variation of the results was very large (CV% = 24.9%).

The consensus mean of aldosterone in sample B1 was in the healthy reference range ($\bar{x} = 218.9$ pmol/L, $n = 11$) and the overall variation in sample B1 was large (B1: CV% = 19.7%). The consensus mean of sample B2 in all participants was well above the adult reference range ($\bar{x} = 1118.6$ pmol/L, $n = 11$). The overall variation in sample B2 was good (B2: CV% = 7.5%).

In DHEAS, the mean concentration of sample B1 was 7.35 μ mol/L, which was in the reference range for young adult men and women. In sample B2, on the other hand, the concentration was strongly elevated (24.43 μ mol/L). The overall variations in both samples were moderate (B1: CV% = 12.9%, $n = 31$, and B2: CV% = 12.2%, $n = 30$).

In estradiol, sample (B1) had estradiol levels of the female follicular/luteal phase for all methods ($\bar{x} = 0.313$ nmol/L, $n = 71$) and sample B2 had a high level of estradiol for ovulation ($\bar{x} = 1.429$ nmol/L, $n = 72$). The overall variation in sample B1 was good (CV% = 8.1%), but in sample B2 it was moderate (CV% = 12.5%). Siemens Immulite gave a lower result level in sample B1 than other methods.

2023-03-13

FINAL REPORT

Product no. 2301, 2301S

| | |
|--------------|------------|
| Samples sent | 2023-02-07 |
| Round closed | 2023-02-27 |
| Final report | 2023-03-13 |

Request for correction

Typing errors in laboratory's result forms are on laboratory's responsibility. Labquality accepts responsibility only for result processing. Requests must be notified by writing within three weeks from the date of this letter.

Authorized by

EQA Coordinator
Päivi Ranta
paivi.ranta@labquality.fi

Expert

Anna Becker, Lic.Sc,
Clinical chemist, HUSLAB,
Helsinki

Labquality Oy

Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200
Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.com



Mean cortisol results represented normal adult concentrations in sample B1. DiaSorin Liaison gave a lower result level at a low concentration than other methods. The results were slightly elevated in sample B2 compared to the reference range for adult cortisol. (B1: $x = 262.3$ nmol/L, $n = 78$ and B2: $x = 672.6$ nmol/L, $n = 78$). The overall variations of the methods and results were good and moderate (B1: CV% = 6.8%, B2: CV% = 10.8%).

The mean concentrations of progesterone in sample B1 corresponded high follicular phase concentration (B1: $x = 2.93$ nmol/L), and the overall variation of the methods was moderately large (B1: $x = 13.1\%$). Siemens Dimension & Vista gave a higher result level than other methods (5.20 nmol/l). In sample B2 the concentration of progesterone was high and corresponded to the women's luteal phase (B2: $x = 57.52$ nmol/L). All participants reported high luteal phase concentrations, but TOSOH AIA gave a higher result level than other methods (75.50 nmol/l). The overall variation of the methods was large (B2: CV% = 18.0 %).

The concentration of 17OH-progesterone in sample B1 corresponded to the normal level of men ($x = 3.38$ nmol/L, $n = 11$). The overall variation of the methods was large, and the concentration varied largely between the groups (2.25 – 5.40 nmol/L, CV%=36.0 %). The concentration in sample B2 was elevated and clearly above the reference values ($x = 15.75$ nmol/L, $n = 11$). The overall variation of the methods was large (B2: CV% = 20.5 %).

In testosterone, the mean concentration of sample B1 was in the normal male concentration ($x = 24.37$ nmol/L, $n = 70$). The concentration of sample B2 corresponded to the lower limit of men's reference range ($x = 9.76$ nmol/L, $n = 70$). The Siemens Advia-Atellica user group received a lower mean compared to other groups. In both samples, the overall variation of the methods was good (B1: CV% = 5.9% and B2: CV% = 6.2%).

Free testosterone was reported by only five to six (5-6) laboratories. Because different methods used measured free testosterone levels differently, the concentrations of the samples obtained varied by 12-35 -fold and the total variations from the samples was consistent (CV% t were B1: 112.9% and B2: 115.0%). It is therefore not appropriate to compare the results with the different methods. The participants of the round can compare their own results with other laboratories that mainly use a similar method. In general, the testosterone and SHBG levels measured by reliable methods and the calculated free testosterone levels determined by them correlate with the results and result levels of known reference methods based on ultrafiltration or equilibrium dialysis.

End of report