

General Clinical Chemistry, Daytrol Rounds 1-12, 2023

Store this instruction, as it applies to the whole year of 2023.

Specimen

Please find enclosed lyophilized human serum samples, Daytrol (DT23) each 5 mL, according to your order for the whole year of 2022. They can be used from the 1st January 2023 onwards.

Caution

Quality control specimens derived from human blood must be handled with the same care as patient samples, i.e. as potential transmitters of serious diseases. The specimens are found to be HBsAg, HCVAb and HIVAgAb negative when tested with licensed reagents, but no known test method can offer complete assurance that the specimens will not transmit these or other infectious diseases.

Examinations

Please see page 3.

Storage and use

After arrival, store the unopened lyophilized samples at +2 ... 8 °C. Before analysis, open the vial carefully to prevent escape of dried material. Add 5 mL of high-grade room temperature laboratory water and immediately close the vial and let stand at room temperature, protected from light, for about half an hour until all the material is completely dissolved. Invert the vial several times to dissolve any material adhering to the stopper. Analyze as a patient sample.

After use, the remaining serum is divided, if necessary, into batches, which are transferred to + 2 ... 8 °C. After reconstitution, the sample is valid for 5 days at + 2 ... 8 °C. Take the serum sample to the room temperature shortly before use, so that it reaches room temperature before analysis.

Result reporting and methods

The round number equals the month in question. Report the results via LabScala (www.labscala.com) on the last day of each month.

Method information will be given in LabScala. The device, reagent, method, chemical principle, and unit are given in the device registry. The reagent manufacturer is selected as the method. We kindly ask you to verify that your methods in LabScala are correct. If necessary, contact the EQA Coordinator, if you cannot find the correct device or reagent in LabScala. An example of the view in the device registry is given on the page two of this instruction letter.

We kindly ask you to notice that we will no longer provide a separate target value table of the results, (which has been provided twice a year before). Laboratories should compare their results to the monthly delivered summary reports and to the assigned values found there. These results are the most accurate and current ones. Therefore, the assigned values and target areas for example of Round 1, January will be available in the beginning of February.

2022-12-13

INSTRUCTIONS

Product no. 1031
LQ720123011-121/UK

If the kit is incomplete or contains damaged specimens, please report immediately to info@labquality.fi

The results should be reported via LabScala no later than **on the last day of each month**.

Inquiries

EQA Coordinator
Liisa Ylilepsa
liisa.ylilepsa@labquality.fi

Labquality

Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200
Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.com



General Clinical Chemistry, Daytrol

Methods in Labscala

My Registry → My Devices

Modify device DEMO DEMO DEMO DEMO DEMO DEMO DEMO DEMO

Help

▼ Modify device

Client: 2399 Test customers

DEVICE

Instrument: Atellica CH 930	* Start date: 01.05.2019
Manufacturer: SIEMENS	Please note! The start date must be earlier or the same as the Measurement date on the Result form.
Nickname: *: Daytrol	Not used after:
Free text field for a device name	This field is not mandatory. The instrument cannot be used after the end-date.
	Location:
	Client system ID:
	LQ 3-digit code:
	Unique devicespecific code
	Integration device:

USED FOR MEASURING SCHEMES

Product	Active		
DayTrol, human serum	<input checked="" type="checkbox"/>		
Analyte	Definition	Active	
		All	<input type="button" value="Q"/>
ALT		<input type="checkbox"/>	
Alb	SIEMENS - ADVIA Chemistry Albumin BCP (ALBP); Siemens Atellica; Bromcresol purple (BCP); photometry; g/l	<input checked="" type="checkbox"/>	
ALP		<input type="checkbox"/>	
Amyl		<input type="checkbox"/>	
AST		<input type="checkbox"/>	
Bil, tot	SIEMENS - Atellica CH Total Bilirubin_2 (TBil_2); Siemens Atellica; Biliverdin, Na nitrite; photometry; µmol/l	<input checked="" type="checkbox"/>	
Phosphorus		<input type="checkbox"/>	
Glucose		<input type="checkbox"/>	
GT		<input type="checkbox"/>	
Potassium		<input type="checkbox"/>	

First Previous 1 2 3 4 Next Last

Add device's secondary information for test Alb DEMO DEMO

▼ Add device's reagent info for Analyte Alb

Nickname: Daytrol
Instrument: Atellica CH 930
Location:
Client system id:
* Reagent manufacturer: SIEMENS
* Reagent: ADVIA Chemistry Alb
* Chemical principle: Bromcresol purple (B)
* Method: Siemens Atellica
Calibrator:
* Measure unit: g/l
Methodics: 1072-037alb,037 Bromcresol purple



Back

General Clinical Chemistry, Daytrol

Examinations

Na	Pi	Fe
K	Uric acid	Transf
Cl	Chol	T4
Lactate	Trigly	GT
Crea	Bil	TSH
Urea	ALP	T4-free
Prot	AMYL	Chol-HDL
Alb	AST	OSMOL
Gluc	ALT	TfR
Ca	CK	Lithium
Mg	LD	

DayTrol | ALT | BTS 310




Coordinator:

	X_{pt}	$X_{pt} \pm 12\%$	Diff%	CV% target	CV50%	CV90%	n devices	n results
IFCC methods without P-5-P (ALT)	74	65 - 83	7.7	4.0	1.9	5.3	3	38
All methods	78	-	1.5	4.0	2.8	4.7	35	524

Round	X_o	$X_{pt} \pm 12\%$	sd	CV%	n results
22/2	69	61 - 78	2	2	11
22/3	70	62 - 79	3	5	11
22/4	69	61 - 78	3	4	11
22/5	68	62 - 79	2	3	11
22/6	68	61 - 78	1	2	8
22/7	69	61 - 78	1	2	11
22/8	68	62 - 79	2	3	10
22/9	70	62 - 79	2	3	9
22/10	70	61 - 78	2	4	11
22/11	69	61 - 77	1	2	9
22/12	70	60 - 77	1	2	12
23/1	80	65 - 83	4	5	10
22/2 - 23/1	70	-	2	3	124

Laboratory's analytical performance

January - Internal CV% (compared to all methodicses)





January - Diff% (bias) (compared to all methodicses)



7 / 10 of your results this month are within the target area of your method group.

DayTrol | AST | BTS 310



Coordinator:

	X_{pt}	$X_{pt} \pm 12\%$	Diff%	CV% target	CV50%	CV90%	n devices	n results
IFCC without P-5-P (AST)	76	67 - 85	-	4.0	4.1	4.1	1	10
All methods	78	-	-3.2	4.0	2.6	3.5	27	400


Round	X_o	$X_{pt} \pm 12\%$	sd	CV%	n results
22/2	60	55 - 70	2	4	11
22/3	62	58 - 73	1	2	10
22/4	62	59 - 75	2	4	10
22/5	61	57 - 73	2	3	11
22/6	61	58 - 74	2	3	9
22/7	61	61 - 78	2	3	11
22/8	62	60 - 76	3	5	10
22/9	66	61 - 78	1	2	10
22/10	67	61 - 77	2	3	11
22/11	67	61 - 78	2	3	9
22/12	69	62 - 79	1	1	12
23/1	76	67 - 85	3	4	10
22/2 - 23/1	64	-	2	3	124

Laboratory's analytical performance

January - Internal CV% (compared to all methodologies)





January - Diff% (bias) (compared to all methodologies)



10 / 10 of your results this month are within the target area of your method group.

DayTrol | Bil, tot | BTS 310




Coordinator:

	X_{pt}	$X_{pt} \pm 12\%$	Diff%	CV% target	CV50%	CV90%	n devices	n results
Photometry (Bil)	35.5	31.2 - 39.7	4.8	3.4	2.8	5.1	16	244
All methods	35.5	-	4.8	3.4	2.8	5.1	16	244

Round	X_o	$X_{pt} \pm 12\%$	sd	CV%	n results
22/2	37.7	33.4 - 42.5	1.6	4.2	11
22/3	37.2	32.9 - 41.8	0.9	2.5	11
22/4	38.1	33.2 - 42.2	1.0	2.6	10
22/5	37.4	32.8 - 41.8	1.4	3.8	10
22/6	37.0	32.7 - 41.6	1.5	4.2	8
22/7	37.2	32.9 - 41.8	1.3	3.6	10
22/8	37.4	33.1 - 42.1	1.4	3.8	9
22/9	37.5	33.2 - 42.2	1.2	3.1	10
22/10	37.7	32.8 - 41.8	1.4	3.7	11
22/11	36.3	33.1 - 42.1	0.5	1.4	9
22/12	36.6	33.1 - 42.1	0.8	2.3	12
23/1	37.2	31.2 - 39.7	0.8	2.2	9
22/2 - 23/1	37.3	-	1.2	3.1	120

Laboratory's analytical performance

January - Internal CV% (compared to all methodicses)




January - Diff% (bias) (compared to all methodicses)



9 / 9 of your results this month are within the target area of your method group.

DayTrol | GT | BTS 310



Coordinator:

	X_{pt}	$X_{pt} \pm 12\%$	Diff%	CV% target	CV50%	CV90%	n devices	n results
IFCC methods (GGT)	49	43 - 55	1.5	4.0	2.7	4.2	32	488
All methods	49	-	1.5	4.0	2.7	4.2	32	488

Round	X_o	$X_{pt} \pm 12\%$	sd	CV%	n results
22/2	48	43 - 54	1	2	11
22/3	48	43 - 54	2	4	11
22/4	48	43 - 54	2	3	11
22/5	48	42 - 54	2	4	11
22/6	49	42 - 54	3	6	8
22/7	49	42 - 54	1	2	10
22/8	49	42 - 54	1	2	9
22/9	48	42 - 54	1	3	10
22/10	49	42 - 54	1	3	11
22/11	49	42 - 54	1	2	9
22/12	49	43 - 55	1	3	12
23/1	50	43 - 55	1	2	10
22/2 - 23/1	49	-	1	3	123

Laboratory's analytical performance

January - Internal CV% (compared to all methodicses)



January - Diff% (bias) (compared to all methodicses)



10 / 10 of your results this month are within the target area of your method group.

Report info

Participants

49 participants from 5 countries.

Report info


Your own result should be compared to others using the same method. The laboratory's own result (X_o) is calculated from all results. Assigned values (X_{pt} , target values) are means of the results where results deviating more than +/- 3*standard deviation from the median are removed. The standard uncertainty (u) of the assigned value is reported as standard error of the mean (SEM). In case the client's result is the only one in the method group, no assigned value will be calculated, no target area shown, and no statistics calculated. In case there are only a few

results in the client's own method group, the result can be compared to all method mean or to a group that is similar to the own method.
Results reported with < or >-signs cannot be included in the statistics.

For information on report interpretation and performance evaluation, please see the " EQAS Interpretation guidelines" LabScala User instructions (top right corner ?Help link).

DayTrol | ALT, U/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
IFCC comparable methods (ALT)	79	79	3	3.4	<1	71	83	1	32	486
IFCC methods without P-5-P (ALT)	74	77	8	10.4	4	65	80	-	3	38
All	78	79	3	3.8	<1	69	83	1	35	524



DayTrol | ALT, U/l, Additional summary

Methodics	Method	X_{pt}	sd	CV%	n devices	n results
IFCC comparable methods (ALT)		79	3	3.4	32	486
	Abbott Alinity	74	7	9.8	2	41
	BioSystems	-	-	-	1	8
	Roche cobas	-	-	-	1	28
	Siemens Atellica	-	-	-	1	8
	Thermo Scientific	79	3	3.4	27	401
IFCC methods without P-5-P (ALT)		74	8	10.4	3	38
	Human	-	-	-	1	10
	Roche cobas	-	-	-	1	20
	Thermo Scientific	-	-	-	1	8


Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
IFCC comparable methods (ALT)		79	3	3.4	32	486
	Alinity c	74	7	9.8	2	41
	Atellica CH 930	-	-	-	1	8
	Biosystems BA-400	-	-	-	1	8
	cobas c501	-	-	-	1	28
	Indiko	78	2	2.4	2	29
	Indiko Plus	78	3	4.0	9	135
	Konelab Prime 30	80	<1	0.8	2	34
	Konelab PRIME 60i	80	2	2.3	4	71
	Konelab 20	-	-	-	1	7
	Konelab 20i	78	4	5.0	6	77
	Konelab 20XT	-	-	-	1	4
	Konelab 20XTi	-	-	-	1	21
	Konelab 30i	-	-	-	1	23
IFCC methods without P-5-P (ALT)		74	8	10.4	3	38
	Biosystems BTS-310	-	-	-	1	10
	cobas c111	-	-	-	1	20
	Konelab 20i	-	-	-	1	8

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
IFCC comparable methods (ALT)		79	3	3.4	32	486
	Alpha-ketoglutarate, Ala, P-5-P / NADH consumption; photometry	78	3	3.9	32	486
IFCC methods without P-5-P (ALT)		74	8	10.4	3	38

	Alpha-ketoglutarate, Ala / NADH consumption; photometry	74	8	10.4	3	38
--	--	----	---	------	---	----

DayTrol | Alb, g/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Bromcresol green	36.1	36.3	0.8	2.3	0.5	35.3	36.9	-	3	37
Bromcresol purple	-	-	-	-	-	36.1	36.1	-	1	12
All	36.1	36.2	0.7	1.9	0.3	35.3	36.9	-	4	49



DayTrol | Alb, g/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Bromcresol green		36.1	0.8	2.29	3	37
	Siemens Atellica	-	-	-	1	8
	Thermo Scientific	35.8	0.7	2.06	2	29
Bromcresol purple		-	-	-	1	12
	Thermo Scientific	-	-	-	1	12

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Bromcresol green		36.1	0.8	2.29	3	37
	Atellica CH 930	-	-	-	1	8
	Konelab PRIME 60i	-	-	-	1	21
	Konelab 20i	-	-	-	1	8
Bromcresol purple		-	-	-	1	12
	Konelab PRIME 60i	-	-	-	1	12

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Bromcresol green		36.1	0.8	2.29	3	37
	Bromcresol green (BCG); photometry	36.1	0.8	2.29	3	37
Bromcresol purple		-	-	-	1	12
	Bromcresol purple (BCP); photometry	-	-	-	1	12

DayTrol | ALP, U/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
IFCC methods ALP (IFCC2011)	109	111	9	7.8	2	86	120	-	27	410
All	109	111	9	7.8	2	86	120	-	27	410



DayTrol | ALP, U/l, Additional summary



Methodics	Method	X_{pt}	sd	CV%	n devices	n results
IFCC methods ALP (IFCC2011)		109	9	7.8	27	410
	BioSystems	-	-	-	1	8
	Roche cobas	115	5	4.2	2	47
	Siemens Atellica	-	-	-	1	8
	Thermo Scientific	112	<1	0.1	2	34
	Thermo Scientific (IFCC) Plus	109	8	7.0	21	313

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
IFCC methods ALP (IFCC2011)		109	9	7.8	27	410
	Atellica CH 930	-	-	-	1	8
	Biosystems BA-400	-	-	-	1	8
	cobas c111	-	-	-	1	19
	cobas c501	-	-	-	1	28
	Indiko	-	-	-	1	21
	Indiko Plus	110	8	7.4	9	130
	Konelab Prime 30	110	5	4.6	2	34
	Konelab PRIME 60i	112	1	1.1	4	57
	Konelab 20	-	-	-	1	7
	Konelab 20i	104	13	12.0	4	54
	Konelab 20XTi	-	-	-	1	21
	Konelab 30i	-	-	-	1	23

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
IFCC methods ALP (IFCC2011)		109	9	7.8	27	410
	pNPP, AMP buffer / pNP; photometry	109	9	7.8	27	410

DayTrol | Amyl, U/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
IFCC methods (Amyl)	134	135	2	1.6	<1	130	137	-	13	225
Other methods, calculated to IFCC (Amyl)	125	123	7	5.2	4	121	133	-	3	49
All	132	133	5	3.4	1	121	137	-	16	274



DayTrol | Amyl, U/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
IFCC methods (Amyl)		134	2	1.6	13	225
	Roche cobas	-	-	-	1	28
	Thermo Scientific	134	2	1.6	12	197
Other methods, calculated to IFCC (Amyl)		125	7	5.2	3	49
	Abbott Alinity	122	2	1.4	2	41
	BioSystems	-	-	-	1	8

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
IFCC methods (Amyl)		134	2	1.6	13	225
	cobas c501	-	-	-	1	28
	Indiko Plus	133	2	1.4	4	64
	Konelab Prime 30	-	-	-	1	21
	Konelab PRIME 60i	134	1	1.0	2	33
	Konelab 20	-	-	-	1	7
	Konelab 20i	132	4	3.0	2	28
	Konelab 20XTi	-	-	-	1	21
	Konelab 30i	-	-	-	1	23
Other methods, calculated to IFCC (Amyl)		125	7	5.2	3	49
	Alinity c	122	2	1.4	2	41
	Biosystems BA-400	-	-	-	1	8

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
IFCC methods (Amyl)		134	2	1.6	13	225
	Etyl.-G7-pNP / pNP; photometry	134	2	1.6	13	225
Other methods, calculated to IFCC (Amyl)		125	7	5.2	3	49
	Cl-G3-pNP / Cl-pNP; photometry	125	7	5.2	3	49

DayTrol | AST, U/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
IFCC comparable methods (ASAT)	78	78	4	4.8	<1	68	85	-	26	390
IFCC without P-5-P (AST)	-	-	-	-	-	76	76	-	1	10
All	78	78	4	4.7	<1	68	85	-	27	400



DayTrol | AST, U/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
IFCC comparable methods (ASAT)		78	4	4.8	26	390
	Abbott Alinity	82	2	1.8	2	42
	BioSystems	-	-	-	1	8
	Roche cobas	75	1	1.8	2	48
	Siemens Atellica	-	-	-	1	8
	Thermo Scientific	78	4	4.5	20	284
IFCC without P-5-P (AST)		-	-	-	1	10
	Human	-	-	-	1	10

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
IFCC comparable methods (ASAT)		78	4	4.8	26	390
	Alinity c	82	2	1.8	2	42
	Atellica CH 930	-	-	-	1	8
	Biosystems BA-400	-	-	-	1	8
	cobas c111	-	-	-	1	20
	cobas c501	-	-	-	1	28
	Indiko	-	-	-	1	21
	Indiko Plus	77	1	1.4	7	93
	Konelab Prime 30	79	5	6.0	2	34
	Konelab PRIME 60i	79	2	1.9	2	33
	Konelab 20	-	-	-	1	7
	Konelab 20i	79	6	7.9	6	75
	Konelab 20XTi	-	-	-	1	21
IFCC without P-5-P (AST)		-	-	-	1	10
	Biosystems BTS-310	-	-	-	1	10

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
IFCC comparable methods (ASAT)		78	4	4.8	26	390
	Alpha-ketoglutarate, Asp, P-5-P / NADH consumption; photometry	78	4	4.8	26	390
IFCC without P-5-P (AST)		-	-	-	1	10
	Alpha-ketoglutarate, Asp / NADH consumption; photometry	-	-	-	1	10

DayTrol | Bil, tot, $\mu\text{mol/l}$

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Photometry (Bil)	35.5	35.0	2.1	5.8	0.5	32.9	40.7	-	16	244
All	35.5	35.0	2.1	5.8	0.5	32.9	40.7	-	16	244

DayTrol | Bil, tot, $\mu\text{mol/l}$, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Photometry (Bil)		35.5	2.1	5.81	16	244
	Abbott Alinity	39.8	1.3	3.26	2	42
	BioSystems	-	-	-	1	8
	Human	-	-	-	1	9
	Roche cobas	-	-	-	1	28
	Thermo Scientific	34.4	1.1	3.21	4	37
	Thermo Scientific NBD	34.8	1.3	3.74	7	120

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Photometry (Bil)		35.5	2.1	5.81	16	244
	Alinity c	39.8	1.3	3.26	2	42
	Biosystems BA-400	-	-	-	1	8
	Biosystems BTS-310	-	-	-	1	9
	cobas c501	-	-	-	1	28
	Indiko Plus	34.9	1.4	3.92	5	85
	Konelab PRIME 60i	34.7	1.2	3.55	2	33
	Konelab 20	-	-	-	1	7
	Konelab 20i	-	-	-	1	8
	Konelab 20XT	-	-	-	1	3
	Konelab 20XTi	-	-	-	1	21

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Photometry (Bil)		35.5	2.1	5.81	16	244
	Acid diazo, DCA; photometry	35.5	2.3	6.38	5	62
	Acid diazo, Detergent (Malloy-Evelyn and applications); photometry	35.3	2.2	6.31	9	145
	Alkaline diazo, Caffeine, Na-benzoate (Jendrassik-Grof applications); photometry	36.0	1.7	4.60	2	37

DayTrol | Pi, mmol/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Photometry (Inorganic Phosphate)	1.42	1.40	0.05	3.4	0.02	1.38	1.51	-	6	88
All	1.42	1.40	0.05	3.4	0.02	1.38	1.51	-	6	88



DayTrol | Pi, mmol/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Photometry (Inorganic Phosphate)		1.42	0.05	3.372	6	88
	Abbott Alinity	1.39	0.01	0.969	2	43
	BioSystems	-	-	-	1	8
	Roche cobas	-	-	-	1	28
	Siemens Atellica	-	-	-	1	8
	Thermo Scientific	-	-	-	1	1

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Photometry (Inorganic Phosphate)		1.42	0.05	3.372	6	88
	Alinity c	1.39	0.01	0.969	2	43
	Atellica CH 930	-	-	-	1	8
	Biosystems BA-400	-	-	-	1	8
	cobas c501	-	-	-	1	28
	Indiko Plus	-	-	-	1	1

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Photometry (Inorganic Phosphate)		1.42	0.05	3.372	6	88
	Ammonium molybdate, phosphomolybdate complex, without reducing agent or deproteinization; photometry	1.42	0.05	3.372	6	88

DayTrol | Glucose, mmol/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Photometry (Gluc)	5.44	5.41	0.10	1.8	0.02	5.30	5.68	1	34	514
All	5.44	5.41	0.10	1.8	0.02	5.30	5.68	1	34	514



DayTrol | Glucose, mmol/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Photometry (Gluc)		5.44	0.10	1.752	34	514
	Abbott Alinity	5.37	<0.01	0.011	2	42
	BioSystems	-	-	-	1	8
	Roche cobas	5.44	0.21	3.837	2	48
	Thermo Scientific	5.46	0.11	2.074	29	416

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Photometry (Gluc)		5.44	0.10	1.752	34	514
	Alinity c	5.37	<0.01	0.011	2	42
	Biosystems BA-400	-	-	-	1	8
	cobas c111	-	-	-	1	20
	cobas c501	-	-	-	1	28
	Indiko	5.49	<0.01	0.161	2	29
	Indiko Plus	5.46	0.11	2.094	9	135
	Konelab Prime 30	5.41	0.02	0.302	2	34
	Konelab PRIME 60i	5.41	0.07	1.335	5	79
	Konelab 20	-	-	-	1	7
	Konelab 20i	5.51	0.15	2.761	7	85
	Konelab 20XT	-	-	-	1	3
	Konelab 20XTi	-	-	-	1	21
	Konelab 30i	-	-	-	1	23

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Photometry (Gluc)		5.44	0.10	1.752	34	514
	Glucose dehydrogenase, mutarotase / NADH; photometry	5.34	0.06	1.087	2	40
	Glucose oxidase, H ₂ O ₂ / chromogen; photometry	5.59	0.21	3.816	3	34
	Hexokinase, glucose-6-P-dehydrogenase / NADH; photometry	5.45	0.09	1.737	29	440

DayTrol | GT, U/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
IFCC methods (GGT)	49	49	2	3.4	<1	45	52	-	32	488
All	49	49	2	3.4	<1	45	52	-	32	488



DayTrol | GT, U/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
IFCC methods (GGT)		49	2	3.4	32	488
	Abbott Alinity	47	2	3.3	2	42
	BioSystems	-	-	-	1	8
	Human	-	-	-	1	10
	Roche cobas	49	2	4.1	2	47
	Siemens Atellica	-	-	-	1	8
	Thermo Scientific	49	2	3.1	25	373

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
IFCC methods (GGT)		49	2	3.4	32	488
	Alinity c	47	2	3.3	2	42
	Atellica CH 930	-	-	-	1	8
	Biosystems BA-400	-	-	-	1	8
	Biosystems BTS-310	-	-	-	1	10
	cobas c111	-	-	-	1	19
	cobas c501	-	-	-	1	28
	Indiko	50	1	2.3	2	16
	Indiko Plus	49	<1	1.9	9	134
	Konelab Prime 30	50	<1	0.9	2	34
	Konelab PRIME 60i	48	1	2.7	4	71
	Konelab 20	-	-	-	1	7
	Konelab 20i	48	1	3.1	5	67
	Konelab 20XTi	-	-	-	1	21
	Konelab 30i	-	-	-	1	23

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
IFCC methods (GGT)		49	2	3.4	32	488
	GLUCANa, glycylglycine / p-nitroanilin; photometry	49	2	3.6	26	402
	GLUCANa, glycylglycine,Tris / p-nitroanilin; photometry	50	1	2.4	6	86

DayTrol | K, mmol/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
ISE direct	4.25	4.26	0.05	1.3	0.01	4.11	4.33	-	27	428
ISE indirect	4.33	4.33	0.08	1.8	0.04	4.25	4.44	-	4	78
All	4.26	4.26	0.06	1.5	0.01	4.11	4.44	-	31	506



DayTrol | K, mmol/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
ISE direct		4.25	0.05	1.260	27	428
	Easylite	-	-	-	1	8
	Thermo Scientific	4.26	0.05	1.184	24	398
	Thermo Scientific electrolyte analysers	4.27	0.04	0.915	2	22
ISE indirect		4.33	0.08	1.798	4	78
	Abbott Alinity	4.28	0.04	1.046	2	42
	Roche cobas	-	-	-	1	28
	Siemens Atellica	-	-	-	1	8

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
ISE direct		4.25	0.05	1.260	27	428
	EasyLyte	-	-	-	1	8
	Indiko	-	-	-	1	21
	Indiko Plus	4.28	0.03	0.816	10	157
	Konelab Prime 30	4.26	0.07	1.613	2	34
	Konelab PRIME 60i	4.26	0.06	1.312	4	71
	Konelab 20	-	-	-	1	7
	Konelab 20i	4.21	0.05	1.229	6	87
	Konelab 20XTi	-	-	-	1	21
	Konelab 30i	-	-	-	1	22
ISE indirect		4.33	0.08	1.798	4	78
	Alinity c	4.28	0.04	1.046	2	42
	Atellica CH 930	-	-	-	1	8
	cobas c501	-	-	-	1	28

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
ISE direct		4.25	0.05	1.260	27	428
	Direct potentiometry	4.25	0.05	1.260	27	428
ISE indirect		4.33	0.08	1.798	4	78
	Indirect potentiometry	4.33	0.08	1.798	4	78

DayTrol | Ca, mmol/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Photometry (Ca)	2.47	2.43	0.08	3.4	0.03	2.41	2.61	-	8	120
All	2.47	2.43	0.08	3.4	0.03	2.41	2.61	-	8	120



DayTrol | Ca, mmol/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Photometry (Ca)		2.47	0.08	3.368	8	120
	Abbott Alinity	2.42	0.02	0.890	2	42
	BioSystems	-	-	-	1	8
	Roche cobas	-	-	-	1	28
	Siemens Atellica	-	-	-	1	8
	Thermo Scientific	2.44	0.03	1.106	3	34

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Photometry (Ca)		2.47	0.08	3.368	8	120
	Alinity c	2.42	0.02	0.890	2	42
	Atellica CH 930	-	-	-	1	8
	Biosystems BA-400	-	-	-	1	8
	cobas c501	-	-	-	1	28
	Indiko Plus	-	-	-	1	1
	Konelab PRIME 60i	2.42	<0.01	0.156	2	33

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Photometry (Ca)		2.47	0.08	3.368	8	120
	Arsenazo III; photometry	2.48	0.09	3.476	7	92
	5-nitro-5'-methyl-BAPTA (NM-BAPTA)	-	-	-	1	28

DayTrol | Cl, mmol/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
ISE direct	111	111	1	1.2	<1	110	112	-	2	9
ISE indirect	113	113	3	2.4	2	110	115	-	3	47
All	112	112	2	2.0	<1	110	115	-	5	56



DayTrol | Cl, mmol/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
ISE direct		111	1	1.2	2	9
	Easylite	-	-	-	1	8
	Thermo Scientific electrolyte analysers	-	-	-	1	1
ISE indirect		113	3	2.4	3	47
	Abbott Alinity	-	-	-	1	11
	Roche cobas	-	-	-	1	28
	Siemens Atellica	-	-	-	1	8

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
ISE direct		111	1	1.2	2	9
	EasyLyte	-	-	-	1	8
	Indiko Plus	-	-	-	1	1
ISE indirect		113	3	2.4	3	47
	Alinity c	-	-	-	1	11
	Atellica CH 930	-	-	-	1	8
	cobas c501	-	-	-	1	28

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
ISE direct		111	1	1.2	2	9
	Direct potentiometry	111	1	1.2	2	9
ISE indirect		113	3	2.4	3	47
	Indirect potentiometry	113	3	2.4	3	47

DayTrol | Chol, mmol/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Photometry, enzymatic (Chol)	4.89	4.91	0.11	2.2	0.02	4.69	5.20	-	34	507
All	4.89	4.91	0.11	2.2	0.02	4.69	5.20	-	34	507



DayTrol | Chol, mmol/l, Additional summary





Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Photometry, enzymatic (Chol)		4.89	0.11	2.231	34	507
	Abbott Alinity	4.82	0.03	0.538	2	44
	BioSystems	-	-	-	1	8
	Dialab reagents	-	-	-	1	8
	Roche cobas	4.89	0.04	0.739	2	48
	Thermo Scientific	4.88	0.09	1.939	28	399

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Photometry, enzymatic (Chol)		4.89	0.11	2.231	34	507
	Alinity c	4.82	0.03	0.538	2	44
	Atellica CH 930	-	-	-	1	8
	Biosystems BA-400	-	-	-	1	8
	cobas c111	-	-	-	1	20
	cobas c501	-	-	-	1	28
	Indiko	4.83	0.12	2.440	3	37
	Indiko Plus	4.86	0.11	2.266	9	134
	Konelab Prime 30	4.93	<0.01	0.032	2	34
	Konelab PRIME 60i	4.86	0.04	0.916	3	54
	Konelab 20	-	-	-	1	7
	Konelab 20i	4.92	0.08	1.556	7	85
	Konelab 20XT	-	-	-	1	4
	Konelab 20XTi	-	-	-	1	21
	Konelab 30i	-	-	-	1	23

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Photometry, enzymatic (Chol)		4.89	0.11	2.231	34	507
	Cholesterol esterase, cholesterol oxidase, H ₂ O ₂ , peroxidase / chromogen; photometry	4.89	0.11	2.231	34	507

DayTrol | Chol-HDL, mmol/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Chol HDL direct, Biosystems	-	-	-	-	-	1.86	1.86	-	1	8
Chol-HDL direct, Abbott	1.75	1.75	<0.01	0.6	<0.01	1.74	1.75	-	2	43
Chol-HDL direct, Roche	-	-	-	-	-	1.60	1.60	-	1	20
Chol-HDL direct, Thermo Scientific	1.59	1.60	0.04	2.4	<0.01	1.52	1.69	-	21	319
All	1.61	1.60	0.06	3.5	0.01	1.52	1.75	1	25	390



DayTrol | Chol-HDL, mmol/l, Additional summary

Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Chol HDL direct, Biosystems	BioSystems	-	-	-	1	8
Chol-HDL direct, Abbott	Abbott Alinity	1.75	<0.01	0.566	2	43
Chol-HDL direct, Roche	Roche cobas	-	-	-	1	20
Chol-HDL direct, Thermo Scientific	Thermo Scientific	1.59	0.04	2.424	21	319
	Thermo Scientific HDL Cholesterol Plus	1.62	0.04	2.430	7	98
		1.58	0.03	1.985	14	221


Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Chol HDL direct, Biosystems	Biosystems BA-400	-	-	-	1	8
Chol-HDL direct, Abbott	Alinity c	1.75	<0.01	0.566	2	43
Chol-HDL direct, Roche		-	-	-	1	20

	cobas c111	-	-	-	1	20
Chol-HDL direct, Thermo Scientific		1.59	0.04	2.424	21	319
	Indiko	1.63	0.05	3.140	3	37
	Indiko Plus	1.59	0.04	2.502	7	108
	Konelab Prime 30	1.61	<0.01	0.537	2	34
	Konelab PRIME 60i	1.55	<0.01	0.081	2	33
	Konelab 20	-	-	-	1	7
	Konelab 20i	1.58	0.03	1.851	4	56
	Konelab 20XTi	-	-	-	1	21
	Konelab 30i	-	-	-	1	23

Methodics	Chemical principle	X _{pt}	sd	CV%	n devices	n results
Chol HDL direct, Biosystems	Direct measurement of HDL-cholesterol after decomposition of cholesterol from other lipoproteines	-	-	-	1	8
Chol-HDL direct, Abbott	Direct measurement of HDL-cholesterol after decomposition of cholesterol from other lipoproteines	1.75	<0.01	0.566	2	43
Chol-HDL direct, Roche	Direct measurement of HDL-cholesterol after decomposition of cholesterol from other lipoproteines	-	-	-	1	20
Chol-HDL direct, Thermo Scientific	Direct measurement of HDL-cholesterol after decomposition of cholesterol from other lipoproteines	1.59	0.04	2.424	21	319

DayTrol | CK, U/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
IFCC methods (CK)	221	219	6	2.8	2	214	233	1	12	151
All	221	219	6	2.8	2	214	233	1	12	151



DayTrol | CK, U/l, Additional summary



Methodics	Method	X_{pt}	sd	CV%	n devices	n results
IFCC methods (CK)		221	6	2.8	12	151
	Abbott Alinity	218	1	0.5	2	41
	BioSystems	-	-	-	1	8
	Roche cobas	-	-	-	1	28
	Siemens Atellica	-	-	-	1	8
	Thermo Scientific	223	7	3.1	7	66

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
IFCC methods (CK)		221	6	2.8	12	151
	Alinity c	218	1	0.5	2	41
	Atellica CH 930	-	-	-	1	8
	Biosystems BA-400	-	-	-	1	8
	cobas c501	-	-	-	1	28
	Indiko Plus	217	2	1.0	2	20
	Konelab PRIME 60i	227	<1	<0.1	2	20
	Konelab 20	-	-	-	1	7
	Konelab 20i	-	-	-	1	15
	Konelab 20XT	-	-	-	1	4

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
IFCC methods (CK)		221	6	2.8	12	151
	Creatine-P, ADP, NAC activator / NADPH; photometry	227	21	9.3	12	151

DayTrol | Crea, $\mu\text{mol/l}$

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Photometry, enzymatic (Crea)	167.2	167.0	3.6	2.2	0.6	161.0	174.8	-	33	521
Photometry, Jaffe	158.0	158.0	1.0	0.6	0.6	156.9	158.9	-	3	37
All	166.4	166.6	4.3	2.6	0.7	156.9	174.8	-	36	558

DayTrol | Crea, $\mu\text{mol/l}$, Additional summary

Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Photometry, enzymatic (Crea)		167.2	3.6	2.17	33	521
	Abbott Alinity	163.0	1.8	1.12	2	43
	Roche cobas	168.1	4.2	2.50	2	47
	Siemens Atellica	-	-	-	1	8
	Siemens Dimension Vista System	-	-	-	1	10
	Thermo Scientific	167.1	3.4	2.05	27	413
Photometry, Jaffe		158.0	1.0	0.63	3	37
	BioSystems	-	-	-	1	8
	Human	-	-	-	1	8
	Thermo Scientific	-	-	-	1	21


Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Photometry, enzymatic (Crea)		167.2	3.6	2.17	33	521
	Alinity c	163.0	1.8	1.12	2	43
	Atellica CH 930	-	-	-	1	8
	cobas c111	-	-	-	1	19
	cobas c501	-	-	-	1	28
	Indiko	165.4	1.6	0.99	2	29
	Indiko Plus	167.6	3.6	2.17	9	136
	Konelab Prime 30	169.2	1.4	0.80	2	34
	Konelab PRIME 60i	170.6	3.1	1.84	4	71
	Konelab 20	-	-	-	1	7
	Konelab 20i	165.6	3.5	2.13	7	98
	Konelab 20XT	-	-	-	1	4
	Konelab 20XTi	-	-	-	1	21
	Konelab 30i	-	-	-	1	23
Photometry, Jaffe		158.0	1.0	0.63	3	37
	Biosystems BA-400	-	-	-	1	8
	Indiko Plus	-	-	-	1	21
	Konelab 20i	-	-	-	1	8

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Photometry, enzymatic (Crea)		167.2	3.6	2.17	33	521
	Creatininase / PAP; photometry	166.6	2.9	1.76	11	208
	Creatinine deiminase / NADPH; photometry	164.9	1.7	1.03	2	31
	Enzyme / H ₂ O ₂ / chromogen; photometry	168.0	4.2	2.48	16	239
	Sarcosine oxidase method	166.9	4.1	2.45	4	43

Photometry, Jaffe		158.0	1.0	0.63	3	37
	Alkaline picrate (Jaffe); photometry	158.0	1.0	0.63	3	37

DayTrol | Lactate, mmol/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Photometry enzymatic (Lactate)	2.13	2.11	0.06	2.6	0.03	2.08	2.19	-	3	46
All	2.13	2.11	0.06	2.6	0.03	2.08	2.19	-	3	46



DayTrol | Lactate, mmol/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Photometry enzymatic (Lactate)		2.13	0.06	2.588	3	46
	Abbott Alinity	-	-	-	1	13
	Roche cobas	-	-	-	1	28
	Thermo Scientific	-	-	-	1	5

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Photometry enzymatic (Lactate)		2.13	0.06	2.588	3	46
	Alinity c	-	-	-	1	13
	cobas c501	-	-	-	1	28
	Konelab 20XT	-	-	-	1	5

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Photometry enzymatic (Lactate)		2.13	0.06	2.588	3	46
	Lactate, NAD / NADH; photometry	-	-	-	1	28
	Lactate oxidase / H ₂ O ₂ / 4-aminoantipyrine; photometry	2.15	0.06	2.589	2	18

DayTrol | LD, U/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
IFCC methods (LDH)	130	128	6	4.4	3	125	138	-	4	71
All	130	128	6	4.4	3	125	138	-	4	71



DayTrol | LD, U/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
IFCC methods (LDH)		130	6	4.4	4	71
	Abbott Alinity	126	1	0.9	2	42
	Roche cobas	-	-	-	1	28
	Thermo Scientific	-	-	-	1	1

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
IFCC methods (LDH)		130	6	4.4	4	71
	Alinity c	126	1	0.9	2	42
	cobas c501	-	-	-	1	28
	Indiko Plus	-	-	-	1	1

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
IFCC methods (LDH)		130	6	4.4	4	71
	Lactate, NAD / NADH; photometry	130	6	4.4	4	71

DayTrol | Li, mmol/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Photometry (Lithium)	1.09	1.09	0.04	3.3	0.03	1.06	1.11	-	2	28
All	1.09	1.09	0.04	3.3	0.03	1.06	1.11	-	2	28



DayTrol | Li, mmol/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Photometry (Lithium)		1.09	0.04	3.266	2	28
	Abbott Alinity	-	-	-	1	13
	Thermo Scientific	-	-	-	1	15

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Photometry (Lithium)		1.09	0.04	3.266	2	28
	Alinity c	-	-	-	1	13
	Indiko Plus	-	-	-	1	15

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Photometry (Lithium)		1.09	0.04	3.266	2	28
	Chromogen	1.09	0.04	3.266	2	28

DayTrol | Mg, mmol/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Photometry (Mg)	0.94	0.94	0.02	1.8	<0.01	0.92	0.96	-	5	58
All	0.94	0.94	0.02	1.8	<0.01	0.92	0.96	-	5	58



DayTrol | Mg, mmol/l, Additional summary



Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Photometry (Mg)		0.94	0.02	1.779	5	58
	Abbott Alinity	-	-	-	1	13
	BioSystems	-	-	-	1	8
	Roche cobas	-	-	-	1	28
	Siemens Atellica	-	-	-	1	8
	Thermo Scientific	-	-	-	1	1

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Photometry (Mg)		0.94	0.02	1.779	5	58
	Alinity c	-	-	-	1	13
	Atellica CH 930	-	-	-	1	8
	Biosystems BA-400	-	-	-	1	8
	cobas c501	-	-	-	1	28
	Indiko Plus	-	-	-	1	1

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Photometry (Mg)		0.94	0.02	1.779	5	58
	Calmagite; photometry	-	-	-	1	8
	Enzymatic method; absorption photometry	-	-	-	1	13
	Xylidyl blue; photometry	0.94	0.02	1.714	3	37

DayTrol | Na, mmol/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
ISE direct	145	145	1	0.7	<1	143	148	-	27	429
ISE indirect	149	148	3	2.0	1	146	153	-	4	78
All	145	145	1	0.8	<1	143	148	1	31	507



■ All method groups ■ ISE direct (X_{pt} : 145 | Target area: 142-148 | Target: $\pm 2\%$)

■ All method groups ■ ISE indirect (X_{pt} : 149 | Target area: 146-151 | Target: $\pm 2\%$)

DayTrol | Na, mmol/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
ISE direct		145	1	0.7	27	429
	Easylite	-	-	-	1	8
	Thermo Scientific	145	1	0.7	24	399
	Thermo Scientific electrolyte analysers	145	1	0.7	2	22
ISE indirect		149	3	2.0	4	78
	Abbott Alinity	147	2	1.0	2	42
	Roche cobas	-	-	-	1	28
	Siemens Atellica	-	-	-	1	8

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
ISE direct		145	1	0.7	27	429
	EasyLyte	-	-	-	1	8
	Indiko	-	-	-	1	21
	Indiko Plus	145	1	0.7	10	157
	Konelab Prime 30	145	2	1.2	2	34
	Konelab PRIME 60i	145	<1	0.6	4	71
	Konelab 20	-	-	-	1	7
	Konelab 20i	144	<1	0.6	6	87
	Konelab 20XTi	-	-	-	1	21
	Konelab 30i	-	-	-	1	23
ISE indirect		149	3	2.0	4	78
	Alinity c	147	2	1.0	2	42
	Atellica CH 930	-	-	-	1	8
	cobas c501	-	-	-	1	28

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
ISE direct		145	1	0.7	27	429
	Direct potentiometry	145	1	0.7	27	429
ISE indirect		149	3	2.0	4	78
	Indirect potentiometry	149	3	2.0	4	78

DayTrol | Prot, g/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Photometry (Prot)	55.9	55.9	1.2	2.2	0.5	54.6	57.6	-	6	62
All	55.9	55.9	1.2	2.2	0.5	54.6	57.6	-	6	62



DayTrol | Prot, g/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Photometry (Prot)		55.9	1.2	2.18	6	62
	Abbott Alinity	-	-	-	1	9
	BioSystems	-	-	-	1	8
	Roche cobas	-	-	-	1	28
	Siemens Atellica	-	-	-	1	8
	Thermo Scientific T PROT Plus	56.4	0.9	1.57	2	9

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Photometry (Prot)		55.9	1.2	2.18	6	62
	Alinity c	-	-	-	1	9
	Atellica CH 930	-	-	-	1	8
	Biosystems BA-400	-	-	-	1	8
	cobas c501	-	-	-	1	28
	Indiko Plus	-	-	-	1	1
	Konelab 20i	-	-	-	1	8

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Photometry (Prot)		55.9	1.2	2.18	6	62
	Biuret method; photometry	55.9	1.2	2.18	6	62

DayTrol | Transf, g/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Turbidometry (Transf)	-	-	-	-	-	2.06	2.06	-	1	1
All	-	-	-	-	-	2.06	2.06	-	1	1



DayTrol | Transf, g/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Turbidometry (Transf)		-	-	-	1	1
	Thermo Scientific	-	-	-	1	1

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Turbidometry (Transf)		-	-	-	1	1
	Indiko Plus	-	-	-	1	1

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Turbidometry (Transf)		-	-	-	1	1
	Antigen-antibody (Ag-Ab) complex; turbidimetry	-	-	-	1	1

DayTrol | TfR, mg/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Thermo Scientific (TfR)	-	-	-	-	-	0.68	0.68	-	1	1
All	-	-	-	-	-	0.68	0.68	-	1	1



DayTrol | TfR, mg/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Thermo Scientific (TfR)		-	-	-	1	1
	Thermo Scientific sTfR	-	-	-	1	1

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Thermo Scientific (TfR)		-	-	-	1	1
	Indiko Plus	-	-	-	1	1

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Thermo Scientific (TfR)		-	-	-	1	1
	Antigen-antibody (Ag-Ab) complex; turbidimetry	-	-	-	1	1

DayTrol | Trigly, mmol/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Photometry, enzymatic (Trigly)	1.51	1.51	0.06	3.7	<0.01	1.40	1.63	-	34	511
All	1.51	1.51	0.06	3.7	<0.01	1.40	1.63	-	34	511



DayTrol | Trigly, mmol/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Photometry, enzymatic (Trigly)		1.51	0.06	3.745	34	511
	Abbott Alinity	1.44	0.02	1.057	2	43
	BioSystems	-	-	-	1	8
	Roche cobas	1.47	0.08	5.258	2	48
	Thermo Scientific	1.52	0.05	3.287	29	412

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Photometry, enzymatic (Trigly)		1.51	0.06	3.745	34	511
	Alinity c	1.44	0.02	1.057	2	43
	Biosystems BA-400	-	-	-	1	8
	cobas c111	-	-	-	1	20
	cobas c501	-	-	-	1	28
	Indiko	1.52	0.09	5.928	3	37
	Indiko Plus	1.48	0.03	2.258	9	134
	Konelab Prime 30	1.54	0.03	1.802	2	34
	Konelab PRIME 60i	1.53	0.04	2.308	4	68
	Konelab 20	-	-	-	1	7
	Konelab 20i	1.56	0.04	2.790	7	84
	Konelab 20XT	-	-	-	1	4
	Konelab 20XTi	-	-	-	1	21
	Konelab 30i	-	-	-	1	23

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Photometry, enzymatic (Trigly)		1.51	0.06	3.745	34	511
	Lipase, GDH / NADH; photometry	1.53	0.07	4.670	5	85
	Lipase, GK/G-3-P-oxidase, H2O2, peroxidase / chromogen; photometry	1.51	0.05	3.620	29	426

DayTrol | Urea, mmol/l

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Photometry, enzymatic (Urea)	5.25	5.29	0.17	3.2	0.06	4.95	5.45	-	7	96
All	5.25	5.29	0.17	3.2	0.06	4.95	5.45	-	7	96



DayTrol | Urea, mmol/l, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Photometry, enzymatic (Urea)		5.25	0.17	3.196	7	96
	Abbott Alinity	5.09	0.19	3.790	2	43
	BioSystems	-	-	-	1	8
	Roche cobas	-	-	-	1	28
	Siemens Atellica	-	-	-	1	8
	Thermo Scientific	5.22	0.12	2.288	2	9

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Photometry, enzymatic (Urea)		5.25	0.17	3.196	7	96
	Alinity c	5.09	0.19	3.790	2	43
	Atellica CH 930	-	-	-	1	8
	Biosystems BA-400	-	-	-	1	8
	cobas c501	-	-	-	1	28
	Indiko Plus	-	-	-	1	1
	Konelab 20i	-	-	-	1	8

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Photometry, enzymatic (Urea)	Urease, glutamate dehydrogenase / NADH; photometry	5.25	0.17	3.196	7	96

DayTrol | Uric acid, $\mu\text{mol/l}$

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Photometry, enzymatic (Uric acid)	390	390	12	3.2	3	371	425	-	18	310
All	390	390	12	3.2	3	371	425	-	18	310

DayTrol | Uric acid, $\mu\text{mol/l}$, Additional summary


Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Photometry, enzymatic (Uric acid)		390	12	3.2	18	310
	Abbott Alinity	373	4	1.1	2	43
	BioSystems	-	-	-	1	8
	Roche cobas	-	-	-	1	28
	Thermo Scientific	390	8	2.0	14	231

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Photometry, enzymatic (Uric acid)		390	12	3.2	18	310
	Alinity c	373	4	1.1	2	43
	Biosystems BA-400	-	-	-	1	8
	cobas c501	-	-	-	1	28
	Indiko Plus	386	6	1.6	4	63
	Konelab Prime 30	-	-	-	1	13
	Konelab PRIME 60i	393	3	0.7	3	54
	Konelab 20i	387	9	2.2	4	57
	Konelab 20XTi	-	-	-	1	21
	Konelab 30i	-	-	-	1	23

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Photometry, enzymatic (Uric acid)		390	12	3.2	18	310
	Uricase / allantoin; photometry	387	6	1.6	6	92
	Uricase, H ₂ O ₂ , peroxidase / chromogen; photometry	392	15	3.7	12	218

DayTrol | Fe, $\mu\text{mol/l}$

Methodics	X_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n devices	n results
Photometry (Fe)	25.6	25.8	1.6	6.4	0.5	21.9	27.5	-	9	117
All	25.6	25.8	1.6	6.4	0.5	21.9	27.5	-	9	117

DayTrol | Fe, $\mu\text{mol/l}$, Additional summary

Methodics	Method	X_{pt}	sd	CV%	n devices	n results
Photometry (Fe)		25.6	1.6	6.45	9	117
	Abbott Alinity	25.6	0.3	1.12	2	43
	BioSystems	-	-	-	1	8
	Roche cobas	-	-	-	1	28
	Siemens Atellica	-	-	-	1	8
	Thermo Scientific	26.7	0.7	2.68	4	30

Methodics	Instrument	X_{pt}	sd	CV%	n devices	n results
Photometry (Fe)		25.6	1.6	6.45	9	117
	Alinity c	25.6	0.3	1.12	2	43
	Atellica CH 930	-	-	-	1	8
	Biosystems BA-400	-	-	-	1	8
	cobas c501	-	-	-	1	28
	Indiko Plus	26.7	0.6	2.18	2	10
	Konelab PRIME 60i	-	-	-	1	12
	Konelab 20i	-	-	-	1	8

Methodics	Chemical principle	X_{pt}	sd	CV%	n devices	n results
Photometry (Fe)		25.6	1.6	6.45	9	117
	Ferene; photometry	26.4	0.8	3.00	6	73
	Ferrozine; photometry	24.0	1.9	8.00	3	44

Report info**Participants**

49 participants from 5 countries.

Report info

Your own result should be compared to others using the same method.

The laboratory's own result (X_0) is calculated from all results. Assigned values (x_{pt} , target values) are means of the results where results deviating more than +/- 3*standard deviation

from the median are removed. The standard uncertainty (u) of the assigned value is reported as standard error of the mean (SEM).

In case the client's result is the only one in the method group, no assigned value will be calculated, no target area shown, and no statistics calculated. In case there are only a few results in the client's own method group, the result can be compared to all method mean or to a group that is similar to the own method.

Results reported with < or >-signs cannot be included in the statistics.

For information on report interpretation and performance evaluation, please see the " EOAS Interpretation guidelines" LabScala User instructions (top right corner ?Help link).

External Quality Assessment Scheme

General Chemistry, Daytrol Round 1, 2023

Specimens

Sample DT23 was lyophilized human serum sample.

Based on the previous tests and the results of this round, the samples are homogeneous, stable, and suitable for the external quality assessment scheme.

The materials were sent without temperature control packaging.

Report info

It is important to read the Final report first, because it contains important information of the samples and results in each round. Laboratory's analytical performance is given when the number of results ≥ 8 and number of laboratories ≥ 6 . In the client specific report, the symbol X_o represents laboratory's own result. If SD and CV% results would be so low that they would be rounded to zero, then their values are presented as a dash.

We kindly ask the customers to review their results and methods and to contact the EQA coordinator if there are any errors in the methods. Customers will no longer receive separate note letters, but notes will be recorded in the client-specific report below the histogram if needed.

Comments – Expert

This year's sample, DT23, was a human serum sample, which electrolyte, glucose, cholesterol values and most of the enzyme activities were in the normal range. The result variations of this round are generally small. For detailed evaluation of the results of this round, please study the summary report.

End of report

2023-02-06

FINAL REPORT

Product no. 1031

Samples sent 2022-12-13
Round closed 2023-01-31
Report released 2023-02-06

Request for correction

Typing errors in laboratory's result forms are on laboratory's responsibility. Labquality accepts responsibility only for result processing. Requests must be notified by writing within three weeks from the date of this letter.

Authorized by

EQA Coordinator
Liisa Ylitepsa
liisa.ylitepsa@labquality.fi

Expert

M.Sc.,Chemist, Ulla-Riitta Nordberg

Labquality Oy

Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200
Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.com

