

External Quality Assessment Scheme

Proteins, immunochemical determinations, Round 1, 2023

Specimens

Please find enclosed 2 human serum samples S001 and S002, each 1 mL.

Caution

Quality control specimens derived from human blood must be handled with the same care as patient samples, i.e. as potential transmitters of serious diseases. The specimens are found to be HBsAg, HCVAb and HIVAgAb negative when tested with licensed reagents, but no known test method can offer complete assurance that the specimens will not transmit these or other infectious diseases.

Examinations

A1-Acid Glycoprotein/Orosomucoid

A1-Antitrypsin

A2-Macroglobulin

Albumin

Ceruloplasmin

Complement C3 (C3c)

Complement C4 (C4c)

Haptoglobin

Hemopexin

IgA

IgG

IgLcKappa, total

IgLcLambda, total

IgLcKappa/Lambda, total

IgLcKappa, free

IgLcLambda, free

IgLcKappa/Lambda, free

IgM

Prealbumin/Transthyretin

Retinol Binding Protein

Transferrin

Transferrin receptor

Storage and use

The samples are ready to use, and they stable at 2...8 °C during the whole round processing time. Let the samples reach room temperature before analysis. Ensure homogeneity by gently inverting the sample. Analyse as patient sample.

Result reporting

Please enter the results and methods via LabScala (www.labscala.com). If you cannot find your instrument or reagent from the registry, please contact the EQA Coordinator.

S001:



S002:



2023-01-23

INSTRUCTIONS

Product no. 2230
LQ735523011-012/UK

If the kit is incomplete or contains damaged specimens, please report immediately to info@labquality.fi

The results should be reported no later than **February 13, 2023**.

Inquiries

EQA Coordinator

Päivi Ranta

paivi.ranta@labquality.fi

Labquality Oy

Kumpulantie 15

FI-00520 HELSINKI

Finland

Tel. + 358 9 8566 8200



Fax + 358 9 8566 8280

info@labquality.fi

www.labquality.com



Ceruloplasmin | Cobas c702



	x_{pt}	sd	SEM	CV%	n
Roche, turbidimetry, Ceruloplasmin	0.282 g/l	0.007	0.002	2.4	8
All methods	0.357 g/l	0.064	0.014	18.0	20

	x_{pt}	sd	SEM	CV%	n
Roche, turbidimetry, Ceruloplasmin	0.442 g/l	0.012	0.004	2.7	8
All methods	0.562 g/l	0.108	0.024	19.1	20

Round	Sample	x_{pt}	Result	diff%	z-score
23/1	Specimen S002	0.442	0.447	1.057%	0.38
23/1	Specimen S001	0.282	0.276	-2.132%	-0.89
22/4	Specimen S002	0.228	0.223	-2.193%	-0.54
22/4	Specimen S001	0.206	0.216	4.727%	1.02
22/3	Specimen S002	0.252	0.248	-1.587%	-0.08
22/3	Specimen S001	0.158	0.146	-7.302%	-1.22
22/2	Specimen S002	0.379	0.367	-3.134%	-7.23
22/2	Specimen S001	0.155	0.148	-4.310%	-1.05
22/1	Specimen S002	0.250	0.241	-3.471%	-1.76
22/1	Specimen S001	0.234	0.225	-4.028%	-1.52
21/4	Specimen S002	0.196	0.187	-4.375%	-1.56
21/4	Specimen S001	0.252	0.250	-0.968%	-0.31

Report info**Participants**

63 participants from 17 countries.

Report info

Your own result should be compared to others using the same method.

Assigned values (x_{pt} , target values) are means of the results where results deviating more than +/- 3*standard deviation from the median are removed. The standard uncertainty (u) of

the assigned value is reported as standard error of the mean (SEM). Additionally, if the measurement uncertainty of the target value is large an automatic text is printed on the report:

"The uncertainty of the assigned value is not negligible, and evaluations could be affected."

In case the client's result is the only one in the method group, no assigned value will be calculated, no target area shown, and no statistics calculated. In case there are only a few

results in the client's own method group, the result can be compared to all method mean or to a group that is similar to the own method.





Results reported with < or >-signs cannot be included in the statistics.

For information on report interpretation and performance evaluation, please see the " EOAS Interpretation guidelines" LabScala User instructions (top right corner ?Help link).

Specimen S001 | A1-Glycoprotein, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Beckman Coulter, turbidimetry, A1-Glycoprotein	-	-	-	-	-	1.070	1.070	-	1
Roche, turbidimetry, A1-Glycoprotein	1.045	1.045	0.006	0.6	0.005	1.040	1.049	-	2
Siemens, nefelometry, A1-Glycoprotein	1.059	1.050	0.085	8.1	0.028	0.980	1.270	-	9
Siemens, turbidimetry, A1-Glycoprotein	-	-	-	-	-	1.130	1.130	-	1
All	1.046	1.050	0.040	3.8	0.011	0.980	1.130	1	13





Specimen S001 | A1-Glycoprotein, g/l histogram summaries in LabScale




Specimen S001 | A1-Antitrypsin, g/l

Methodics		x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, A1-Antitrypsin		1.29	1.26	0.06	4.6	0.03	1.26	1.36	-	3
Beckman Coulter, turbidimetry, A1-Antitrypsin	-	-	-	-	-	-	1.26	1.26	-	1
Binding site, turbidometry, A1-Antitrypsin	-	-	-	-	-	-	1.11	1.11	-	1
Roche, turbidimetry, A1-Antitrypsin		1.25	1.24	0.08	6.5	0.03	1.18	1.42	-	7
Siemens, nefelometry, A1-Antitrypsin		1.31	1.34	0.07	5.4	0.02	1.14	1.40	-	13
Siemens, turbidimetry, A1-Antitrypsin		1.11	1.10	0.03	2.4	0.01	1.09	1.15	-	5
All		1.25	1.26	0.10	7.8	0.02	1.09	1.42	-	30


Specimen S001 | A1-Antitrypsin, g/l histogram summaries in LabScala



Siemens, nephelometry, A1-Antitrypsin




Siemens, turbidimetry, A1-Antitrypsin



Specimen S001 | A2-Makroglobulin, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Siemens, nefelometry, A2-Macroglobulin	1.94	1.95	0.05	2.6	0.02	1.88	1.99	-	6
All	1.94	1.95	0.05	2.6	0.02	1.88	1.99	-	6


Specimen S001 | A2-Makroglobulin, g/l| histogram summaries in LabScala



Specimen S001 | Albumin, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Albumin chemical	43.3	44.0	1.8	4.2	0.5	40.0	45.3	1	16
Albumin immunochemical	42.7	42.6	1.5	3.5	0.4	40.5	46.7	1	19
All	42.6	42.7	2.3	5.4	0.4	36.0	46.7	-	35





Specimen S001 | Albumin, g/l histogram summaries in LabScala



Specimen S001 | Ceruloplasmin, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, Ceruloplasmin	0.395	0.395	0.007	1.8	0.005	0.390	0.400	-	2
Roche, turbidimetry, Ceruloplasmin	0.282	0.279	0.007	2.4	0.002	0.275	0.290	-	8
Siemens, nefelometry, Ceruloplasmin	0.411	0.405	0.019	4.7	0.006	0.375	0.434	-	9
Siemens, turbidimetry, Ceruloplasmin	-	-	-	-	-	0.390	0.390	-	1
All	0.357	0.390	0.064	18.0	0.014	0.275	0.434	-	20





Specimen S001 | Ceruloplasmin, g/l histogram summaries in LabScala





Specimen S001 | Complement C3, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, C3	1.617	1.620	0.023	1.4	0.010	1.590	1.640	-	6
Beckman Coulter, turbidimetry, C3	1.520	1.520	0.071	4.7	0.050	1.470	1.570	-	2
Roche, turbidimetry, C3	1.635	1.649	0.064	3.9	0.021	1.520	1.730	-	9
Siemens, nefelometry, C3	1.598	1.600	0.065	4.1	0.018	1.430	1.680	-	13
Siemens, turbidimetry, C3	1.570	1.570	0.060	3.8	0.035	1.510	1.630	-	3
Thermo Scientific, turbidimetry, C3	-	-	-	-	-	1.560	1.560	-	1
All	1.603	1.600	0.062	3.9	0.011	1.430	1.730	-	34

Specimen S001 | Complement C3, g/l histogram summaries in LabScala











Specimen S001 | Complement C4, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, C4	0.326	0.327	0.006	1.8	0.002	0.320	0.334	-	6
Beckman Coulter, turbidimetry, C4	-	-	-	-	-	0.360	0.360	-	1
Roche, turbidimetry, C4	0.302	0.300	0.014	4.6	0.005	0.290	0.330	-	9
Siemens, nefelometry, C4	0.326	0.328	0.016	5.0	0.005	0.300	0.360	-	13
Siemens, turbidimetry, C4	0.317	0.320	0.006	1.8	0.003	0.310	0.320	-	3
Thermo Scientific, turbidimetry, C4	-	-	-	-	-	0.340	0.340	-	1
All	0.320	0.320	0.018	5.7	0.003	0.290	0.360	-	33

Specimen S001 | Complement C4, g/l histogram summaries in LabScala







Specimen S001 | Haptoglobin, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, Haptoglobin	0.90	0.90	0.01	1.5	<0.01	0.87	0.91	-	6
Beckman Coulter, turbidimetry, Haptoglobin	-	-	-	-	-	0.91	0.91	-	1
Roche, turbidimetry, Haptoglobin	0.89	0.89	0.03	2.9	<0.01	0.85	0.93	-	11
Siemens, nefelometry, Haptoglobin	0.87	0.88	0.02	2.7	<0.01	0.83	0.90	-	11
Siemens, turbidimetry, Haptoglobin	0.86	0.86	0.02	2.3	<0.01	0.83	0.88	-	5
All	0.88	0.89	0.03	2.9	<0.01	0.83	0.93	-	34

Specimen S001 | Haptoglobin, g/l histogram summaries in LabScala






Specimen S001 | Hemopexin, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Siemens, nefelometry, Hemopexin	0.60	0.60	<0.01	1.2	<0.01	0.60	0.61	-	2
All	0.60	0.60	<0.01	1.2	<0.01	0.60	0.61	-	2





Specimen S001 | Hemopexin, g/l histogram summaries in LabScala






Specimen S001 | IgA, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, IgA	2.96	2.98	0.05	1.8	0.02	2.84	3.02	-	10
Beckman Coulter, turbidimetry, IgA	2.99	2.99	0.04	1.2	0.03	2.96	3.01	-	2
Binding site, turbidometry, IgA	-	-	-	-	-	2.90	2.90	-	1
Roche, turbidimetry, IgA	3.01	3.01	0.09	3.1	0.03	2.89	3.15	-	12
Siemens, nefelometry, IgA	3.10	3.07	0.13	4.3	0.03	2.85	3.31	-	15
Siemens, turbidimetry, IgA	3.05	3.06	0.04	1.3	0.02	3.01	3.09	-	3
Thermo Scientific, turbidimetry, IgA	-	-	-	-	-	3.01	3.01	-	1
All	3.03	3.01	0.11	3.7	0.02	2.84	3.31	-	44

Specimen S001 | IgA, g/l histogram summaries in LabScala












Specimen S001 | IgG, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, IgG	15.2	15.3	0.4	2.6	0.1	14.6	15.8	-	10
Beckman Coulter, turbidimetry, IgG	14.3	14.3	<0.1	0.6	<0.1	14.3	14.4	-	2
Binding site, turbidometry, IgG	-	-	-	-	-	14.9	14.9	-	1
Roche, turbidimetry, IgG	14.9	15.1	0.8	5.6	0.3	12.9	16.0	-	11
Siemens, nefelometry, IgG	15.1	14.9	0.7	4.5	0.2	14.0	16.3	-	18
Siemens, turbidimetry, IgG	15.4	15.4	0.2	1.4	0.1	15.2	15.6	-	3
Thermo Scientific, turbidimetry, IgG	-	-	-	-	-	14.0	14.0	-	1
All	15.1	15.1	0.6	3.8	<0.1	14.0	16.3	1	46

Specimen S001 | IgG, g/l histogram summaries in LabScala








Specimen S001 | IgLCKappa, total, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidometry (Light chains)	-	-	-	-	-	3.18	3.18	-	1
Roche, turbidimetry	3.94	3.94	0.05	1.4	0.04	3.90	3.98	-	2
Siemens, nefelometry	3.56	3.62	0.20	5.7	0.09	3.33	3.83	-	5
All	3.61	3.63	0.29	8.0	0.10	3.18	3.98	-	8





Specimen S001 | IgLCKappa, total, g/l histogram summaries in LabScala



Specimen S001 | IgLC kappa, free, mg/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Freelite, nephelometry	18.85	18.90	0.96	5.1	0.39	17.40	20.40	-	6
Freelite, turbidimetry	15.49	14.80	1.88	12.2	0.63	13.50	18.67	-	9
Siemens, nephelometry	10.55	10.55	0.37	3.5	0.11	9.95	11.10	-	12
Siemens turbidimetry	11.14	11.10	0.10	0.9	0.06	11.07	11.26	-	3
All	13.75	12.38	3.53	25.7	0.64	9.95	20.40	-	30




Specimen S001 | IgLC kappa, free, mg/l histogram summaries in LabScala



Specimen S001 | IgLCLambda, total, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidometry (Light chains)	-	-	-	-	-	1.66	1.66	-	1
Roche, turbidimetry	2.20	2.20	0.11	5.2	0.08	2.12	2.28	-	2
Siemens, nefelometry	1.94	1.94	0.07	3.5	0.03	1.85	2.03	-	5
All	1.97	1.96	0.18	9.4	0.07	1.66	2.28	-	8





Specimen S001 | IgLCLambda, total, g/l histogram summaries in LabScala



Specimen S001 | IgLCLambda, free, mg/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Freelite, nephelometry	15.20	15.40	1.87	12.3	0.77	12.60	17.30	-	6
Freelite, turbidimetry	14.08	14.65	1.44	10.3	0.46	11.46	15.70	-	10
Siemens, nephelometry	21.77	21.90	1.33	6.1	0.38	19.00	23.30	-	12
Siemens turbidimetry	24.42	25.07	2.27	9.3	1.31	21.90	26.30	-	3
All	18.27	17.30	4.31	23.6	0.77	11.46	26.30	-	31


Specimen S001 | IgLCLambda, free, mg/l histogram summaries in LabScala




Specimen S001 | IgM, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, IgM	1.20	1.20	0.03	2.9	0.01	1.16	1.26	-	10
Beckman Coulter, turbidimetry, IgM	1.16	1.16	<0.01	<0.1	<0.01	1.16	1.16	-	2
Binding site, turbidometry, IgM	-	-	-	-	-	1.21	1.21	-	1
Roche, turbidimetry, IgM	1.20	1.20	0.03	2.4	<0.01	1.15	1.25	-	12
Siemens, nefelometry, IgM	1.26	1.24	0.04	3.2	0.01	1.20	1.33	-	14
Siemens, turbidimetry, IgM	1.21	1.23	0.03	2.8	0.02	1.17	1.23	-	3
Thermo Scientific, turbidimetry, IgM	-	-	-	-	-	1.19	1.19	-	1
All	1.22	1.22	0.04	3.6	<0.01	1.15	1.33	-	43


Specimen S001 | IgM, g/l histogram summaries in LabScala




All method groups Abbott, turbidimetry, IgM (x_{pt} : 1.20 | Target area: 1.02-1.38 | Target: ±15%)






All method groups Beckman Coulter, turbidimetry, IgM (x_{pt} : 1.16 | Target area: 0.99-1.33 | Target: ±15%)



All method groups Binding site, turbidometry, IgM







All method groups Roche, turbidimetry, IgM (x_{pt} : 1.20 | Target area: 1.02-1.38 | Target: ±15%)





Specimen S001 | Prealbumin / Transtiretin, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, Prealbumin	-	-	-	-	-	0.240	0.240	-	1
Beckman Coulter, turbidimetry, Prealbumin	0.250	0.250	<0.001	<0.1	<0.001	0.250	0.250	-	2
Binding site, turbidometry, Prealbumin	0.228	0.228	0.012	5.3	0.009	0.219	0.236	-	2
Roche, turbidimetry, Prealbumin	0.236	0.237	0.006	2.5	0.002	0.228	0.244	-	8
Siemens, nefelometry, Prealbumin	0.237	0.240	0.016	6.7	0.005	0.202	0.258	-	9
Siemens, turbidimetry, Prealbumin	0.235	0.235	0.007	3.0	0.005	0.230	0.240	-	2
All	0.238	0.240	0.009	3.7	0.002	0.219	0.258	1	24

Specimen S001 | Prealbumin / Transtiretin, g/l histogram summaries in LabScala






Specimen S001 | Retinol Binding Prot, mg/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Siemens, nefelometry, Retinol binding protein	39.3	40.0	2.8	7.1	1.2	36.2	43.2	-	5
All	39.3	40.0	2.8	7.1	1.2	36.2	43.2	-	5





Specimen S001 | Retinol Binding Prot, mg/l histogram summaries in LabScala





Specimen S001 | Transferrin, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, Transferrin	2.88	2.87	0.08	2.7	0.03	2.80	3.07	-	9
Beckman Coulter, turbidimetry, Transferrin	2.88	2.88	<0.01	0.2	<0.01	2.87	2.88	-	2
Roche, turbidimetry, Transferrin	2.96	2.95	0.05	1.8	0.01	2.89	3.08	1	14
Siemens, nefelometry, Transferrin	2.82	2.78	0.14	4.9	0.05	2.67	3.10	-	8
Siemens, turbidimetry, Transferrin	2.88	2.87	0.05	1.7	0.03	2.83	2.93	-	3
Thermo Scientific, turbidimetry, Transferrin	-	-	-	-	-	2.87	2.87	-	1
All	2.89	2.88	0.11	3.8	0.02	2.59	3.10	-	37

Specimen S001 | Transferrin, g/l histogram summaries in LabScala










Specimen S001 | Transferrin receptor, mg/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Randox, turbidimetry, TfR	0.6	0.6	0.1	20.2	<0.1	0.5	0.7	-	2
Roche, turbidimetry, TfR	2.6	2.5	0.2	9.3	<0.1	2.4	3.0	-	7
Siemens, nefelometry, TfR	1.0	1.0	<0.1	3.5	<0.1	1.0	1.1	-	9
Thermo Scientific, turbidimetry, TfR	-	-	-	-	-	0.8	0.8	-	1
Trimero Diagnostics (TfR)	-	-	-	-	-	1.0	1.0	-	1
All	1.5	1.0	0.8	55.2	0.2	0.5	3.0	-	20

Specimen S001 | Transferrin receptor, mg/l histogram summaries in LabScala








Specimen S001 | IgLKappa/Lambda, total, Ratio

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidometry	-	-	-	-	-	1.92	1.92	-	1
Roche, turbidimetry	-	-	-	-	-	1.88	1.88	-	1
Siemens, nefelometry	1.88	1.91	0.13	6.8	0.06	1.67	2.01	-	5
All	1.89	1.91	0.11	5.6	0.04	1.67	2.01	-	7





Specimen S001 | IgLKappa/Lambda, total, Ratio| histogram summaries in LabScala



Specimen S001 | IgLCKappa/Lambda, free, Ratio

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Freelite, nephelometry	1.23	1.15	0.17	14.2	0.08	1.07	1.50	-	5
Freelite, turbidimetry	1.08	1.03	0.15	14.3	0.06	0.94	1.38	-	7
Siemens, nephelometry	0.48	0.47	0.02	4.6	<0.01	0.45	0.51	-	6
Siemens, turbidometry	0.46	0.44	0.04	8.4	0.02	0.43	0.50	-	3
All	0.85	0.94	0.36	42.7	0.08	0.43	1.50	-	21





Specimen S001 | IgLCKappa/Lambda, free, Ratio| histogram summaries in LabScala



Specimen S002 | A1-Glykoprotein, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Beckman Coulter, turbidimetry, A1-Glycoprotein	-	-	-	-	-	1.650	1.650	-	1
Roche, turbidimetry, A1-Glycoprotein	1.556	1.556	0.035	2.2	0.025	1.531	1.580	-	2
Siemens, nefelometry, A1-Glykoprotein	1.624	1.610	0.108	6.6	0.036	1.490	1.810	-	9
Siemens, turbidimetry, A1-Glycoprotein	-	-	-	-	-	1.680	1.680	-	1
All	1.620	1.610	0.094	5.8	0.026	1.490	1.810	-	13





Specimen S002 | A1-Glykoprotein, g/l histogram summaries in LabScala




Specimen S002 | A1-Antitrypsin, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, A1-Antitrypsin	2.00	1.96	0.08	3.9	0.05	1.95	2.09	-	3
Beckman Coulter, turbidimetry, A1-Antitrypsin	-	-	-	-	-	1.85	1.85	-	1
Binding site, turbidometry, A1-Antitrypsin	-	-	-	-	-	1.63	1.63	-	1
Roche, turbidimetry, A1-Antritrypsin	1.87	1.84	0.10	5.1	0.04	1.80	2.08	-	7
Siemens, nefelometry, A1-Antitrypsin	2.11	2.12	0.08	3.8	0.02	1.90	2.21	-	13
Siemens, turbidimetry, A1-Antitrypsin	1.72	1.72	0.04	2.4	0.02	1.67	1.77	-	5
All	1.95	1.96	0.18	9.0	0.03	1.63	2.21	-	30


Specimen S002 | A1-Antitrypsin, g/l histogram summaries in LabScala



Siemens, nephelometry, A1-Antitrypsin




Siemens, turbidimetry, A1-Antitrypsin



Specimen S002 | A2-Makroglobulin, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Siemens, nefelometry, A2-Macroglobulin	3.02	3.03	0.05	1.7	0.02	2.94	3.07	-	6
All	3.02	3.03	0.05	1.7	0.02	2.94	3.07	-	6



Specimen S002 | A2-Makroglobulin, g/l| histogram summaries in LabScala



Specimen S002 | Albumin, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Albumin chemical	64.4	63.8	6.3	9.8	1.6	53.0	74.6	-	16
Albumin immunochemical	64.7	65.1	3.4	5.3	0.8	55.0	72.7	-	19
All	64.5	65.0	4.9	7.5	0.8	53.0	74.6	-	35





Specimen S002 | Albumin, g/l histogram summaries in LabScala



Specimen S002 | Ceruloplasmin, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, Ceruloplasmin	0.585	0.585	0.007	1.2	0.005	0.580	0.590	-	2
Roche, turbidimetry, Ceruloplasmin	0.442	0.443	0.012	2.7	0.004	0.419	0.460	-	8
Siemens, nephelometry, Ceruloplasmin	0.665	0.666	0.033	4.9	0.011	0.600	0.707	-	9
Siemens, turbidimetry, Ceruloplasmin	-	-	-	-	-	0.560	0.560	-	1
All	0.562	0.585	0.108	19.1	0.024	0.419	0.707	-	20





Specimen S002 | Ceruloplasmin, g/l histogram summaries in LabScala





Specimen S002 | Complement C3, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, C3	2.467	2.470	0.041	1.7	0.017	2.410	2.520	-	6
Beckman Coulter, turbidimetry, C3	2.305	2.305	0.078	3.4	0.055	2.250	2.360	-	2
Roche, turbidimetry, C3	2.475	2.480	0.127	5.1	0.042	2.290	2.660	-	9
Siemens, nefelometry, C3	2.422	2.430	0.104	4.3	0.029	2.200	2.570	-	13
Siemens, turbidimetry, C3	2.407	2.390	0.038	1.6	0.022	2.380	2.450	-	3
Thermo Scientific, turbidimetry, C3	-	-	-	-	-	2.400	2.400	-	1
All	2.435	2.430	0.101	4.1	0.017	2.200	2.660	-	34

Specimen S002 | Complement C3, g/l histogram summaries in LabScala











Specimen S002 | Complement C4, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, C4	0.488	0.487	0.012	2.4	0.005	0.470	0.500	-	6
Beckman Coulter, turbidimetry, C4	-	-	-	-	-	0.530	0.530	-	1
Roche, turbidimetry, C4	0.458	0.460	0.026	5.6	0.009	0.420	0.510	-	9
Siemens, nefelometry, C4	0.490	0.490	0.029	6.0	0.008	0.440	0.540	-	13
Siemens, turbidimetry, C4	0.477	0.480	0.006	1.2	0.003	0.470	0.480	-	3
Thermo Scientific, turbidimetry, C4	-	-	-	-	-	0.510	0.510	-	1
All	0.481	0.480	0.028	5.9	0.005	0.420	0.540	-	33

Specimen S002 | Complement C4, g/l histogram summaries in LabScala







Specimen S002 | Haptoglobin, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, Haptoglobin	1.35	1.35	0.04	3.2	0.02	1.30	1.40	-	6
Beckman Coulter, turbidimetry, Haptoglobin	-	-	-	-	-	1.28	1.28	-	1
Roche, turbidimetry, Haptoglobin	1.33	1.35	0.05	3.7	0.02	1.25	1.42	-	11
Siemens, nefelometry, Haptoglobin	1.34	1.33	0.05	3.5	0.01	1.29	1.45	-	11
Siemens, turbidimetry, Haptoglobin	1.27	1.27	0.04	2.8	0.02	1.23	1.32	-	5
All	1.33	1.32	0.05	3.9	<0.01	1.23	1.45	-	34

Specimen S002 | Haptoglobin, g/l histogram summaries in LabScala






Specimen S002 | Hemopexin, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Siemens, nefelometry, Hemopexin	0.95	0.95	<0.01	0.2	<0.01	0.95	0.95	-	2
All	0.95	0.95	<0.01	0.2	<0.01	0.95	0.95	-	2





Specimen S002 | Hemopexin, g/l histogram summaries in LabScala




Specimen S002 | IgA, g/l


Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, IgA	4.52	4.53	0.09	2.0	0.03	4.32	4.61	-	10
Beckman Coulter, turbidimetry, IgA	4.31	4.31	0.13	3.1	0.10	4.21	4.40	-	2
Binding site, turbidometry, IgA	-	-	-	-	-	4.39	4.39	-	1
Roche, turbidimetry, IgA	4.49	4.49	0.10	2.3	0.03	4.33	4.67	-	12
Siemens, nefelometry, IgA	4.70	4.64	0.26	5.5	0.07	4.22	5.16	-	15
Siemens, turbidimetry, IgA	4.60	4.69	0.18	3.9	0.10	4.39	4.71	-	3
Thermo Scientific, turbidimetry, IgA	-	-	-	-	-	4.54	4.54	-	1
All	4.55	4.54	0.18	4.0	0.03	4.21	5.07	1	44

Specimen S002 | IgA, g/l histogram summaries in LabScala






All method groups Siemens, nefelometry, IgA (x_{pt} : 4.70 | Target area: 4.00-5.41 | Target: $\pm 15\%$)



All method groups Siemens, turbidimetry, IgA (x_{pt} : 4.60 | Target area: 3.91-5.29 | Target: $\pm 15\%$)









All method groups Thermo Scientific, turbidimetry, IgA

Specimen S002 | IgG, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, IgG	23.1	23.3	0.7	2.9	0.2	22.0	24.0	-	10
Beckman Coulter, turbidimetry, IgG	20.7	20.7	0.2	0.9	0.1	20.6	20.8	-	2
Binding site, turbidometry, IgG	-	-	-	-	-	22.1	22.1	-	1
Roche, turbidimetry, IgG	23.0	22.9	1.0	4.1	0.3	21.8	24.9	-	12
Siemens, nefelometry, IgG	22.6	22.4	1.1	4.8	0.3	20.9	24.6	-	18
Siemens, turbidimetry, IgG	23.1	23.1	0.1	0.6	<0.1	23.0	23.3	-	3
All	22.8	22.9	1.0	4.4	0.1	20.6	24.9	-	46

Specimen S002 | IgG, g/l histogram summaries in LabScala








Specimen S002 | IgLCKappa, total, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidometry (Light chains)	-	-	-	-	-	4.80	4.80	-	1
Roche, turbidimetry	5.82	5.82	0.26	4.5	0.18	5.64	6.01	-	2
Siemens, nefelometry	5.42	5.44	0.24	4.4	0.11	5.15	5.76	-	5
All	5.44	5.47	0.38	7.0	0.13	4.80	6.01	-	8


Specimen S002 | IgLCKappa, total, g/l histogram summaries in LabScala



Specimen S002 | IgLCKappa, free, mg/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Freelite, nephelometry	25.02	24.30	2.18	8.7	0.89	23.20	29.00	-	6
Freelite, turbidimetry	21.63	20.39	2.63	12.1	0.88	19.00	26.18	-	9
Siemens, nephelometry	13.84	13.75	0.59	4.3	0.17	13.10	15.10	-	12
Siemens turbidimetry	14.97	14.97	0.13	0.9	0.07	14.84	15.10	-	3
All	18.53	17.05	5.00	27.0	0.91	13.10	29.00	-	30




Specimen S002 | IgLCKappa, free, mg/l histogram summaries in LabScala



Specimen S002 | IgLCLambda, total, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidometry (Light chains)	-	-	-	-	-	2.56	2.56	-	1
Roche, turbidimetry	3.31	3.31	0.19	5.9	0.14	3.17	3.44	-	2
Siemens, nefelometry	2.84	2.83	0.13	4.6	0.06	2.68	3.04	-	5
All	2.92	2.86	0.28	9.7	0.10	2.56	3.44	-	8





Specimen S002 | IgLCLambda, total, g/l histogram summaries in LabScala



Specimen S002 | IgLCLambda, free, mg/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Freelite, nephelometry	21.80	22.10	1.30	6.0	0.53	19.60	23.00	-	6
Freelite, turbidimetry	19.70	20.20	1.88	9.5	0.59	16.29	22.40	-	10
Siemens, nephelometry	26.29	25.95	2.17	8.3	0.63	22.70	30.30	-	12
Siemens turbidimetry	32.50	32.81	2.16	6.6	1.25	30.20	34.48	-	3
All	23.90	22.90	4.45	18.6	0.80	16.29	34.48	-	31





Specimen S002 | IgLCLambda, free, mg/l histogram summaries in LabScala






Specimen S002 | IgM, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, IgM	1.79	1.80	0.05	2.8	0.02	1.71	1.88	-	10
Beckman Coulter, turbidimetry, IgM	1.69	1.69	0.06	3.8	0.05	1.64	1.73	-	2
Binding site, turbidometry, IgM	-	-	-	-	-	1.79	1.79	-	1
Roche, turbidimetry, IgM	1.79	1.80	0.04	2.2	0.01	1.72	1.86	-	12
Siemens, nefelometry, IgM	1.92	1.91	0.08	4.0	0.02	1.80	2.11	-	14
Siemens, turbidimetry, IgM	1.83	1.86	0.08	4.4	0.05	1.74	1.89	-	3
Thermo Scientific, turbidimetry, IgM	-	-	-	-	-	1.77	1.77	-	1
All	1.83	1.82	0.08	4.2	0.01	1.64	2.03	1	43

Specimen S002 | IgM, g/l histogram summaries in LabScala











Specimen S002 | Prealbumin / Transtyrelin, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, Prealbumin	-	-	-	-	-	0.360	0.360	-	1
Beckman Coulter, turbidimetry, Prealbumin	0.365	0.365	0.007	1.9	0.005	0.360	0.370	-	2
Binding site, turbidometry, Prealbumin	0.342	0.342	0.013	3.7	0.009	0.333	0.351	-	2
Roche, turbidimetry, Prealbumin	0.358	0.355	0.011	3.0	0.004	0.350	0.381	-	8
Siemens, nefelometry, Prealbumin	0.391	0.390	0.040	10.1	0.013	0.316	0.452	-	9
Siemens, turbidimetry, Prealbumin	0.355	0.355	0.007	2.0	0.005	0.350	0.360	-	2
All	0.366	0.360	0.025	6.8	0.005	0.316	0.430	1	24

Specimen S002 | Prealbumin / Transtyrelin, g/l histogram summaries in LabScala






Specimen S002 | Retinol Binding Prot, mg/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Siemens, nefelometry, Retinol binding protein	61.9	61.0	1.8	2.9	0.8	60.0	63.9	-	5
All	61.9	61.0	1.8	2.9	0.8	60.0	63.9	-	5


Specimen S002 | Retinol Binding Prot, mg/l histogram summaries in LabScala




Specimen S002 | Transferrin, g/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidimetry, Transferrin	4.30	4.29	0.12	2.7	0.04	4.17	4.54	-	9
Beckman Coulter, turbidimetry, Transferrin	4.33	4.33	<0.01	0.2	<0.01	4.32	4.33	-	2
Roche, turbidimetry, Transferrin	4.48	4.50	0.20	4.4	0.05	3.93	4.82	-	14
Siemens, nefelometry, Transferrin	4.18	4.12	0.22	5.3	0.08	3.96	4.67	-	8
Siemens, turbidimetry, Transferrin	4.30	4.38	0.15	3.4	0.09	4.13	4.39	-	3
All	4.34	4.34	0.21	4.8	0.03	3.93	4.82	-	36

Specimen S002 | Transferrin, g/l histogram summaries in LabScala










Specimen S002 | Transferrin receptor, mg/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Randox, turbidimetry, TfR	1.2	1.1	0.2	14.1	<0.1	1.1	1.3	-	3
Roche, turbidimetry, TfR	3.8	3.8	0.3	6.8	<0.1	3.6	4.4	-	7
Siemens, nefelometry, TfR	1.5	1.5	<0.1	3.0	<0.1	1.5	1.6	-	9
Thermo Scientific, turbidimetry, TfR	-	-	-	-	-	1.3	1.3	-	1
Trimero Diagnostics (TfR)	-	-	-	-	-	1.5	1.5	-	1
All	2.2	1.5	1.2	52.8	0.3	1.1	4.4	-	21

Specimen S002 | Transferrin receptor, mg/l histogram summaries in LabScala








Specimen S002 | IgLKappa/Lambda, total, Ratio

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Abbott, turbidometry	-	-	-	-	-	1.88	1.88	-	1
Roche, turbidimetry	-	-	-	-	-	1.78	1.78	-	1
Siemens, nefelometry	1.94	1.95	0.14	7.0	0.06	1.72	2.09	-	5
All	1.91	1.94	0.13	6.7	0.05	1.72	2.09	-	7





Specimen S002 | IgLKappa/Lambda, total, Ratio| histogram summaries in LabScala



Specimen S002 | IgLCKappa/Lambda, free, Ratio

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Freelite, nephelometry	1.16	1.11	0.15	12.6	0.07	1.01	1.33	-	5
Freelite, turbidimetry	1.08	1.05	0.17	15.3	0.06	0.90	1.34	-	7
Siemens, nephelometry	0.52	0.52	0.04	7.1	0.01	0.45	0.56	-	6
Siemens, turbidometry	0.46	0.45	0.03	7.4	0.02	0.43	0.50	-	3
All	0.85	0.94	0.33	39.3	0.07	0.43	1.34	-	21

Specimen S002 | IgLCKappa/Lambda, free, Ratio| histogram summaries in LabScala



Report info**Participants**

63 participants from 17 countries.

Report info

Your own result should be compared to others using the same method.

Assigned values (x_{pt} , target values) are means of the results where results deviating more than +/- 3*standard deviation from the median are removed. The standard uncertainty (u) of

the assigned value is reported as standard error of the mean (SEM). Additionally, if the measurement uncertainty of the target value is large an automatic text is printed on the report:

"The uncertainty of the assigned value is not negligible, and evaluations could be affected."

In case the client's result is the only one in the method group, no assigned value will be calculated, no target area shown, and no statistics calculated. In case there are only a few

results in the client's own method group, the result can be compared to all method mean or to a group that is similar to the own method.

Results reported with < or >-signs cannot be included in the statistics.

For information on report interpretation and performance evaluation, please see the " EOAS Interpretation guidelines" LabScala User instructions (top right corner ?Help link).

External Quality Assessment Scheme

Proteins, Immunochemical Determinations Round 1, 2023

Specimens

Sample S001 (LQ735523011) and sample S002 (LQ735523012) were commercial liquid human serum samples.

Based on the previous tests and the results of this round, the samples are homogeneous, stable and suitable for the external quality assessment scheme.

The materials were sent without temperature control packaging.

Report info

Please see the description of the data analysis on the last page of the laboratory-specific histogram and Global report. It is important to read the Final report first, because it contains important information of the samples and results in each round.

Comments – EQA Coordinator

Albumin is divided into chemical and immunochemical groups, and other analytes into reagent manufacturer-specific turbidimetric and nephelometric groups. Nine results that differed from the other results were excluded from the result processing. We kindly ask customers to pay attention to result filling in LabScala, so that the selected unit for the analyte in LabScala is correct.

The distribution of results for ceruloplasmin is bimodal, with Roche's results being the lowest and Siemens' results being the highest in both samples. This is more visible in sample S001. Roche transferrin receptor results are 2-fold compared to Randox turbidometry, Siemens nephelometry, Thermo Scientific and Trimero Diagnostics results due to different standardizations.

End of report

2023-02-14

FINAL REPORT

Product no. 2230

Samples sent 2023-01-23
Round closed 2023-02-13
Report released 2023-02-14

Request for correction

Typing errors in laboratory's result forms are on laboratory's responsibility. Labquality accepts responsibility only for result processing.

Requests must be notified by writing within three weeks from the date of this letter.

Authorized by

EQA Coordinator
Päivi Ranta
paivi.ranta@labquality.fi

Expert

Kari Åkerman, Petra Anttila
Seinäjoki Central Hospital, Finland

Labquality Oy

Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200
Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.com



Copyright © Labquality Oy

Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copyright. Labquality prohibits any use of its name, or reference to Labquality EQA program, or material in this report in any advertising, brochures or other commercial publications. Labquality EQA data do not necessarily indicate the superiority of instruments, reagents, testing equipments or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of equipments or materials may be deceptive and misleading. Proficiency test results are handled confidentially. Labquality will not issue any statements to third parties of the performance of laboratories in external quality assessment schemes unless otherwise agreed.