

External Quality Assessment Scheme

Prostate specific antigen Round 1, 2023

Specimens

Please find enclosed 2 liquid pooled human sera samples S001 and S002, each 1 mL. No preservatives are added.

Caution

Quality control specimens derived from human blood must be handled with the same care as patient samples, i.e. as potential transmitters of serious diseases. The specimens are found to be HBsAg, HCVAb and HIVAgAb negative when tested with licensed reagents, but no known test method can offer complete assurance that the specimens will not transmit these or other infectious diseases.

Examinations

Total PSA
Free PSA
Complexed PSA
Free/total PSA ratio

Storage and use

Samples are ready to use. Please analyse the samples as soon as possible. If you analyse the samples later than on the day of the sample arrival, please store them frozen. A little before analysing, thaw, allow the samples to reach the room temperature and mix until homogenous. Analyse as patient samples.

Result reporting

Please enter the results and methods via LabScala (www.labscala.com). If you cannot find your instrument or reagent from the registry, please contact the EQA Coordinator.

S001



S002



2023-02-06

INSTRUCTIONS

Product no. 2226
LQ741123011-012/FI
Subcontracting: Sample preparation,
sample pretesting

If the kit is incomplete or contains damaged specimens, please report immediately to info@labquality.fi.

The results should be reported no later than **March 2, 2023.**

Inquiries

EQA Coordinator
Päivi Ranta
paivi.ranta@labquality.fi


Labquality Oy

Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200
Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.com





Participants

34 participants from 9 countries.

Report info

Assigned value (target value) calculation and its uncertainty

Your own result should be compared to others using the same method.

The assigned values (X_{rob}) are calculated according to the robust procedure described in the standard ISO 13528 (Statistical methods for use in proficiency testing by interlaboratory comparisons, Annex C, Algorithm A).

The standard uncertainty of the assigned value is expressed as $1.25 \times$ the standard error of mean (SEM) and marked as "u" in numerical summary. Due to its iterative mode algorithm A adds the uncertainty of the assigned value and with this factor we want to adjust uncertainty accordingly.

In case there are 2-12 results in a method group, the robust calculation is not used but assigned values (X_{pt}) are means of the results where results deviating more than $\pm 3 \times$ standard deviation from the median are removed. The standard uncertainty (u) of the assigned value is reported as standard error of the mean (SEM). Additionally, if the measurement uncertainty of the target value is too large ($u > 0.1 \times$ maximum allowable error) an automatic text is printed on the report: "The uncertainty of the assigned value is not negligible, and evaluations could be affected."

Please notice also that for groups that have only 1 result only the client's own result is shown. No target value (except for reference method values) is calculated, no target areas are shown.

Z score

In case there are 2-5 results in a method group, no z-score is calculated, and a text is printed on the report: "Due to the small number of results, the z score is not calculated." In case there are 6-12 results, the report has a text: "Z score is uncertain due to the small number of observations."

Results reported with $< \text{tai} >$ -signs cannot be included in the statistics.

For information on report interpretation and performance evaluation, please see the "EQAS Interpretation guidelines" in LabScala User instructions (top right corner ?Help link).

NUMERICAL SUMMARY
Prostate specific antigen 2023/01

Analyte	Method group	x	med	s	CV%	u	Min	Max	Number
Sample 001									
PSA, µg/l									
	Abbott Architect	17.90	17.9	0.66	3.7	0.38	17.3	18.6	3
	Autobio AutoLumo	17.31	17.0	1.21	7.0	0.60	16.2	19.0	4
	bioMerieux Vidas	22.45	22.5	-	-	-	-	-	1
	Brahms Kryptor	23.20	23.2	0.00	0.0	0.00	23.2	23.2	2
	maccura	16.14	17.2	2.79	17.3	1.06	10.2	18.2	7
	Roche cobas e, Elecsys, & Modular E	18.42	18.4	1.01	5.5	0.28	16.6	20.2	20
	Siemens Advia Centaur & Atellica	17.44	17.0	1.35	7.7	0.51	16.4	20.4	7
	All	18.01	18.1	1.39	7.7	0.21	10.2	23.2	44
PSA free, µg/L									
	Abbott Architect	1.420	1.42	-	-	-	-	-	1
	Brahms Kryptor	1.730	1.73	0.042	2.5	0.030	1.70	1.76	2
	Roche cobas e, Elecsys, & Modular E	1.451	1.47	0.093	6.4	0.033	1.28	1.58	8
	Siemens Advia Centaur & Atellica	1.665	1.67	0.064	3.8	0.045	1.62	1.71	2
	Siemens Dimension & Vista	1.360	1.36	-	-	-	-	-	1
	All	1.513	1.48	0.163	10.8	0.044	1.28	1.76	14
PSA free/total, %									
	Abbott Architect	8.00	8.0	-	-	-	-	-	1
	Brahms Kryptor	7.45	7.5	0.21	2.8	0.15	7.3	7.6	2
	Calculated	8.00	8.0	-	-	-	-	-	1
	Roche cobas e, Elecsys, & Modular E	7.58	7.6	0.18	2.4	0.08	7.3	7.8	5
	All	7.64	7.6	0.29	3.8	0.10	7.3	8.0	9
PSA, complexed, µg/L									
	Roche cobas e, Elecsys, & Modular E	16.87	17.3	1.19	7.1	0.69	15.5	17.8	3
	Siemens Dimension & Vista	26.78	26.8	-	-	-	-	-	1
	All	19.35	17.6	5.72	29.6	2.86	15.5	26.8	4
Sample 002									
PSA, µg/l									
	Abbott Architect	3.32	3.3	0.14	4.1	0.08	3.2	3.4	3
	Autobio AutoLumo	3.46	3.4	0.33	9.6	0.15	3.1	3.8	5
	bioMerieux Vidas	4.21	4.2	-	-	-	-	-	1
	Brahms Kryptor	4.50	4.5	0.06	1.3	0.04	4.5	4.5	2
	maccura	3.27	3.4	0.44	13.3	0.16	2.4	3.8	7
	Roche cobas e, Elecsys, & Modular E	3.49	3.5	0.18	5.2	0.05	3.1	3.8	20
	Siemens Advia Centaur & Atellica	3.38	3.3	0.24	7.1	0.09	3.2	3.9	7
	All	3.46	3.4	0.27	7.7	0.04	2.4	4.5	45
PSA free, µg/L									
	Abbott Architect	0.300	0.30	-	-	-	-	-	1
	Brahms Kryptor	0.335	0.34	0.007	2.1	0.005	0.33	0.34	2
	Roche cobas e, Elecsys, & Modular E	0.314	0.32	0.015	4.9	0.005	0.29	0.34	8
	Siemens Advia Centaur & Atellica	0.345	0.35	0.007	2.0	0.005	0.34	0.35	2
	Siemens Dimension & Vista	0.240	0.24	-	-	-	-	-	1
	All	0.318	0.32	0.023	7.3	0.006	0.24	0.35	14
PSA free/total, %									
	Abbott Architect	9.00	9.0	-	-	-	-	-	1
	Brahms Kryptor	7.45	7.5	0.07	0.9	0.05	7.4	7.5	2
	Calculated	9.00	9.0	-	-	-	-	-	1

NUMERICAL SUMMARY Prostate specific antigen 2023/01, Sample 002

Analyte	Method group	x	med	s	CV%	u	Min	Max	Number
PSA free/total, %									
	Roche cobas e, Elecsys, & Modular E	8.69	8.8	0.28	3.2	0.13	8.4	9.0	5
	Siemens Advia Centaur & Atellica	10.00	10.0	0.00	0.0	0.00	10.0	10.0	2
	All	8.76	8.9	0.95	10.8	0.29	7.4	10.0	11
PSA, complexed, µg/L									
	Roche cobas e, Elecsys, & Modular E	3.14	3.2	0.20	6.4	0.12	2.9	3.3	3
	Siemens Dimension & Vista	5.14	5.1	-	-	-	-	-	1
	All	3.64	3.3	1.15	31.5	0.57	2.9	5.1	4

Participants

34 participants from 9 countries.

Report info

Assigned value (target value) calculation and its uncertainty

Your own result should be compared to others using the same method.

The assigned values (X_{rob}) are calculated according to the robust procedure described in the standard ISO 13528 (Statistical methods for use in proficiency testing by interlaboratory comparisons, Annex C, Algorithm A).

The standard uncertainty of the assigned value is expressed as $1.25 \times$ the standard error of mean (SEM) and marked as "u" in numerical summary. Due to its iterative mode algorithm A adds the uncertainty of the assigned value and with this factor we want to adjust uncertainty accordingly.

In case there are 2-12 results in a method group, the robust calculation is not used but assigned values (X_{pt}) are means of the results where results deviating more than $\pm 3 \times$ standard deviation from the median are removed. The standard uncertainty (u) of the assigned value is reported as standard error of the mean (SEM). Additionally, if the measurement uncertainty of the target value is too large ($u > 0.1 \times$ maximum allowable error) an automatic text is printed on the report: "The uncertainty of the assigned value is not negligible, and evaluations could be affected."

Please notice also that for groups that have only 1 result only the client's own result is shown. No target value (except for reference method values) is calculated, no target areas are shown.

Z score

In case there are 2-5 results in a method group, no z-score is calculated, and a text is printed on the report: "Due to the small number of results, the z score is not calculated." In case there are 6-12 results, the report has a text: "Z score is uncertain due to the small number of observations."

Results reported with $< \text{tai} >$ -signs cannot be included in the statistics.

For information on report interpretation and performance evaluation, please see the "EQAS Interpretation guidelines" in LabScala User instructions (top right corner ?Help link).

External Quality Assessment Scheme

Prostate specific antigen Round 1, 2023

Specimens

Sample S001 (LQ741123011) and sample S002 (LQ741123012) were prepared by dilution of plasma pools with elevated PSA level into a serum matrix.

Based on the previous tests and the results of this round, the samples are homogeneous, stable and suitable for the external quality assessment scheme.

The materials were sent without temperature control packaging.

Report info

Please see the description of the data analysis on the last page of the laboratory-specific histogram and Global report. It is important to read the Final report first, because it contains important information of the samples and results in each round.

Comments - Expert

In sample S001 the total PSA level was elevated in all method groups (mean total PSA of all method groups 18 µg/L) which in early diagnosis setting would likely lead to further evaluations of the patient. For complexed PSA all results in two method groups were also elevated (range 14 – 27 µg/L).

In sample S002 the mean of all total PSA method groups was 3.5 µg/L (range of the mean levels 3.3 – 4.5 µg/L). Thus, total PSA level was close to the commonly used prostate cancer early diagnosis cut-off 4 µg/L. The proportion of free PSA was fairly low (the mean of all total PSA method groups 8.8 %), which can indicate elevated risk of prostate cancer. Thus, these results may lead to further evaluations of the patient. For complexed PSA there were results in two method groups (range 2.6 – 5.1 µg/L).

End of report

2023-03-10

FINAL REPORT

Product no. 2226
Subcontracting: Sample preparation,
sample pretesting

Samples sent	2023-02-06
Round closed	2023-03-02
Final report	2023-03-10

Request for correction

Typing errors in laboratory's result forms are on laboratory's responsibility. Labquality accepts responsibility only for result processing. Requests must be notified by writing within three weeks from the date of this letter.

Authorized by

EQA Coordinator
Päivi Ranta
paivi.ranta@labquality.fi

Expert

PhD Jari Leinonen
HUSLAB, Helsinki, Finland.

Labquality Oy

Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200
Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.com



Copyright © Labquality Oy

Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copyright. Labquality prohibits any use of its name, or reference to Labquality EQA program, or material in this report in any advertising, brochures or other commercial publications. Labquality EQA data do not necessarily indicate the superiority of instruments, reagents, testing equipments or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of equipments or materials may be deceptive and misleading. Proficiency test results are handled confidentially. Labquality will not issue any statements to third parties of the performance of laboratories in external quality assessment schemes unless otherwise agreed.