

External Quality Assessment Scheme

Antibody Screening and Compatibility Testing Round 1, 2023

Specimens

Please find enclosed 2 whole blood samples S001 and S002, each 4 mL and four 15% red cell suspension in preservation medium S003, S004, S005 and S006, each 3 mL.

Caution

Quality control specimens derived from human blood must be handled with the same care as patient samples, i.e. as potential transmitters of serious diseases. The specimens are found to be HBsAg, HCVAb and HIVAgAb negative when tested with licensed reagents, but no known test method can offer complete assurance that the specimens will not transmit these or other infectious diseases.

Examinations

Antibody screening
Compatibility testing

Storage and use

Analyse the samples on the arrival day, if possible. Otherwise store in a refrigerator (+2 ... 8 °C) until the tests can be performed. The samples will be stable for 4 weeks from the date of this letter.

Samples S001 and S002 will be screened for unexpected antibodies. Samples S003, S004, S005 and S006 are for compatibility testing. The cells should be centrifuged to the bottom of the tube and the laboratory should prepare a suspension appropriate to the method. Compatibility tests simulate situation where four units of donor red blood cells (samples S003, S004, S005 and S006) are each crossmatched for two patients (samples S001 and S002).

Result reporting

Please enter the results and methods via LabScala (www.labscala.com). The compatibility tests should always be answered according to the serological tests. At the request of customers, the possibility of responding to a possibly identified antibody has been added to this round. This section is completely optional and will not be scored. Both reaction strength and interpretation are scored for the antibody screening. Reaction strengths are not scored in compatibility testing, only crossmatches are scored.

S001



S004



S002



S005



S003



S006



2023-02-27

INSTRUCTIONS

Product no. 4460
LQ714323011-012/BE
LQ714223013-016/BE

If the kit is incomplete or contains damaged specimens, please report immediately to info@labquality.fi.

The results should be reported no later than
March 20, 2023.

Inquiries

EQA Coordinator
Iida Silvo
iida.silvo@labquality.fi

Labquality Oy

Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200
Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.com




Client report

	No of participants	No of responded participants	Response percentage
Antibody screening and compatibility testing, February, 1-2023	142	140	98.6 %

Summary

Summary	Own score	Max score	Own success rate
Sample S001	44	44	100 %
Sample S002	44	44	100 %
Average:			100 %


Sample S001







ANTIBODY SCREENING

Antibody screening	Method	Reaction strength	n	Reaction strength Score	Interpretation: Positive	Interpretation: Positive Score
	DiaMed column agglutination (Bio-Rad)	2+	7	1	7	2
		⊙ 3+	65	2	⊙ 65	2
		4+	11	2	11	2
	Total:		83			

Sample S001 Antibody identification



Antibody identification	Result	Result count
	Anti-c	1
	Anti-C	1
	⊙ Anti-D	43
	Anti-E	1
	Anti-Fya	4
	Anti-Fyb	4
	Anti-s	3
	Anti-S	1
	Total:	58




Compatibility 1 / 3	Method	Reaction strength	Interpretation: Compatible	Interpretation: Incompatible	Interpretation: Compatible Score	Interpretation: Incompatible Score	n
	DiaMed column agglutination (Bio-Rad)	3+		9		10	9
		4+		70		10	70
		Negative	1		0		1
	Total:						80

Compatibility 1 / 4	Method	Reaction strength	Interpretation: Incompatible	Interpretation: Incompatible Score	n
	DiaMed column agglutination (Bio-Rad)	2+	6	10	6
		3+	68	10	68
		4+	6	10	6
	Total:				80

Compatibility 1 / 5	Method	Reaction strength	Interpretation: Compatible	Interpretation: Compatible Score	n
	DiaMed column agglutination (Bio-Rad)	Negative	80	10	80
	Total:				80





Compatibility 1 / 6	Method	Reaction strength	Interpretation: Incompatible	Interpretation: Incompatible Score	n
	DiaMed column agglutination (Bio-Rad)	3+	14	10	14
		4+	66	10	66
	Total:				80

Sample S002



ANTIBODY SCREENING

Antibody screening	Method	Reaction strength	n	Reaction strength Score	Interpretation: Negative	Interpretation: Negative Score
	DiaMed column agglutination (Bio-Rad)	⊙ Negative	83	2	⊙ 82	2
	Total:		83			



Compatibility 2 / 3	Method	Reaction strength	Interpretation: Compatible	Interpretation: Incompatible	Interpretation: Compatible Score	Interpretation: Incompatible Score	n
	DiaMed column agglutination (Bio-Rad)	<input checked="" type="radio"/> Negative	<input checked="" type="radio"/> 80	1	10	5	81
	Total:						81

Compatibility 2 / 4	Method	Reaction strength	Interpretation: Compatible	Interpretation: Incompatible	Interpretation: Compatible Score	Interpretation: Incompatible Score	n
	DiaMed column agglutination (Bio-Rad)	<input checked="" type="radio"/> Negative	<input checked="" type="radio"/> 80	1	10	5	81
	Total:						81

Compatibility 2 / 5	Method	Reaction strength	Interpretation: Compatible	Interpretation: Incompatible	Interpretation: Compatible Score	Interpretation: Incompatible Score	n
	DiaMed column agglutination (Bio-Rad)	<input checked="" type="radio"/> Negative	<input checked="" type="radio"/> 80	1	10	5	81
	Total:						81

Compatibility 2 / 6	Method	Reaction strength	Interpretation: Compatible	Interpretation: Incompatible	Interpretation: Compatible Score	Interpretation: Incompatible Score	n
	DiaMed column agglutination (Bio-Rad)	<input checked="" type="radio"/> Negative	<input checked="" type="radio"/> 80	1	10	5	81
	Total:						81

Report Info

PARTICIPANTS

Altogether 142 laboratories from 12 countries participated in this EQA round.

REPORT INFO

The principles of the scoring will be as follows:

Antibody screening

1. Reaction strengths

Correct agglutination reaction and grade: 2 points/reaction

Correct reaction, but large difference in reaction grade from the expected: 1 point (eg. + for an expected +++)

Wrong reaction: 0 points / reaction

2. Interpretation

Correct interpretation 2 points / specimen

Wrong interpretation 0 points / specimen, maximum points / specimen = 4

Compatibility testing

Correct compatibility 10 points / specimen

Correct incompatibility 10 points / specimen

False compatibility 0 points / specimen

False incompatibility 5 points / specimen

There are 8 compatibility tests in the survey-> maximum points / survey = 80


GLOBAL REPORT

	No of participants	No of responded participants	Response percentage
Antibody screening and compatibility testing, February, 1-2023	142	140	98.6 %

Summary

Summary	AVR success rate
Sample S001	99.7 %
Sample S002	99.7 %
Average:	99.7 %


Sample S001



ANTIBODY SCREENING

Antibody screening	Method	Reaction strength	n	Reaction strength Score	Interpretation: Positive	Interpretation: Positive Score
	DiaMed column agglutination (Bio-Rad)	2+	7	1	7	2
		3+	65	2	65	2
		4+	11	2	11	2
	DiaPro/Across Gel	2+	1	2	1	2
		3+	1	2	1	2
		4+	1	2	1	2
	Grifols DG Gel	3+	13	2	13	2
		4+	3	2	3	2
	LISS, tube method	3+	1	2	1	2
	Microplate	4+	3	2	3	2
	ORTHO BioVue column agglutinaatio	3+	14	2	14	2
		4+	21	2	21	2
	Saline, tube method	2+	1	2	1	2
	Total:		142			

Sample S001 Antibody identification







- Anti-c
- Anti-C
- Anti-D
- Anti-E
- Anti-Fya
- Anti-Fyb
- Anti-s
- Anti-S

Antibody identification	Result	Result count
	Anti-c	1
	Anti-C	1
	Anti-D	43
	Anti-E	1
	Anti-Fya	4

LABQUALITY Antibody screening and compatibility testing, February, 1-2023

	Anti-Fyb	4
	Anti-s	3
	Anti-S	1
	Total:	58



Compatibility 1 / 3	Method	Reaction strength	Interpretation: Compatible	Interpretation: Incompatible	Interpretation: Compatible Score	Interpretation: Incompatible Score	n
	DiaMed column agglutination (Bio-Rad)	3+		9		10	9
		4+		70		10	70
		Negative	1		0		1
	DiaPro/Across Gel	4+		4		10	4
		3+		2		10	2
	Grifols DG Gel	4+		13		10	13
		4+		37		10	37
		Mixed field		1		10	1
	Saline, tube method	3+		1		10	1
	Total:						138

Compatibility 1 / 4	Method	Reaction strength	Interpretation: Incompatible	Interpretation: Incompatible Score	n
	DiaMed column agglutination (Bio-Rad)	2+	6	10	6
		3+	68	10	68
		4+	6	10	6
	DiaPro/Across Gel	3+	3	10	3
		4+	1	10	1
	Grifols DG Gel	3+	13	10	13
		4+	2	10	2
	ORTHO BioVue column agglutinaatio	3+	13	10	13


LABQUALITY Antibody screening and compatibility testing, February, 1-2023

		4+	24	10	24
		Mixed field	1	10	1
	Saline, tube method	2+	1	10	1
	Total:				138

Compatibility 1 / 5	Method	Reaction strength	Interpretation: Compatible	Interpretation: Compatible Score	n
	DiaMed column agglutination (Bio-Rad)	Negative	80	10	80
	DiaPro/Across Gel	Negative	4	10	4
	Grifols DG Gel	Negative	15	10	15
	ORTHO BioVue column agglutinaatio	Negative	38	10	38
	Saline, tube method	Negative	1	10	1
	Total:				138





Compatibility 1 / 6	Method	Reaction strength	Interpretation: Incompatible	Interpretation: Incompatible Score	n
	DiaMed column agglutination (Bio-Rad)	3+	14	10	14
		4+	66	10	66
	DiaPro/Across Gel	4+	4	10	4
	Grifols DG Gel	3+	2	10	2
		4+	13	10	13
	ORTHO BioVue column agglutinaatio	3+	16	10	16
		4+	21	10	21
		Mixed field	1	10	1
	Saline, tube method	3+	1	10	1
	Total:				138

Sample S002



ANTIBODY SCREENING

Antibody screening	Method	Reaction strength	n	Reaction strength Score	Interpretation: Negative	Interpretation: Negative Score
	DiaMed column agglutination (Bio-Rad)	Negative	83	2	82	2
	DiaPro/Across Gel	Negative	2	2	2	2
	Grifols DG Gel	Negative	16	2	16	2
	LISS, tube method	Negative	1	2	1	2
	Microplate	Negative	3	2	3	2
	ORTHO BioVue column agglutination	Negative	35	2	35	2
	Saline, tube method	Negative	1	2	1	2
	Total:		141			



■ Negative

■ Negative

Compatibility 2 / 3	Method	Reaction strength	Interpretation: Compatible	Interpretation: Incompatible	Interpretation: Compatible Score	Interpretation: Incompatible Score	n
	DiaMed column agglutination (Bio-Rad)	Negative	80	1	10	5	81
	DiaPro/Across Gel	Negative	3		10		3
	Grifols DG Gel	Negative	15		10		15
	ORTHO BioVue column agglutinaatio	Negative	37		10		37
	Saline, tube method	Negative	1		10		1
	Total:						137

Compatibility 2 / 4	Method	Reaction strength	Interpretation: Compatible	Interpretation: Incompatible	Interpretation: Compatible Score	Interpretation: Incompatible Score	n
	DiaMed column agglutination (Bio-Rad)	Negative	80	1	10	5	81
	DiaPro/Across Gel	Negative	3		10		3
	Grifols DG Gel	Negative	15		10		15
	ORTHO BioVue column agglutinaatio	Negative	37		10		37
	Saline, tube method	Negative	1		10		1
	Total:						137

Compatibility 2 / 5	Method	Reaction strength	Interpretation: Compatible	Interpretation: Incompatible	Interpretation: Compatible Score	Interpretation: Incompatible Score	n
	DiaMed column agglutination (Bio-Rad)	Negative	80	1	10	5	81
	DiaPro/Across Gel	Negative	3		10		3
	Grifols DG Gel	Negative	15		10		15
	ORTHO BioVue column agglutinaatio	Negative	37		10		37
	Saline, tube method	Negative	1		10		1
	Total:						137

Compatibility 2 / 6	Method	Reaction strength	Interpretation: Compatible	Interpretation: Incompatible	Interpretation: Compatible Score	Interpretation: Incompatible Score	n
	DiaMed column agglutination (Bio-Rad)	Negative	80	1	10	5	81
	DiaPro/Across Gel	Negative	3		10		3
	Grifols DG Gel	Negative	15		10		15
	ORTHO BioVue column agglutinaatio	Negative	37		10		37
	Saline, tube method	Negative	1		10		1
	Total:						137

Report Info

PARTICIPANTS

Altogether 142 laboratories from 12 countries participated in this EQA round.

REPORT INFO

The principles of the scoring will be as follows:

Antibody screening

1. Reaction strengths

Correct agglutination reaction and grade: 2 points/reaction

Correct reaction, but large difference in reaction grade from the expected: 1 point (eg. + for an expected +++)

Wrong reaction: 0 points / reaction

2. Interpretation

Correct interpretation 2 points / specimen

Wrong interpretation 0 points / specimen, maximum points / specimen = 4

Compatibility testing

Correct compatibility 10 points / specimen

Correct incompatibility 10 points / specimen

False compatibility 0 points / specimen

False incompatibility 5 points / specimen

There are 8 compatibility tests in the survey-> maximum points / survey = 80

External Quality Assessment Scheme

Compatibility test and antibody screening Round 1, 2023

Specimens

Sample S001 and sample S002 were human-based whole blood samples. Samples S003-S006 were red-cell suspensions.

Based on the pre-testing and the results obtained in the round, the samples were homogeneous, stable and suitable for the external quality assessment scheme. The materials were sent without any temperature control packaging.

Sample S001 (LQ714323011): Blood from A RhD negative patient, anti-D and anti-B in plasma.

Sample S002 (LQ714323012): Blood from AB RhD positive patient.

Sample S003 (LQ714223013): Cells from a donor with blood group B RhD positive.

Sample S004 (LQ714223014): Cells from a donor with blood group O RhD positive.

Sample S005 (LQ714223015): Cells from a donor with blood group A RhD negative.

Sample S006 (LQ714223016): Cells from a donor with blood group AB RhD negative.

Report info

Please see the description of the scoring on the last page of the laboratory-specific reports and global reports. It is important to read the Final report first, because it contains important information of the samples and results in each round.

Comments - Expert

Antibody screening

Sample S001: The expected result was a positive screening due to anti-D.

Sample S002: The expected result was a negative screening.

Compatibility test

Sample S001: The expected results were a negative compatibility test 1/5 and positive compatibility tests 1/3 (anti-B and anti-D and a B RhD pos donor), 1/4 (anti-D and a RhD pos donor) and 1/6 (anti-B and AB donor). Only one false compatibility 1/3 was found (Diamed)

Sample S002: The expected results were negative compatibility tests 2/3, 2/4, 2/5 and 2/6.

Comment

The results were excellent.

This scheme is designed for antibody screening and compatibility testing. However, the participants may also do the antibody identification. For patient 1 an almost 25 % of the participant reported other antibodies than the correct (anti-D), this is alarming if the participants perform antibody identification routinely.

2023-04-17

FINAL REPORT

Product no. 4460

Samples sent	2023-02-27
Round closed	2023-03-20
Final report	2023-04-17

Request for correction

Typing errors in laboratory's result forms are on laboratory's responsibility. Labquality accepts responsibility only for result processing. Requests must be notified by writing within three weeks from the date of this letter.

Authorized by

EQA Coordinator
Iida Silvo
iida.silvo@labquality.fi

Expert

MD, Head of Department
Tomi Koski
Fimlab Medical Laboratories
Finland

Labquality Oy

Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200

Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.com



Exceptions in scoring

No exceptions.

End of report

Copyright © Labquality Oy

Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copyright. Labquality prohibits any use of its name, or reference to Labquality EQA program, or material in this report in any advertising, brochures or other commercial publications. Labquality EQA data do not necessarily indicate the superiority of instruments, reagents, testing equipment or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of equipment or materials may be deceptive and misleading. Proficiency test results are handled confidentially. Labquality will not issue any statements to third parties of the performance of laboratories in external quality assessment schemes unless otherwise agreed.