

External Quality Assessment Scheme

Reticulocytes, manual methods Round 1, 2023

Specimens

Please find enclosed 1 human origin blood cell suspension S001, 3 mL.

Caution

Quality control specimens derived from human blood must be handled with the same care as patient samples, i.e. as potential transmitters of serious diseases. The specimens are found to be HBsAg, HCVAb and HIVAgAb negative when tested with licensed reagents, but no known test method can offer complete assurance that the specimens will not transmit these or other infectious diseases.

Examinations

Reticulocytes

Storage and use

Please analyze the sample on arrival day. If this is not possible, store the sample in a refrigerator (+2...+8 °C) and analyze as soon as possible. Do not freeze. Let the vial warm up in room temperature at least 15 minutes. Gently invert the vial until the cells are completely and uniformly suspended. Do not use mechanical mixer. Prepare staining and smear.

Result reporting

Please enter the results via LabScala (www.labscala.com). In balanced sampling method (Dacie & Lewis: Practical Haematology) at least 100 reticulocytes are counted in successive fields and the count of all red cells in every tenth field. The number of all red cells should be counted in at least ten fields. The calculation is as follows (an example): Number of the reticulocytes seen in 150 fields = 100; Number of all red cells present 15 fields = 300; the approximate number of all red cells in 150 fields = 3000; the reticulocyte percentage = $100/3000 \times 100 = 3.3\%$.

S001



2023-04-03

INSTRUCTIONS

Product no. 4140
LQ711323011/US

If the kit is incomplete or contains damaged specimens, please report immediately to info@labquality.fi

The results should be reported no later than
April 20, 2023.

Inquiries

EQA Coordinator
Iida Silvo
iida.silvo@labquality.fi

Labquality Oy

Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200
Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.com



External Quality Assessment Scheme

Reticulocytes, automated methods Round 1, 2023

Specimens

Please find enclosed 2 blood cell suspensions 3 mL each according to your order. Samples S002 and S003 are for Siemens Advia analyzers, S004 and S005 are for Sysmex analyzers, S006 and S007 are for Cell-Dyn 3200, 3500, 3700 and Ruby analyzers. Samples S008 and S009 are for Coulter GenS and Coulter LH750 analyzers. Samples S010 and S011 are for Cell-Dyn 4000, Sapphire and ABX analyzers. Samples S012 and S013 are for Mindray analysers.

Caution

Quality control specimens derived from human blood must be handled with the same care as patient samples, i.e. as potential transmitters of serious diseases. The specimens are found to be HBsAg, HCVAb and HIVAgAb negative when tested with licensed reagents, but no known test method can offer complete assurance that the specimens will not transmit these or other infectious diseases.

Examinations

Reticulocytes

Storage and use

Store the samples in refrigerator (+ 2...8 °C), do not freeze. Allow samples to warm up at ambient room temperature about 15 minutes. Gently invert the vials until the cells are completely suspended. Do not shake the vial or use a mechanical mixer. Analyse as patient samples.

Result reporting

Please enter the results and methods via LabScala (www.labscale.com). If you cannot find your instrument or reagent from the registry, please contact the EQA Coordinator. Results should be given in absolute values.

2023-04-03

INSTRUCTIONS

Product no. 4150-4156
LQ711323012-113/US

If the kit is incomplete or contains damaged specimens, please report immediately to info@labquality.fi.

The results should be reported no later than
April 20, 2023.

Inquiries

EQA Coordinator
Iida Silvo
iida.silvo@labquality.fi

Labquality Oy

Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200
Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.com



Reticulocytes, automated methods

S002



S008



S003



S009



S004



S010



S005



S011



S006



S012





S007



S013




Reticulocytes | Lionel 45156 Retic



	x _{pt}	sd	SEM	CV%	n
System XN	277.4 x10E9/l	23.6	1.9	8.5	147
All methods	279.1 x10E9/l	23.9	1.9	8.6	158

	x _{pt}	sd	SEM	CV%	n
System XN	310.0 x10E9/l	17.0	1.4	5.5	147
All methods	313.3 x10E9/l	20.4	1.6	6.5	158

History



Round	Sample	x _{pt}	Result	diff%	z-score
23/1	Sample S005	310.0	302.0	-2.6%	-0.47
23/1	Sample S004	277.4	274.0	-1.2%	-0.14
22/4	Sample S005	243.9	219.2	-10.1%	-1.74
22/4	Sample S004	80.8	75.3	-6.8%	-0.69
22/3	Sample S005	329.1	318.0	-3.4%	-0.63
22/3	Sample S004	284.0	296.0	4.2%	0.60
22/2	Sample S005	85.1	84.0	-1.3%	-0.15
22/2	Sample S004	195.8	183.0	-6.5%	-0.93
22/1	Sample S005	59.9	61.0	1.8%	0.16
22/1	Sample S004	243.7	241.0	-1.1%	-0.18

Report info**Participants**

180 participants from 20 countries.

Report info

Your own result should be compared to others using the same method.

Assigned values (\bar{x}_p , target values) are means of the results where results deviating more than ± 3 standard deviation from the median are removed. The standard uncertainty (u) of

the assigned value is reported as standard error of the mean (SEM). Additionally, if the measurement uncertainty of the target value is large an automatic text is printed on the report: "The uncertainty of the assigned value is not negligible, and evaluations could be affected."

In case the client's result is the only one in the method group, no assigned value will be calculated, no target area shown, and no statistics calculated. In case there are only a few results in the client's own method group, the result can be compared to all method mean or to a group that is similar to the own method.



Results reported with $<$ or $>$ -signs cannot be included in the statistics.

For information on report interpretation and performance evaluation, please see the "EOAS Interpretation guidelines" LabScala User instructions (top right corner ?Help link).

Sample S001 | Reticulocytes, %

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Reticulocytes, Balanced sampling	5.0	4.4	2.5	51.0	0.7	0.7	10.0	-	13
Reticulocytes, other manual counting	5.5	5.2	2.4	43.8	0.7	1.9	10.2	-	12
All	5.2	4.7	2.4	46.7	0.5	0.7	10.2	-	25


Sample S001 | Reticulocytes, % | histogram summaries in LabScala



Sample S002 | Reticulocytes, x10E9/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Siemens / Advia	302.3	316.0	40.7	13.5	12.3	235.4	350.0	-	11
All	302.3	316.0	40.7	13.5	12.3	235.4	350.0	-	11


Sample S002 | Reticulocytes, x10E9/l| histogram summaries in LabScala



Sample S003 | Reticulocytes, x10E9/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Siemens / Advia	200.8	209.8	22.1	11.0	6.7	163.0	229.0	-	11
All	200.8	209.8	22.1	11.0	6.7	163.0	229.0	-	11




Sample S003 | Reticulocytes, x10E9/l| histogram summaries in LabScala



Sample S004 | Reticulocytes, x10E9/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Sysmex XN	277.4	278.4	23.6	8.5	1.9	213.0	325.4	1	147
Sysmex XN-L	301.9	298.0	18.0	6.0	5.7	277.0	338.1	-	10
Sysmex XT 4000, 4000i	-	-	-	-	-	295.1	295.1	-	1
All	279.1	279.3	23.9	8.6	1.9	213.0	338.1	1	158


Sample S004 | Reticulocytes, x10E9/l| histogram summaries in LabScala




Sample S005 | Reticulocytes, x10E9/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Sysmex XN	310.0	309.2	17.0	5.5	1.4	273.0	365.4	1	147
Sysmex XN-L	357.6	353.8	9.7	2.7	3.1	347.6	374.8	-	10
Sysmex XT 4000, 4000i	-	-	-	-	-	345.1	345.1	-	1
All	313.3	310.6	20.4	6.5	1.6	273.0	374.8	1	158


Sample S005 | Reticulocytes, x10E9/l| histogram summaries in LabScala



■ All method groups ■ Sysmex XN (x_{pt} : 310.0 | Target area: 248.0-372.0 | Target: $\pm 20\%$)



■ All method groups ■ Sysmex XN-L (x_{pt} : 357.6 | Target area: 286.1-429.1 | Target: $\pm 20\%$)




■ All method groups ■ Sysmex XT 4000, 4000i

Sample S006 | Reticulocytes, x10E9/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Cell-Dyn Ruby	-	-	-	-	-	84.6	84.6	-	1
All	-	-	-	-	-	84.6	84.6	-	1


Sample S006 | Reticulocytes, x10E9/l| histogram summaries in LabScala



Sample S007 | Reticulocytes, x10E9/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Cell-Dyn Ruby	-	-	-	-	-	270.5	270.5	-	1
All	-	-	-	-	-	270.5	270.5	-	1


Sample S007 | Reticulocytes, x10E9/l| histogram summaries in LabScala



Sample S008 | Reticulocytes, x10E9/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Coulter STKS, GenS, LH750, MaxM, DxH	278.6	279.2	1.3	0.5	0.7	277.1	279.4	-	3
All	278.6	279.2	1.3	0.5	0.7	277.1	279.4	-	3


Sample S008 | Reticulocytes, x10E9/l| histogram summaries in LabScala



Sample S009 | Reticulocytes, x10E9/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Coulter STKS, GenS, LH750, MaxM, DxH	434.8	440.1	15.3	3.5	8.8	417.6	446.8	-	3
All	434.8	440.1	15.3	3.5	8.8	417.6	446.8	-	3




Sample S009 | Reticulocytes, x10E9/l| histogram summaries in LabScala



Sample S010 | Reticulocytes, x10E9/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
ABX	338.0	338.0	9.9	2.9	7.0	331.0	345.0	-	2
Alinity hq	310.0	306.0	14.4	4.7	8.3	298.0	326.0	-	3
Yumizen	-	-	-	-	-	382.0	382.0	-	1
All	331.3	328.5	30.1	9.1	12.3	298.0	382.0	-	6


Sample S010 | Reticulocytes, x10E9/l| histogram summaries in LabScala




Sample S011 | Reticulocytes, x10E9/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
ABX	179.5	179.5	7.8	4.3	5.5	174.0	185.0	-	2
Alinity hq	167.3	169.0	3.8	2.3	2.2	163.0	170.0	-	3
Yumizen	-	-	-	-	-	187.0	187.0	-	1
All	174.7	172.0	9.5	5.4	3.9	163.0	187.0	-	6


Sample S011 | Reticulocytes, x10E9/l| histogram summaries in LabScala



■ All method groups ■ ABX (x_{pt} : 179.5 | Target area: 143.6-215.4 | Target: $\pm 20\%$)



■ All method groups ■ Alinity hq (x_{pt} : 167.3 | Target area: 133.9-200.8 | Target: $\pm 20\%$)




■ All method groups ■ Yumizen

Sample S012 | Reticulocytes, x10E9/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Mindray	139.1	147.7	15.3	11.0	5.8	118.9	156.5	-	7
All	139.1	147.7	15.3	11.0	5.8	118.9	156.5	-	7


Sample S012 | Reticulocytes, x10E9/l| histogram summaries in LabScala



Sample S013 | Reticulocytes, x10E9/l

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Mindray	77.9	81.1	6.9	8.9	2.6	64.2	83.6	-	7
All	77.9	81.1	6.9	8.9	2.6	64.2	83.6	-	7

Sample S013 | Reticulocytes, x10E9/l| histogram summaries in LabScala



Report info**Participants**

180 participants from 20 countries.

Report info

Your own result should be compared to others using the same method.

Assigned values (\bar{x}_p , target values) are means of the results where results deviating more than ± 3 standard deviation from the median are removed. The standard uncertainty (u) of

the assigned value is reported as standard error of the mean (SEM). Additionally, if the measurement uncertainty of the target value is large an automatic text is printed on the report: "The uncertainty of the assigned value is not negligible, and evaluations could be affected."

In case the client's result is the only one in the method group, no assigned value will be calculated, no target area shown, and no statistics calculated. In case there are only a few results in the client's own method group, the result can be compared to all method mean or to a group that is similar to the own method.

Results reported with $<$ or $>$ -signs cannot be included in the statistics.

For information on report interpretation and performance evaluation, please see the "EOAS Interpretation guidelines" LabScala User instructions (top right corner ?Help link).

External Quality Assessment Scheme

Reticulocytes Round 1, 2023

Specimens

All samples were blood cell suspensions. Sample S001 (LQ711323011) was for manual methods. Laboratories were asked to stain the sample themselves. Samples S002 (LQ711323012) and S003 (LQ711323013) were for Siemens Advia analysers. Samples S004 (LQ711323014) and S005 (LQ711323015) were for Sysmex analysers. Samples S006 (LQ711323016) and S007 (LQ711323017) were for Cell-Dyn 3200, 3500, 3700 and Ruby analysers. Samples S008 (LQ711323018) and S009 (LQ711323019) were for Coulter analysers. Sample S010 (LQ7113230110) and S011 (LQ7113230111) were for ABX, Cell-Dyn 4000 and Sapphire analysers. Sample S012 (LQ7113230112) and S013 (LQ7113230113) were for Mindray analysers. Note! You can only compare your own results with the results of samples intended for your own analyzer.

Based on the previous tests and the results of this round, the samples were homogeneous, stable and suitable for the external quality assessment scheme.

The materials were sent without temperature control packaging.

Report info

Please see the description of the data analysis on the last page of the laboratory-specific histograms and Numerical Summary reports. It is important to read the Final report first, because it contains important information of the samples and results in each round.

Comments – EQA Coordinator

The results from automated analysers were asked in absolute values. The results have been reported in unit $\times 10^9/L$. Different analysers can give results in different units. There were results given in wrong units. These results have been discarded from target value calculation. Please pay attention to choosing the correct unit, so that the results do not have to be unnecessarily removed from result processing. One customer had also clearly mixed the samples with each other, these results have been removed from the results processing. Apart from this and the results given in clearly wrong units, the round went well.

End of Report

2023-04-24

FINAL REPORT

Product no. 4140, 4150-4156

Samples sent	2023-04-03
Round closed	2023-04-20
Final report	2023-04-24

Request for correction

Typing errors in laboratory's result forms are on laboratory's responsibility. Labquality accepts responsibility only for result processing. Requests must be notified by writing within three weeks from the date of this letter.

Authorized by

EQA Coordinator
Iida Silvo
iida.silvo@labquality.fi

Expert

MD, PhD
Anri Tienhaara
TYKSLAB, Turku, Finland

Labquality Oy

Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200
Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.com

