

External Quality Assessment Scheme

## ***Helicobacter pylori*, nucleic acid detection** **Round 1, 2023 (pilot)**

*Thank you for participating in the pilot round. There will be a feedback survey attached to the reports. Please respond to the survey and give us important information on how we can make the EQA scheme better for you.*

### **Specimens**

Please find enclosed 2 artificial swab samples S001 and S002 and 1 lyophilized faecal sample S003.

### **Caution**

Each specimen simulates a clinical specimen and should therefore be handled with the same care as normal patient samples, capable of transmitting infectious disease.

Not for patient use.

### **Examinations**

*Helicobacter pylori*, nucleic acid detection  
Clarithromycin susceptibility

### **Storage and use**

After arrival, the samples should be stored at +2 ... 8°C.

Handling instructions for swabs (S001, S002):

1. Open the foil package. Pay attention to avoid contamination.
2. Remove the swab from the foil package and place the swab into a sample collection tube / transport media (1-3 mL). Rotate the swab several times in the tube to dislodge as much material as possible. Remove the swab by pressing it towards the walls of the elution vial. You can also cut the swab into the sample collection tube.
3. Recap the collection tube and mix thoroughly.
4. Follow the test manufacturer's instructions for sample extraction and detection.

Handling instructions for lyophilized sample (S003):

1. Let the sample vial warm up to room temperature before rehydration.
2. Carefully add 500 µL of distilled or deionized water into sample vial.
3. Let the content of the vial to dissolve for 10 minutes in room temperature and vortex the vial for 30 seconds. Ensure that the mixture is homogenous before performing the assay.
4. Follow the test manufacturer's instructions for sample extraction and detection.

### **Result reporting**

Please enter the results and methods via LabScala ([www.labscala.com](http://www.labscala.com)). If you cannot find your test kit from the registry, please contact the EQA Coordinator. This round has the possibility to report three results from the same sample, if the sample volume is sufficient for several analyses. To open a new result form, press the "Add result +" button on the right side of the blue bar for each sample in LabScala. Report only the results of examinations that are in use in your laboratory. All reported examinations will be scored.

2023-10-17

### **INSTRUCTIONS**

Product no. 5253  
LQ763323011-013/CA, DE  
UN3373

If the kit is incomplete or contains damaged specimens, please report immediately to [info@labquality.fi](mailto:info@labquality.fi).

---

The results should be reported no later than  
**November 8, 2023.**

---

### **Inquiries**

EQA Coordinator  
Kati Lairo  
[kati.lairo@labquality.fi](mailto:kati.lairo@labquality.fi)

### **Labquality Oy**

Kumpulantie 15  
FI-00520 HELSINKI  
Finland

Tel. + 358 9 8566 8200

[info@labquality.fi](mailto:info@labquality.fi)  
[www.labquality.com](http://www.labquality.com)



*Helicobacter pylori*, nucleic acid detection

S001



S002




S003



Client report

	No of participants	No of responded participants	Response percentage
Helicobacter pylori, nucleic acid detection - Pilot, October, 1-2023	17	13	76.5 %

Summary




Summary	Own score	Max score	Own success rate	Difference	AVR success rate
S001	4	4	100 %	0 %	100 %
S002	2	2	100 %	5 %	95 %
S003	2	2	100 %	33.3 %	66.7 %
Average:			100 %	12.8 %	87.2 %

History	Test nr.	Own success rate	Difference	AVR success rate
History not found				

S001

S001 success rate



S001 results	Responded	Own score	Max score	Own success rate	Difference	AVR success rate	Count
	Helicobacter pylori, NAT	2	2	100 %	0 %	100 %	13
	Clarithromycin susceptibility, NAT	2	2	100 %	0 %	100 %	9
	Total:	4	4	100 %	0 %	100 %	34

S001 Helicobacter pylori, NAT



S001 Clarithromycin susceptibility, NAT




Helicobacter pylori, NAT	Interpretation	Test kit	Interpretation count	Test kit count	Own score	Max score	Own success rate	Difference	AVR success rate
	<input checked="" type="radio"/> H. pylori detected		13		2	2	100 %	0 %	100 %
		BioCorp BC-H. pylori		2					
		DNA-Technology Helicobacter pylori REAL-TIME PCR Detection Kit		1					
		In house		2					
		LightGene H. pylori		1					
		Mobidiag Amplidiag® H. pylori+ClariR		3					
		<input checked="" type="radio"/> R-Biopharm RIDAGENE Helicobacter pylori		3					
		Seegene Allplex™ H. pylori & ClariR Assay		1					
	Total:		13		2	2	100 %	0 %	100 %

Clarithromycin susceptibility, NAT	Interpretation	Test kit	Interpretation count	Test kit count	Own score	Max score	Own success rate	Difference	AVR success rate
	<input checked="" type="radio"/> Resistance to clarithromycin not detected		9		2	2	100 %	0 %	100 %
		In house		2					
		Mobidiag Amplidiag® H. pylori+ClariR		3					
		<input checked="" type="radio"/> R-Biopharm RIDAGENE Helicobacter pylori		3					
		Seegene Allplex™ H. pylori & ClariR Assay		1					
	Total:		9		2	2	100 %	0 %	100 %

S002

S002 success rate



S002 results	Responded	Own score	Max score	Own success rate	Difference	AVR success rate	Count
	Helicobacter pylori, NAT	2	2	100 %	7.7 %	92.3 %	13
	Clarithromycin susceptibility, NAT	-	-	-	-	100 %	7
Total:		2	2	100 %	5 %	95 %	31

S002 Helicobacter pylori, NAT



S002 Clarithromycin susceptibility, NAT




■ H. pylori not detected  
■ Invalid/unclear test result

■ Resistance to clarithromycin not detected

Helicobacter pylori, NAT	Interpretation	Test kit	Interpretation count	Test kit count	Own score	Max score	Own success rate	Difference	AVR success rate
	<input checked="" type="radio"/> H. pylori not detected		12		2	2	100 %	0 %	100 %
		BioCorp BC-H. pylori		1					
		DNA-Technology Helicobacter pylori REAL-TIME PCR Detection Kit		1					
		In house		2					
		LightGene H. pylori		1					
		Mobidiag Amplidiag® H. pylori+ClariR		3					
		<input checked="" type="radio"/> R-Biopharm RIDAGENE Helicobacter pylori		3					
		Seegene Allplex™ H. pylori & ClariR Assay		1					
	Invalid/unclear test result		1		-				0 %
		BioCorp BC-H. pylori		1					
Total:			13		2	2	100 %	7.7 %	92.3 %

Clarithromycin susceptibility, NAT	Interpretation	Test kit	Interpretation count	Test kit count	Own score	Max score	Own success rate	Difference	AVR success rate
	Resistance to clarithromycin not detected		7		-				100 %
		In house		1					
		Mobidiag Amplidiag® H. pylori+ClariR		3					
		R-Biopharm RIDAGENE Helicobacter pylori		2					
		Seegene Allplex™ H. pylori & ClariR Assay		1					
Total:			7		-	-	-	-	100 %

S003



S003 results	Responded	Own score	Max score	Own success rate	Difference	AVR success rate	Count
	Helicobacter pylori, NAT	2	2	100 %	33.3 %	66.7 %	12
	Clarithromycin susceptibility, NAT	-	-	-	-	-	8
	<b>Total:</b>	<b>2</b>	<b>2</b>	<b>100 %</b>	<b>33.3 %</b>	<b>66.7 %</b>	<b>20</b>

S003 Helicobacter pylori, NAT



S003 Clarithromycin susceptibility, NAT



Legend: H. pylori detected (orange), H. pylori not detected (dark grey)

Legend: Resistance to clarithromycin detected (orange), Resistance to clarithromycin not detected (dark grey)

Helicobacter pylori, NAT	Interpretation	Test kit	Interpretation count	Test kit count	Own score	Max score	Own success rate	Difference	AVR success rate
	<input checked="" type="radio"/> H. pylori detected		8		2	2	100 %	0 %	100 %
		BioCorp BC-H. pylori		1					
		In house		2					
		LightGene H. pylori		1					
		Mobidiag Amplidiag® H. pylori+ClariR		3					
		<input checked="" type="radio"/> R-Biopharm RIDAGENE Helicobacter pylori		1					
	H. pylori not detected		4		-				0 %
		BioCorp BC-H. pylori		1					
		DNA-Technology Helicobacter pylori REAL-TIME PCR Detection Kit		1					
		R-Biopharm RIDAGENE Helicobacter pylori		1					
		Seegene Allplex™ H. pylori & ClariR Assay		1					
	<b>Total:</b>		<b>12</b>		<b>2</b>	<b>2</b>	<b>100 %</b>	<b>33.3 %</b>	<b>66.7 %</b>


Clarithromycin susceptibility, NAT	Interpretation	Test kit	Interpretation count	Test kit count	Own score	Max score	Own success rate	Difference	AVR success rate
	<input checked="" type="radio"/> Resistance to clarithromycin detected		3		-				-
		Mobidiag Amplidiag® H. pylori+ClariR		2					
		<input checked="" type="radio"/> R-Biopharm RIDAGENE Helicobacter pylori		1					
	Resistance to clarithromycin not detected		5		-				-

		In house		2					
		Mobidiag Amplidiag® H. pylori+ClariR		1					
		R-Biopharm RIDAGENE Helicobacter pylori		1					
		Seegene Allplex™ H. pylori & ClariR Assay		1					
	Total:			8		-	-	-	-

GLOBAL REPORT

	No of participants	No of responded participants	Response percentage
Helicobacter pylori, nucleic acid detection - Pilot, October, 1-2023	17	13	76.5 %

Summary




Summary	AVR success rate
S001	100 %
S002	95 %
S003	66.7 %
Average:	87.2 %



S001

S001 success rate



S001 results	Responded	AVR success rate	Count
	Helicobacter pylori, NAT	100 %	13
	Clarithromycin susceptibility, NAT	100 %	9
	Total:	100 %	34

S001 Helicobacter pylori, NAT



■ H. pylori detected

S001 Clarithromycin susceptibility, NAT




■ Resistance to clarithromycin not detected

Helicobacter pylori, NAT	Interpretation	Test kit	Interpretation count	Test kit count	AVR success rate	Interpretation Score
	H. pylori detected		13		100 %	2
		BioCorp BC-H. pylori		2		
		DNA-Technology Helicobacter pylori REAL-TIME PCR Detection Kit		1		
		In house		2		
		LightGene H. pylori		1		
		Mobidiag Amplidiag® H. pylori+ClariR		3		
		R-Biopharm RIDAGENE Helicobacter pylori		3		
		Seegene Allplex™ H. pylori & ClariR Assay		1		
	Total:		13		100 %	

Clarithromycin susceptibility, NAT	Interpretation	Test kit	Interpretation count	Test kit count	AVR success rate	Interpretation Score
	Resistance to clarithromycin not detected		9		100 %	2
		In house		2		
		Mobidiag Amplidiag® H. pylori+ClariR		3		
		R-Biopharm RIDAGENE Helicobacter pylori		3		
		Seegene Allplex™ H. pylori & ClariR Assay		1		
	Total:		9		100 %	

S002

S002 success rate



S002 results	Responded	AVR success rate	Count
	Helicobacter pylori, NAT	92.3 %	13
	Clarithromycin susceptibility, NAT	100 %	7
	Total:	95 %	31

S002 Helicobacter pylori, NAT



S002 Clarithromycin susceptibility, NAT



■ H. pylori not detected  
■ Invalid/unclear test result


■ Resistance to clarithromycin not detected

Helicobacter pylori, NAT	Interpretation	Test kit	Interpretation count	Test kit count	AVR success rate	Interpretation Score
	H. pylori not detected		12		100 %	2
		BioCorp BC-H. pylori		1		
		DNA-Technology Helicobacter pylori REAL-TIME PCR Detection Kit		1		
		In house		2		
		LightGene H. pylori		1		
		Mobidiag Amplidiag® H. pylori+ClariR		3		
		R-Biopharm RIDAGENE Helicobacter pylori		3		
		Seegene Allplex™ H. pylori & ClariR Assay		1		
	Invalid/unclear test result		1		0 %	0
		BioCorp BC-H. pylori		1		
	Total:		13		92.3 %	

Clarithromycin susceptibility, NAT	Interpretation	Test kit	Interpretation count	Test kit count	AVR success rate	Interpretation Score
	Resistance to clarithromycin not detected		7		100 %	2
		In house		1		
		Mobidiag Amplidiag® H. pylori+ClariR		3		
		R-Biopharm RIDAGENE Helicobacter pylori		2		
		Seegene Allplex™ H. pylori & ClariR Assay		1		
	Total:		7		100 %	

S003

S003 success rate



S003 results	Responded	AVR success rate	Count
	Helicobacter pylori, NAT	66.7 %	12
	Clarithromycin susceptibility, NAT	-	8
	Total:	66.7 %	20

S003 Helicobacter pylori, NAT

S003 Clarithromycin susceptibility, NAT



■ H. pylori detected ■ H. pylori not detected

■ Resistance to clarithromycin detected ■ Resistance to clarithromycin not detected

Helicobacter pylori, NAT	Interpretation	Test kit	Interpretation count	Test kit count	AVR success rate	Interpretation Score
	H. pylori detected		8		100 %	2
		BioCorp BC-H. pylori		1		
		In house		2		
		LightGene H. pylori		1		
		Mobidiag Amplidiag® H. pylori+ClariR		3		
		R-Biopharm RIDAGENE Helicobacter pylori		1		
	H. pylori not detected		4		0 %	0
		BioCorp BC-H. pylori		1		
		DNA-Technology Helicobacter pylori REAL-TIME PCR Detection Kit		1		
		R-Biopharm RIDAGENE Helicobacter pylori		1		
		Seegene Allplex™ H. pylori & ClariR Assay		1		
	Total:		12		66.7 %	

Clarithromycin susceptibility, NAT	Interpretation	Test kit	Interpretation count	Test kit count	AVR success rate	Interpretation Score
	Resistance to clarithromycin detected		3		-	-
		Mobidiag Amplidiag® H. pylori+ClariR		2		
		R-Biopharm RIDAGENE Helicobacter pylori		1		
	Resistance to clarithromycin not detected		5		-	-
		In house		2		
		Mobidiag Amplidiag® H. pylori+ClariR		1		
		R-Biopharm RIDAGENE Helicobacter pylori		1		

		Seegene Allplex™ H. pylori & ClariR Assay		1		
	Total:		8			

External Quality Assessment Scheme

## ***Helicobacter pylori*, nucleic acid detection** **Round 1, 2023 (pilot)**

### **Specimens**

Samples S001 and S002 of this EQA round were artificial swab samples and sample S003 was a lyophilized faecal sample. Based on the quality controls conducted by the sample material manufacturer and the results obtained in the round, the sample lots are to be considered as homogeneous, stable and suitable for external quality assessment. The materials were sent without temperature control packaging.

The expected results were as follows:

Sample S001 (LQ763323011)

*Helicobacter pylori* *H. pylori* detected  
Clarithromycin susceptibility Resistance to clarithromycin not detected

Sample S002 (LQ763323012)

*Helicobacter pylori* *H. pylori* not detected  
Clarithromycin susceptibility Resistance to clarithromycin not detected

Sample S003 (LQ763323013)

*Helicobacter pylori* *H. pylori* detected  
Clarithromycin susceptibility Resistance to clarithromycin not analysed

### **Report info**

Please see the description of the data analysis on the last page of the laboratory-specific reports and global reports. It is important to read the Final report first, because it contains important information of the samples and results in each round.

### **Comments – Expert**

In this pilot round 13/17 participants (12 participants in sample S003) had reported the results on time.

Eight different methods had been used and with five methods it was possible to determine clarithromycin resistance. The success rate for this round was 87.2%.

Sample S001 was *Helicobacter pylori* positive, and no clarithromycin resistance was detected. The analysis of this sample went very well, and all the laboratories reported expected results.

Sample S002 was negative. 95% of the laboratories reported expected negative results for *H. pylori*. One laboratory reported “Unclear test result”. No clear reason for this was found. All the laboratories reported expected negative results for clarithromycin.

Sample S003 was *H. pylori* positive. The sample was challenging and only 8/12 (66.7%) of the laboratories reported the expected result. Two methods yielded both positive and negative results. Clarithromycin resistance test was not scored although over 60% of laboratories reported “Resistance to clarithromycin not detected”. Two methods out of five yielded both “Resistance to clarithromycin not detected” and “Resistance to clarithromycin detected” results.

Sample S003 was a low positive sample which might have affected the variation of the results of different laboratories. Unlike samples S001 and S002, sample S003 was a lyophilized faecal sample.

2023-11-30

### **REPORT**

Product no. 5253

Samples sent 2023-10-17  
Round closed 2023-11-16  
Final report 2023-11-30

### **Request for correction**

Typing errors in laboratory’s result forms are on laboratory’s responsibility. Labquality accepts responsibility only for result processing. Requests must be notified by writing within three weeks from the date of this letter.

### **Authorized by**

EQA Coordinator  
Kati Lairo  
kati.lairo@labquality.fi

### **Expert**

Clinical microbiologist  
Santra Karumaa,  
SYNLAB Suomi Oy

### **Labquality Oy**

Kumpulantie 15  
FI-00520 HELSINKI  
Finland

Tel. + 358 9 8566 8200

info@labquality.fi  
www.labquality.com



Both the pre-treatment of this sample and the deviant sample type may have also affected the results. This type of sample matrix is not routinely validated in many laboratories or for many nucleic acid detection tests (in-house or commercial).

#### **Exceptions in scoring**

For sample S003, the clarithromycin resistance results were not scored due to low *H. pylori* concentration in the sample, and also the deviant sample type and success of the pre-treatment could have affected the variation of the results.

**End of report**