

External Quality Assessment Scheme

LMW-Heparin/anti-FXa Round 1, 2023

Specimens

Please find enclosed 2 lyophilized heparin plasma samples S001 and S002.

Caution

Quality control specimens derived from human blood must be handled with the same care as patient samples, i.e. as potential transmitters of serious diseases. The specimens are found to be HBsAg, HCVAb and HIVAgAb negative when tested with licensed reagents, but no known test method can offer complete assurance that the specimens will not transmit these or other infectious diseases.

Examinations

Anti-FXa

Storage and use

After arrival store the unopened vials in a refrigerator (+2...8 °C). Open the vials carefully to prevent escape of dried material and reconstitute the contents in 1.0 mL of distilled water. Allow samples to stand for 15 minutes at room temperature. After that mix the samples gently by inverting the tubes several times. Avoid foam formation. Samples are stable for 4 hours after reconstitution. Analyse as a patient sample.

Result reporting

Please enter the results and methods via LabScala (www.labscala.com). If you cannot find your instrument or reagent from the registry, please contact the EQA Coordinator.

S001



S002



2023-02-06

INSTRUCTIONS

Product no. 4387
LQ708623011-012/AT

If the kit is incomplete or contains damaged specimens, please report immediately to info@labquality.fi

The results should be reported no later than **February 21, 2023.**

Inquiries

EQA Coordinator
Iida Silvo
iida.silvo@labquality.fi

Labquality Oy




Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200
Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.fi



LMW heparin |koagulologia



	x_{pt}	sd	SEM	CV%	n
HemosIL Liquid Anti-Xa	0.37 IU/ml	0.05	0.01	12.4	18
All methods	0.41 IU/ml	0.07	<0.01	17.3	66

	x_{pt}	sd	SEM	CV%	n
HemosIL Liquid Anti-Xa	0.91 IU/ml	0.06	0.02	7.1	18
All methods	1.04 IU/ml	0.11	0.01	10.8	66

Round	Sample	x_{pt}	Result	diff%	z-score
23/1	Sample S002	0.91	0.93	1.95%	0.28
23/1	Sample S001	0.37	0.39	6.04%	0.49
22/4	Sample S002	0.37	0.37	0.16%	0.02
22/4	Sample S001	0.54	0.54	0.66%	0.13
22/3	Sample S002	0.90	0.88	-2.43%	-0.73
22/3	Sample S001	0.37	0.36	-3.52%	-0.58
22/2	Sample S002	0.89	0.92	2.88%	0.53
22/2	Sample S001	0.37	0.42	14.37%	1.18
22/1	Sample S002	0.65	0.67	2.84%	0.58
22/1	Sample S001	0.37	0.38	3.12%	0.39

Report info**Participants**

41 participants from 13 countries.

Report info

Your own result should be compared to others using the same method.

Assigned values (\bar{x}_p , target values) are means of the results where results deviating more than ± 3 standard deviation from the median are removed. The standard uncertainty (u) of

the assigned value is reported as standard error of the mean (SEM). Additionally, if the measurement uncertainty of the target value is large an automatic text is printed on the report: "The uncertainty of the assigned value is not negligible, and evaluations could be affected."

In case the client's result is the only one in the method group, no assigned value will be calculated, no target area shown, and no statistics calculated. In case there are only a few results in the client's own method group, the result can be compared to all method mean or to a group that is similar to the own method.





Results reported with $<$ or $>$ -signs cannot be included in the statistics.



For information on report interpretation and performance evaluation, please see the "EOAS Interpretation guidelines" LabScala User instructions (top right corner ?Help link).

Sample S001 | LMW heparin, IU/ml

Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Berichrom Heparin	-	-	-	-	-	0.47	0.47	-	1
Biophen Heparin	-	-	-	-	-	0.52	0.52	-	1
HemosIL Liquid Anti-Xa	0.37	0.38	0.05	12.4	0.01	0.25	0.41	-	18
Innovance Anti-Xa	-	-	-	-	-	0.44	0.44	-	1
Innovance Heparin	0.46	0.47	0.05	10.0	<0.01	0.34	0.55	-	32
Stago Liquid anti-Xa	0.33	0.33	0.03	9.0	<0.01	0.27	0.38	1	13
All	0.41	0.40	0.07	17.3	<0.01	0.25	0.55	1	66

Sample S001 | LMW heparin, IU/ml| histogram summaries in LabScala










Sample S002 | LMW heparin, IU/ml


Methodics	x_{pt}	Median	sd	CV%	SEM	min	max	Outliers	n
Berichrom Heparin	-	-	-	-	-	1.16	1.16	-	1
Biophen Heparin	-	-	-	-	-	1.23	1.23	-	1
HemosIL Liquid Anti-Xa	0.91	0.93	0.06	7.1	0.02	0.77	1.02	-	18
Innovance Anti-Xa	-	-	-	-	-	1.15	1.15	-	1
Innovance Heparin	1.12	1.12	0.04	3.9	<0.01	1.05	1.22	1	32
Stago Liquid anti-Xa	0.98	0.98	0.07	7.6	0.02	0.80	1.08	-	13
All	1.04	1.07	0.11	10.8	0.01	0.77	1.23	-	66

Sample S002 | LMW heparin, IU/ml| histogram summaries in LabScala





■ All method groups ■ Innovance Heparin (x_{pt} : 1.12 | Target area: 0.95-1.29 | Target: $\pm 15\%$)



■ All method groups ■ Stago Liquid anti-Xa (x_{pt} : 0.98 | Target area: 0.83-1.12 | Target: $\pm 15\%$)

Report info**Participants**

41 participants from 13 countries.

Report info

Your own result should be compared to others using the same method.

Assigned values (\bar{x}_p , target values) are means of the results where results deviating more than ± 3 standard deviation from the median are removed. The standard uncertainty (u) of

the assigned value is reported as standard error of the mean (SEM). Additionally, if the measurement uncertainty of the target value is large an automatic text is printed on the report: "The uncertainty of the assigned value is not negligible, and evaluations could be affected."

In case the client's result is the only one in the method group, no assigned value will be calculated, no target area shown, and no statistics calculated. In case there are only a few results in the client's own method group, the result can be compared to all method mean or to a group that is similar to the own method.

Results reported with $<$ or $>$ -signs cannot be included in the statistics.

For information on report interpretation and performance evaluation, please see the "EOAS Interpretation guidelines" LabScala User instructions (top right corner ?Help link).

External Quality Assessment Scheme

LMWH-Heparin/anti-FXa Round 1, 2023

Specimens

Sample S001 (LQ7086223011) and sample S002 (LQ708623012) were lyophilized plasma samples.

Based on the previous tests and the results of this round, the samples are homogeneous, stable and suitable for the external quality assessment scheme.

The materials were sent without temperature control packaging.

Report info

Please see the description of the data analysis on the last page of the laboratory-specific histograms and Numerical Summary reports. It is important to read the Final report first, because it contains important information of the samples and results in each round.

Comments – EQA Coordinator

Sample S001 was low-level sample, the average of all results was 0.41 IU/mL. Results ranged from 0.25 to 0.44 IU/mL. Sample S002 was a high-level sample. The mean of all results for sample S002 was 1.04 IU/mL. The results ranged from 0.77 to 1.23 IU/mL. The results were uniform within the method groups, with the exception of a few deviating results.

End of report

2023-02-23

FINAL REPORT

Product no. 4387

Samples sent	2023-02-06
Round closed	2023-02-21
Final report	2023-02-23

Request for correction

Typing errors in laboratory's result forms are on laboratory's responsibility. Labquality accepts responsibility only for result processing. Requests must be notified by writing within three weeks from the date of this letter.

Authorized by

EQA Coordinator
Iida Silvo
iida.silvo@labquality.fi

Expert

Chief physician, Docent
Lotta Joutsu-Korhonen
HUSLAB, Helsinki, Finland

Labquality Oy

Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200

Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.com



Copyright © Labquality Oy

Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copyright. Labquality prohibits any use of its name, or reference to Labquality EQA program, or material in this report in any advertising, brochures or other commercial publications. Labquality EQA data do not necessarily indicate the superiority of instruments, reagents, testing equipments or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of equipments or materials may be deceptive and misleading. Proficiency test results are handled confidentially. Labquality will not issue any statements to third parties of the performance of laboratories in external quality assessment schemes unless otherwise agreed.