

External Quality Assessment Scheme

***Mycobacterium tuberculosis*, drug resistance**

Round 1, 2023

Specimens

Please find enclosed 2 artificial sputum samples (primary samples) S001 and S002, each 1 mL.

Caution

Quality control specimens must be handled with the same care as patient samples, i.e. as potential transmitters of serious diseases. According to the sample manufacturer the specimens do not contain viable microbes and the specimens are found to be HBsAg, HCVAb and HIVAb negative when tested with licensed reagents. However, no known test method can offer complete assurance that the specimens will not transmit these or other infectious diseases.

Examinations

Mycobacterium tuberculosis, nucleic acid detection
Rifampicin susceptibility
Isoniazid susceptibility

Storage and use

After arrival, the samples should be stored at +2 ... 8 °C, and used as soon as possible, preferably within a week. The samples are artificial primary samples (do not contain viable microbes) and cannot be enriched by culturing. The samples should not be decontaminated. Perform the analysis according to the manufacturer's instructions for a liquid specimen.

Result reporting

Please enter the results and methods via LabScala (www.labscala.com). If you cannot find your test kit from the registry, please contact the EQA Coordinator. This round has the possibility to report three results from the same sample, if the sample volume is sufficient for several analyses. To open a new result form, press the "Add result +" button on the right side of the blue bar for each sample in LabScala. Report only the results of examinations that are in use in your laboratory. All reported examinations will be scored.

S001



S002



2023-03-14

INSTRUCTIONS

Product no. 5230
LQ762523011-012/US

Subcontracting: Sample pretesting

If the kit is incomplete or contains damaged specimens, please report immediately to info@labquality.fi.

The results should be reported no later than **April 6, 2023.**

The expected results of the round are published in LabScala in the View Reports section by April 12, 2023.

Inquiries

EQA Coordinator
Jaana Paakkanen
jaana.paakkanen@labquality.fi

Labquality Oy

Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200
Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.com




Client report

	No of participants	No of responded participants	Response percentage
Mycobacterium tuberculosis, drug resistance, March, 1-2023	21	19	90.5 %

Summary

Overall success rate by samples




Summary	Own score	Max score	Own success rate	Difference	AVR success rate
S001	2	2	100 %	7.1 %	92.9 %
S002	4	4	100 %	0 %	100 %
Average:			100 %	3.6 %	96.4 %

History	Test nr.	Own success rate	Difference	AVR success rate
History not found				


S001

S001 success rate




S001 results	Responded	Own score	Max score	Own success rate	Difference	AVR success rate	Count
	Mycobacterium tuberculosis	2	2	100 %	10.5 %	89.5 %	19
	Rifampicin susceptibility	-	-	-	-	-	19
	Isoniazid susceptibility	-	-	-	-	100 %	9
	Total:	2	2	100 %	7.1 %	92.9 %	47

S001 Mycobacterium tuberculosis



S001 Rifampicin susceptibility



S001 Isoniazid susceptibility



■ MTBC detected ■ MTBC not detected
 ■ Resistance to rifampicin detected ■ Resistance to rifampicin inferred ■ Resistance to rifampicin indeterminate
 ■ Resistance to isoniazid not detected ■ Resistance to isoniazid indeterminate


Mycobacterium tuberculosis	Interpretation	Test kit	Interpretation count	Test kit count	Own score	Max score	Own success rate	Difference	AVR success rate
	<input checked="" type="radio"/> MTBC detected		17		2	2	100 %	0 %	100 %
		Xpert MTB/RIF (Cepheid)		1					
		<input checked="" type="radio"/> Xpert MTB/RIF Ultra (Cepheid)		16					
	MTBC not detected		2		-				0 %
		GenoType MTBDRplus 2.0 (Hain Lifescience)		2					
	Total:		19		2	2	100 %	10.5 %	89.5 %

Rifampicin susceptibility	Interpretation	Test kit	Interpretation count	Test kit count	Own score	Max score	Own success rate	Difference	AVR success rate
	<input checked="" type="radio"/> Resistance to rifampicin detected		10		-				-
		Xpert MTB/RIF (Cepheid)		1					
		<input checked="" type="radio"/> Xpert MTB/RIF Ultra (Cepheid)		9					
	Resistance to rifampicin inferred		1		-				-
		Xpert MTB/RIF Ultra (Cepheid)		1					
	Resistance to rifampicin indeterminate		8		-				-
		GenoType MTBDRplus 2.0 (Hain Lifescience)		2					
		Xpert MTB/RIF (Cepheid)		1					
		Xpert MTB/RIF Ultra (Cepheid)		5					
	Total:		19		-	-	-	-	

Isoniazid susceptibility	Interpretation	Test kit	Interpretation count	Test kit count	Own score	Max score	Own success rate	Difference	AVR success rate
	Resistance to isoniazid not detected		5		-				100 %
		GenoType MTBDRplus 2.0 (Hain Lifescience)		1					
		Xpert MTB/XDR (Cepheid)		4					
	Resistance to isoniazid indeterminate		4		-				100 %
		GenoType MTBDRplus 2.0 (Hain Lifescience)		2					
		Xpert MTB/XDR (Cepheid)		2					
	Total:		9		-	-	-	-	100 %

S002

S002 success rate



S002 results	Responded	Own score	Max score	Own success rate	Difference	AVR success rate	Count
	Mycobacterium tuberculosis	2	2	100 %	0 %	100 %	19
	Rifampicin susceptibility	2	2	100 %	0 %	100 %	19
	Isoniazid susceptibility	-	-	-	-	100 %	9
Total:		4	4	100 %	0 %	100 %	47

S002 Mycobacterium tuberculosis



MTBC detected

S002 Rifampicin susceptibility



Resistance to rifampicin detected
Resistance to rifampicin inferred

S002 Isoniazid susceptibility



Resistance to isoniazid not detected

Mycobacterium tuberculosis	Interpretation	Test kit	Interpretation count	Test kit count	Own score	Max score	Own success rate	Difference	AVR success rate
	MTBC detected		19		2	2	100 %	0 %	100 %
		GenoType MTBDRplus 2.0 (Hain Lifescience)		2					
		Xpert MTB/RIF (Cepheid)		2					
		Xpert MTB/RIF Ultra (Cepheid)		15					
Total:			19		2	2	100 %	0 %	100 %

Rifampicin susceptibility	Interpretation	Test kit	Interpretation count	Test kit count	Own score	Max score	Own success rate	Difference	AVR success rate
	Resistance to rifampicin detected		18		2	2	100 %	0 %	100 %
		GenoType MTBDRplus 2.0 (Hain Lifescience)		2					
		Xpert MTB/RIF (Cepheid)		2					
		Xpert MTB/RIF Ultra (Cepheid)		14					
	Resistance to rifampicin inferred		1		-				100 %
		Xpert MTB/RIF Ultra (Cepheid)		1					
Total:			19		2	2	100 %	0 %	100 %

Isoniazid susceptibility	Interpretation	Test kit	Interpretation count	Test kit count	Own score	Max score	Own success rate	Difference	AVR success rate
	Resistance to isoniazid not detected		9		-				100 %


		GenoType MTBDRplus 2.0 (Hain Lifescience)		3					
		Xpert MTB/XDR (Cepheid)		6					
	Total:		9		-	-	-	-	100 %

GLOBAL REPORT

	No of participants	No of responded participants	Response percentage
Mycobacterium tuberculosis, drug resistance, March, 1-2023	21	19	90.5 %

Summary


Overall success rate by samples



Summary	AVR success rate
S001	92.9 %
S002	100 %
Average:	96.4 %


S001

S001 success rate




S001 results	Responded	AVR success rate	Count
	Mycobacterium tuberculosis	89.5 %	19
	Rifampicin susceptibility	-	19
	Isoniazid susceptibility	100 %	9
	Total:	92.9 %	47

S001 Mycobacterium tuberculosis



S001 Rifampicin susceptibility



S001 Isoniazid susceptibility



■ MTBC detected ■ MTBC not detected
 ■ Resistance to rifampicin detected ■ Resistance to rifampicin inferred ■ Resistance to rifampicin indeterminate
 ■ Resistance to isoniazid not detected ■ Resistance to isoniazid indeterminate

Mycobacterium tuberculosis	Interpretation	Test kit	Interpretation count	Test kit count	AVR success rate	Interpretation Score
	MTBC detected		17		100 %	2
		Xpert MTB/RIF (Cepheid)		1		
		Xpert MTB/RIF Ultra (Cepheid)		16		
	MTBC not detected		2		0 %	0
		GenoType MTBDRplus 2.0 (Hain Lifescience)		2		
	Total:		19		89.5 %	


Rifampicin susceptibility	Interpretation	Test kit	Interpretation count	Test kit count	AVR success rate	Interpretation Score
	Resistance to rifampicin detected		10		-	-
		Xpert MTB/RIF (Cepheid)		1		
		Xpert MTB/RIF Ultra (Cepheid)		9		
	Resistance to rifampicin inferred		1		-	-
		Xpert MTB/RIF Ultra (Cepheid)		1		
	Resistance to rifampicin indeterminate		8		-	-
		GenoType MTBDRplus 2.0 (Hain Lifescience)		2		
		Xpert MTB/RIF (Cepheid)		1		
		Xpert MTB/RIF Ultra (Cepheid)		5		
	Total:		19			

Isoniazid susceptibility	Interpretation	Test kit	Interpretation count	Test kit count	AVR success rate	Interpretation Score
--------------------------	----------------	----------	----------------------	----------------	------------------	----------------------

	Resistance to isoniazid not detected		5		100 %	2
		GenoType MTBDRplus 2.0 (Hain Lifescience)		1		
		Xpert MTB/XDR (Cepheid)		4		
	Resistance to isoniazid indeterminate		4		100 %	2
		GenoType MTBDRplus 2.0 (Hain Lifescience)		2		
		Xpert MTB/XDR (Cepheid)		2		
	Total:		9		100 %	

S002

S002 success rate



S002 results	Responded	AVR success rate	Count
	Mycobacterium tuberculosis	100 %	19
	Rifampicin susceptibility	100 %	19
	Isoniazid susceptibility	100 %	9
	Total:	100 %	47

S002 Mycobacterium tuberculosis



MTBC detected

S002 Rifampicin susceptibility



Resistance to rifampicin detected
Resistance to rifampicin inferred

S002 Isoniazid susceptibility



Resistance to isoniazid not detected

Mycobacterium tuberculosis	Interpretation	Test kit	Interpretation count	Test kit count	AVR success rate	Interpretation Score
	MTBC detected		19		100 %	2
		GenoType MTBDRplus 2.0 (Hain Lifescience)		2		
		Xpert MTB/RIF (Cepheid)		2		
		Xpert MTB/RIF Ultra (Cepheid)		15		
	Total:		19		100 %	

Rifampicin susceptibility	Interpretation	Test kit	Interpretation count	Test kit count	AVR success rate	Interpretation Score
	Resistance to rifampicin detected		18		100 %	2
		GenoType MTBDRplus 2.0 (Hain Lifescience)		2		
		Xpert MTB/RIF (Cepheid)		2		
		Xpert MTB/RIF Ultra (Cepheid)		14		
	Resistance to rifampicin inferred		1		100 %	2
		Xpert MTB/RIF Ultra (Cepheid)		1		
	Total:		19		100 %	

Isoniazid susceptibility	Interpretation	Test kit	Interpretation count	Test kit count	AVR success rate	Interpretation Score
	Resistance to isoniazid not detected		9		100 %	2
		GenoType MTBDRplus 2.0 (Hain Lifescience)		3		
		Xpert MTB/XDR (Cepheid)		6		
	Total:		9		100 %	

External Quality Assessment Scheme

***Mycobacterium tuberculosis*, drug resistance** **Round 1, 2023**

Specimens

Samples of this EQA round were artificial sputum samples (primary samples). Based on the quality controls conducted by the sample material manufacturer and the results obtained in the round, the sample lots are to be considered as homogeneous, stable and suitable for external quality assessment. The materials were sent without temperature control packaging.

The expected results were as follows:

Sample S001 (LQ762523011)
Mycobacterium tuberculosis MTBC detected
Rifampicin susceptibility Resistance to rifampicin detected (low)
Isoniazid susceptibility Resistance to isoniazid not detected

Sample S002 (LQ762523012)
Mycobacterium tuberculosis MTBC detected
Rifampicin susceptibility Resistance to rifampicin detected (high)
Isoniazid susceptibility Resistance to isoniazid not detected

Pre-test methods: Xpert MTB/RIF (Cepheid) and GenoType MTBDRplus 2.0 (Hain Lifescience).

Report info

Please see the description of the data analysis on the last page of this letter as an annex (Annex 1). It is important to read the Final report first, because it contains important information of the samples and results in each round.

Comments – Expert

This round consisted of a single rifampicin resistant and isoniazid susceptible sample in low (sample S001) and high (sample S002) concentration.

The performance for the high concentration sample S002 was excellent with 100% success rates for all targets (MTBC detection, susceptibility to rifampicin and isoniazid). For the low concentration sample S001, the results varied, which was expected due to low template number. There were 2/19 false negative results for MTBC detection and 8/19 indeterminate results for rifampicin and 4/9 for isoniazid compared with none in sample S002. The high proportion of indeterminate results by all used tests refers to proximity to the limit of detection, where result may remain unavailable. For samples that are close to the limit of detection, optimal testing conditions are crucial.

The pretesting with GenoType MTBDRplus was performed with PCR cycle number 50 as instructed for primary samples, instead of 30 cycles for cultures. Testing yielded expected results, though inhA WT1 was weak, and on re-testing there was no katG or inhA product at all (competition in PCR).

Exceptions in scoring

Rifampicin susceptibility in sample S001 was not scored as the success rate was below 60%.

Annex

Annex 1. Report info.

2023-04-13

FINAL REPORT

Product no. 5230

Subcontracting: Sample pretesting

Samples sent	2023-03-14
Round closed	2023-04-06
Expected results	2023-04-12
Final report	2023-04-13

Request for correction

Typing errors in laboratory's result forms are on laboratory's responsibility. Labquality accepts responsibility only for result processing. Requests must be notified by writing within three weeks from the date of this letter.

Authorized by

EQA Coordinator
Jaana Paakkanen
jaana.paakkanen@labquality.fi

Expert

Microbiologist, Ph.D, Docent
Silja Mentula,
National Institute for Health
and Welfare, Finland

Labquality Oy

Kumpulantie 15
FI-00520 HELSINKI
Finland

Tel. + 358 9 8566 8200
Fax + 358 9 8566 8280

info@labquality.fi
www.labquality.com



End of report

Copyright © Labquality Oy

Labquality does not permit any reproduction for commercial purposes of any portion of the material subject to this copyright. Labquality prohibits any use of its name, or reference to Labquality EQA program, or material in this report in any advertising, brochures or other commercial publications. Labquality EQA data do not necessarily indicate the superiority of instruments, reagents, testing equipments or materials used by participating laboratories. Use of Labquality EQA data to suggest superiority or inferiority of equipments or materials may be deceptive and misleading. Proficiency test results are handled confidentially. Labquality will not issue any statements to third parties of the performance of laboratories in external quality assessment schemes unless otherwise agreed.

Annex 1.

Participants

Altogether 21 laboratories from 8 countries participated in this EQA round.

Report info

The results are divided into groups according to the method stated by the laboratory and presented in laboratory-specific tables. In general, the expected results are marked with green color. Laboratory's own result is marked with a black radio button (Ⓞ). In the scoring report you will find summaries of overall success rate and sample specific success rates (%). Sample specific interpretations are shown in pie diagrams as percentages and the total interpretation and methodic counts in the tables. If you have not reported any results you will get a note: "You have not responded in time, only global report is available."

For information on report interpretation and performance evaluation, please see the "EQAS Interpretation guidelines" in LabScala User instructions. In case you have any questions regarding the reports, please contact the EQA Coordinator.

Scoring

The round is scored based on test results when 60% or more of the participants report the expected/correct result and when at least three results are reported.

The following general rules are applied:

- 2/2 points is given to results that are correct/accepted regarding the expected result
- 1/2 point can be given to results that are partly correct/insufficient regarding the expected result
- 0/2 point is given to results that are incorrect/false regarding the expected result

The performance of the laboratory is assessed by the Own success rate (%). The target is 100%. The examination-specific scores obtained by the laboratory in the round are converted to Own success rate per sample (scores/maximum scores*100). The Laboratory's Own success rate is the average of the sample success rates.

The success rate for the entire round (AVR success rate) is calculated from the total number of scores given to the results per sample (all scores/maximum scores*100). The AVG success rate of the entire round is the average of the sample success rates. The difference in the Laboratory's Own success rate (%) to the corresponding numbers for the entire round is shown in the table.